



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93412

Temp.: 23.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/20/2018	15:35
2	CAL2	1	Water			7.00	02/20/2018	15:37
3	CAL3	1	Water			10.01	02/20/2018	15:39
4	ICV	1	Water			7.00	02/20/2018	15:41
5	CCV1	1	Water			2.00	02/20/2018	15:43
6	J1626-01	1	Water			9.01	02/20/2018	15:45
7	J1626-01DUP	1	Water			9.02	02/20/2018	15:47
8	CCV2	1	Water			12.00	02/20/2018	15:50

WORKLIST(Hardcopy Internal Chain)

13 93412

Worklist Name : ph j1626

Worklist ID : 108752

Date : 2/20/2018 3:25:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2018	Water	J1626-01-1F	pH	Cool 4 deg C	PSEG01	P63	SILVERLAKE-LAYDOWN-YAF	02/20/2018	9040C

Date/Time 02.20.18 3:30PM
 Received by: JP
 Relinquished by: C

Date/Time 02.20.18 5:10PM
 Received by: C
 Relinquished by: JP