


7471B

LB93416
INSTRUMENT ID: CY1

Linear

 μ Abs.:

43844



A= 0.0000e+000

B= 2.3101e-004

C= -3.3820e-002

Rho= 0.9996848

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.047	-0.047	-56	0.000	-56				
0.2	0.200	0.212	0.012	1066	0.0 %	1066				
2.5	2.500	2.619	0.119	11485	0.0 %	11485				
5.0	5.000	4.980	-0.020	21706	0.0 %	21706				
7.5	7.500	7.340	-0.160	31920	0.0 %	31920				
10	10.000	10.095	0.095	43844	0.0 %	43844				

LB93416 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Feb 2018 10:18:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-56	-	0.00007471B	PPB		21 Feb 2018 10:36:12	S	Std
0.2 - 1	50.2	1066	-	0.20007471B	PPB		21 Feb 2018 10:38:29	S	Std
2.5 - 1	52.5	11485	-	2.50007471B	PPB		21 Feb 2018 10:40:47	S	Std
5.0 - 1	55	21706	-	5.00007471B	PPB		21 Feb 2018 10:43:04	S	Std
7.5 - 1	57.5	31920	-	7.50007471B	PPB		21 Feb 2018 10:45:21	S	Std
10 - 1	510	43844	-	10.00007471B	PPB		21 Feb 2018 10:47:38	S	Std
ICV10 - 1	ICV10	17053	3.9056	-7471B	PPB		21 Feb 2018 10:50:43	U	SMPL
ICB10 - 1	ICB10	155	0.0020	-7471B	PPB		21 Feb 2018 10:52:59	U	SMPL
CCV86 - 1	CCV86	21635	4.9641	-7471B	PPB		21 Feb 2018 10:55:14	U	SMPL
CCB86 - 1	CCB86	148	0.0004	-7471B	PPB		21 Feb 2018 10:57:30	U	SMPL
CRA - 1	CRA	1065	0.2122	-7471B	PPB		21 Feb 2018 10:59:46	U	SMPL
HighStd - 1	HighStd	44520	10.2507	-7471B	PPB		21 Feb 2018 11:02:02	U	SMPL
ChkStd - 1	ChkStd	29665	6.8191	-7471B	PPB		21 Feb 2018 11:04:18	U	SMPL
PB106707BL - 1	PBS	45	-0.0234	-7471B	PPB		21 Feb 2018 11:06:36	U	SMPL
PB106707BS - 1	LCSS	14832	3.3925	-7471B	PPB		21 Feb 2018 11:08:54	U	SMPL
J1391-01 - 1	AU-05-022018-A	1162	0.2346	-7471B	PPB		21 Feb 2018 11:11:11	U	SMPL
J1391-03 - 1	AU-05-022018-B	1248	0.2545	-7471B	PPB		21 Feb 2018 11:13:28	U	SMPL
J1391-03DUP - 1	AU-05-022018-BDUP	1171	0.2367	-7471B	PPB		21 Feb 2018 11:15:46	U	SMPL
J1391-03MS - 1	AU-05-022018-BMS	18028	4.1308	-7471B	PPB		21 Feb 2018 11:18:04	U	SMPL
J1391-03MSD - 1	AU-05-022018-BMSD	18518	4.2440	-7471B	PPB		21 Feb 2018 11:20:19	U	SMPL
CCV87 - 1	CCV87	21691	4.9770	-7471B	PPB		21 Feb 2018 11:22:36	U	SMPL
CCB87 - 1	CCB87	149	0.0006	-7471B	PPB		21 Feb 2018 11:24:51	U	SMPL
J1398-03 - 1	OK-02-022018	2353	0.5097	-7471B	PPB		21 Feb 2018 11:27:07	U	SMPL
J1403-05 - 1	SU-04-022018	2385	0.5171	-7471B	PPB		21 Feb 2018 11:29:24	U	SMPL
J1404-05 - 1	TR-05-022018	3170	0.6985	-7471B	PPB		21 Feb 2018 11:31:40	U	SMPL
J1622-01 - 1	MAD-SUB-TRENCH	30952	7.1164	-7471B	PPB		21 Feb 2018 11:33:56	U	SMPL
J1623-01 - 1	TP-1-2	5625	1.2656	-7471B	PPB		21 Feb 2018 11:36:13	U	SMPL
J1623-05 - 1	TP-3-4	1272	0.2600	-7471B	PPB		21 Feb 2018 11:38:31	U	SMPL
J1623-09 - 1	TP-5-6	2265	0.4894	-7471B	PPB		21 Feb 2018 11:40:48	U	SMPL
J1623-13 - 1	TP-7-8	20963	4.8088	-7471B	PPB		21 Feb 2018 11:43:06	U	SMPL
J1623-17 - 1	TP-9-10	24933	5.7259	-7471B	PPB		21 Feb 2018 11:45:23	U	SMPL
J1624-01 - 1	TP-1	1623	0.3411	-7471B	PPB		21 Feb 2018 11:47:39	U	SMPL
CCV88 - 1	CCV88	21609	4.9581	-7471B	PPB		21 Feb 2018 11:49:56	U	SMPL
CCB88 - 1	CCB88	157	0.0024	-7471B	PPB		21 Feb 2018 11:52:13	U	SMPL
J1624-03 - 1	TP-2	858	0.1644	-7471B	PPB		21 Feb 2018 11:54:29	U	SMPL
J1624-05 - 1	TP-3	972	0.1907	-7471B	PPB		21 Feb 2018 11:56:45	U	SMPL
J1391-03LX5 - 1		424	0.0641	-7471B	PPB		21 Feb 2018 11:59:01	U	SMPL
J1391-03A - 1		17960	4.1151	-7471B	PPB		21 Feb 2018 12:01:17	U	SMPL
CCV89 - 1	CCV89	21843	5.0121	-7471B	PPB		21 Feb 2018 12:03:34	U	SMPL
CCB89 - 1	CCB89	139	-0.0017	-7471B	PPB		21 Feb 2018 12:05:51	U	SMPL