



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93420

Temp.: 23.3 °C

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/21/2018	09:10
2	CAL2	1	Water			7.00	02/21/2018	09:12
3	CAL3	1	Water			10.01	02/21/2018	09:14
4	ICV	1	Water			7.00	02/21/2018	12:30
5	CCV1	1	Water			2.00	02/21/2018	12:32
6	J1632-01	1	Water			9.51	02/21/2018	12:34
7	J1632-01DUP	1	Water			9.52	02/21/2018	12:36
8	CCV2	1	Water			12.00	02/21/2018	12:38

WORKLIST(Hardcopy Internal Chain)

LB 93420

Worklist Name : ph j1632

Worklist ID : 108793

Date : 2/21/2018 12:20:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2018	Water	J1632-01-E	pH	Cool 4 deg C	VIAS01	P43	EFF-180220-01	02/20/2018	SM 4500-PH B

Date/Time

02/21/18 12:25 PM

Received by:

SP

Relinquished by:

CP

Date/Time

02/21/18 12:45 PM

Received by:

CP

Relinquished by:

SP