

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AS

2/21/2018 13:11

Test: ^{Reactive}~~Total~~ CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.257	0.0	0.089	
ICB1	-0.274	0.0	0.003	
CCV1	240.865	0.0	0.216	
CCB1	0.238	0.0	0.004	
PB106720BL	-0.064	0.0	0.004	
J1391-02	0.024	0.0	0.004	
J1391-02DUP	-0.080	0.0	0.004	
J1403-06	0.080	0.0	0.004	
J1404-06	0.439	0.0	0.004	
J1622-02	-0.169	0.0	0.004	
J1623-04	-0.214	0.0	0.003	
J1623-08	-0.024	0.0	0.004	
CCV2	241.568	0.0	0.217	
CCB2	0.067	0.0	0.004	
J1623-12	-0.479	0.0	0.003	
J1623-16	-0.360	0.0	0.003	
J1623-20	-0.308	0.0	0.003	
J1624-02	-0.306	0.0	0.003	
J1624-04	-0.217	0.0	0.003	
J1624-06	-0.319	0.0	0.003	
PB106721BL	-0.115	0.0	0.004	
J1626-02	-0.133	0.0	0.004	
J1626-02DUP	-0.095	0.0	0.004	
CCV3	258.018	0.0	0.232	
CCB3	0.406	0.0	0.004	
J1391-04	-0.217	0.0	0.003	
J1398-04	-0.267	0.0	0.003	
CCV4	256.915	0.0	0.231	
CCB4	0.109	0.0	0.004	

N	29
Mean	37.633
SD	88.0638
CV%	234.01

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Feb 21 11:59:28 2018

Wed Feb 21 13:04:06 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.3338	µg/l	2/21/2018 10:36:00	
5.OPPBCN	A	Reactive CN	P	4.8199	µg/l	2/21/2018 10:36:01	
10PPBCN	A	Reactive CN	P	9.8464	µg/l	2/21/2018 10:36:02	
50PPBCN	A	Reactive CN	P	51.1198	µg/l	2/21/2018 10:36:03	
100PPBCN	A	Reactive CN	P	98.8163	µg/l	2/21/2018 10:36:04	
250PPBCN	A	Reactive CN	P	251.2034	µg/l	2/21/2018 10:36:05	
500PPBCN	A	Reactive CN	P	499.5279	µg/l	2/21/2018 10:36:06	
ICV1	S	Reactive CN	P	96.2566	µg/l	2/21/2018 11:59:28	
ICB1	S	Reactive CN	P	-0.2739	µg/l	2/21/2018 11:59:29	
CCV1	S	Reactive CN	P	240.8648	µg/l	2/21/2018 12:32:04	
CCB1	S	Reactive CN	P	0.2381	µg/l	2/21/2018 12:32:05	
PB106720BL	S	Reactive CN	P	-0.0641	µg/l	2/21/2018 12:32:06	
J1391-02	S	Reactive CN	P	0.0238	µg/l	2/21/2018 12:32:07	
J1391-02DUP	S	Reactive CN	P	-0.0799	µg/l	2/21/2018 12:32:08	
J1403-06	S	Reactive CN	P	0.0804	µg/l	2/21/2018 12:32:11	
J1404-06	S	Reactive CN	P	0.4391	µg/l	2/21/2018 12:32:12	
J1622-02	S	Reactive CN	P	-0.1692	µg/l	2/21/2018 12:39:33	
J1623-04	S	Reactive CN	P	-0.2143	µg/l	2/21/2018 12:39:34	
J1623-08	S	Reactive CN	P	-0.0239	µg/l	2/21/2018 12:39:35	
CCV2	S	Reactive CN	P	241.5682	µg/l	2/21/2018 12:39:36	
CCB2	S	Reactive CN	P	0.0673	µg/l	2/21/2018 12:39:37	
J1623-12	S	Reactive CN	P	-0.4789	µg/l	2/21/2018 12:39:38	
J1623-16	S	Reactive CN	P	-0.3603	µg/l	2/21/2018 12:39:39	
J1623-20	S	Reactive CN	P	-0.3078	µg/l	2/21/2018 12:39:40	
J1624-02	S	Reactive CN	P	-0.3058	µg/l	2/21/2018 12:39:41	
J1624-04	S	Reactive CN	P	-0.217	µg/l	2/21/2018 12:39:42	
J1624-06	S	Reactive CN	P	-0.3187	µg/l	2/21/2018 12:39:43	
PB106721BL	S	Reactive CN	P	-0.115	µg/l	2/21/2018 12:44:20	
J1626-02	S	Reactive CN	P	-0.1332	µg/l	2/21/2018 12:44:21	
J1626-02DUP	S	Reactive CN	P	-0.0954	µg/l	2/21/2018 12:44:22	
CCV3	S	Reactive CN	P	258.0177	µg/l	2/21/2018 12:44:23	
CCB3	S	Reactive CN	P	0.4062	µg/l	2/21/2018 12:44:24	
J1391-04	S	Reactive CN	P	-0.217	µg/l	2/21/2018 13:04:03	
J1398-04	S	Reactive CN	P	-0.2669	µg/l	2/21/2018 13:04:04	
CCV4	S	Reactive CN	P	256.9152	µg/l	2/21/2018 13:04:05	
CCB4	S	Reactive CN	P	0.1086	µg/l	2/21/2018 13:04:06	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/21/2018 10:37


Test Total CN

Accepted 2/21/2018 10:37

Factor 1132
Bias 0.004

Coeff. of det. 0.999978

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.3338	0.0000	
2	5.0PPBCN	0.008	4.8199	5.0000	
3	10PPBCN	0.012	9.8464	10.0000	
4	50PPBCN	0.049	51.1198	50.0000	
5	100PPBCN	0.091	98.8163	100.0000	
6	250PPBCN	0.226	251.2034	250.0000	
7	500PPBCN	0.445	499.5279	500.0000	