

ISM02.4


LB93423

INSTRUMENT ID: CX1

Linear

 μ Abs.:

41202



A= 0.0000e+000

B= 2.4286e-004

C= -8.3376e-002

Rho= 0.9998807

Accept= Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	-0.023	-0.023	247	0.000	247					-2
0.2	0.200	0.197	-0.003	1153	0.0 %	1153					1
2.5	2.500	2.517	0.017	10708	0.0 %	10708					0
5.0	5.000	4.976	-0.024	20834	0.0 %	20834					1
7.5	7.500	7.610	0.110	31679	0.0 %	31679					1
10.0	10.000	9.923	-0.077	41202	0.0 %	41202					-1

MB
2/21/18

LB93423
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	247	-		0 ISM02.4	PPB	2/21/2018 14:21	Std
	0.2 S0.2	1153	-		0.2 ISM02.4	PPB	2/21/2018 14:24	Std
	2.5 S2.5	10708	-		2.5 ISM02.4	PPB	2/21/2018 14:26	Std
	5 S5	20834	-		5 ISM02.4	PPB	2/21/2018 14:28	Std
	7.5 S7.5	31679	-		7.5 ISM02.4	PPB	2/21/2018 14:31	Std
	10 S10	41202	-		10 ISM02.4	PPB	2/21/2018 14:33	Std
ICV	ICV	16269	3.8677 -		ISM02.4	PPB	2/21/2018 14:36	SMPL
ICB	ICB	130	-0.0518 -		ISM02.4	PPB	2/21/2018 14:38	SMPL
CCV090	CCV090	20667	4.9358 -		ISM02.4	PPB	2/21/2018 14:40	SMPL
CCB090	CCB090	118	-0.0547 -		ISM02.4	PPB	2/21/2018 14:42	SMPL
PB106697BL	PBW003	165	-0.0433 -		ISM02.4	PPB	2/21/2018 14:45	SMPL
J1571-01	ME62G2	-59	-0.0977 -		ISM02.4	PPB	2/21/2018 14:47	SMPL
J1571-02	ME62G4	-35	-0.0919 -		ISM02.4	PPB	2/21/2018 14:49	SMPL
J1571-03	ME62H8	-73	-0.1011 -		ISM02.4	PPB	2/21/2018 14:51	SMPL
J1571-04	ME62K3	-41	-0.0933 -		ISM02.4	PPB	2/21/2018 14:54	SMPL
J1571-05	ME62K8	-69	-0.1001 -		ISM02.4	PPB	2/21/2018 14:56	SMPL
J1571-06	ME62K8D	-70	-0.1004 -		ISM02.4	PPB	2/21/2018 14:58	SMPL
J1571-07	ME62K8S	4387	0.9821 -		ISM02.4	PPB	2/21/2018 15:01	SMPL
J1571-08	ME62L9	-109	-0.1098 -		ISM02.4	PPB	2/21/2018 15:03	SMPL
J1571-09	ME62M5	-86	-0.1043 -		ISM02.4	PPB	2/21/2018 15:05	SMPL
PB106698BL	PBW003	137	-0.0501 -		ISM02.4	PPB	2/21/2018 15:07	SMPL
J1572-01	ME62G3	-82	-0.1033 -		ISM02.4	PPB	2/21/2018 15:10	SMPL
J1572-02	ME62H4	-71	-0.1006 -		ISM02.4	PPB	2/21/2018 15:12	SMPL
J1572-03	ME62H9	-101	-0.1079 -		ISM02.4	PPB	2/21/2018 15:14	SMPL
J1572-04	ME62K4	-122	-0.113 -		ISM02.4	PPB	2/21/2018 15:17	SMPL
J1572-05	ME62K9	-43	-0.0938 -		ISM02.4	PPB	2/21/2018 15:19	SMPL
J1572-06	ME62K9D	-50	-0.0955 -		ISM02.4	PPB	2/21/2018 15:21	SMPL
J1572-07	ME62K9S	4449	0.9971 -		ISM02.4	PPB	2/21/2018 15:23	SMPL
J1572-08	ME62M6	-109	-0.1098 -		ISM02.4	PPB	2/21/2018 15:26	SMPL
J1572-09	ME62M7	-115	-0.1113 -		ISM02.4	PPB	2/21/2018 15:28	SMPL
CCV091	CCV091	20846	4.9793 -		ISM02.4	PPB	2/21/2018 15:30	SMPL
CCB091	CCB091	118	-0.0547 -		ISM02.4	PPB	2/21/2018 15:33	SMPL