



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93435

Temp.: 22.3 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/21/2018	15:45
2	CAL2	1	Water			7.00	02/21/2018	15:47
3	CAL3	1	Water			10.01	02/21/2018	15:49
4	ICV	1	Water			7.00	02/21/2018	15:51
5	CCV1	1	Water			2.00	02/21/2018	15:53
6	J1634-06	1	Water			6.84	02/21/2018	15:55
7	J1634-06DUP	1	Water			6.85	02/21/2018	15:57
8	J1637-03	1	Water			9.83	02/21/2018	15:59
9	CCV2	1	Water			12.00	02/21/2018	16:02

WORKLIST(Hardcopy Internal Chain)

1893435

Worklist Name : ph J1634

Worklist ID : 108802

Date : 2/21/2018 3:43:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/21/2018	Water	J1634-06-F	pH	Cool 4 deg C	PSEG01	P61	FRAC-TANK-257928	02/21/2018	9040C
02/21/2018	Water	J1637-03-F	pH	Cool 4 deg C	PSEG01	F11	FRAC-TANK-A124	02/21/2018	9040C

Date/Time 02-21-18 3:40 PM
 Received by: JP
 Relinquished by: CO

Date/Time 02-21-18 6:40 PM
 Received by: JP
 Relinquished by: CO