



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93437

Temp.: 22.9 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/21/2018	16:35
2	CAL2	1	Water			7.00	02/21/2018	16:36
3	CAL3	1	Water			10.02	02/21/2018	16:37
4	ICV	1	Water			7.01	02/21/2018	16:39
5	CCV1	1	Water			2.01	02/21/2018	16:40
6	J1396-02	1	Solid	20.02	20	7.56	02/21/2018	16:41
7	J1396-02DUP	1	Solid	20.03	20	7.57	02/21/2018	16:42
8	J1634-05	1	Solid	20.03	20	6.44	02/21/2018	16:44
9	J1635-02	1	Solid	20.02	20	6.65	02/21/2018	16:45
10	J1635-04	1	Solid	20.03	20	8.34	02/21/2018	16:46
11	CCV2	1	Water			12.02	02/21/2018	16:47

WORKLIST(Hardcopy Internal Chain)

1043437

WorkList Name : corrosivity j1635

WorkList ID : 108800

Date : 2/21/2018 3:42:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/21/2018	Solid	J1396-02	Corrosivity	Cool 4 deg C	PSEG05	P53	HD-01-022118	02/21/2018	9045C
02/21/2018	Solid	J1634-05	Corrosivity	Cool 4 deg C	PSEG01	P61	RO-72-11931	02/21/2018	9045C
02/21/2018	Solid	J1635-02	Corrosivity	Cool 4 deg C	PSEG01	P62	TP-1-3FT	02/21/2018	9045C
02/21/2018	Solid	J1635-04	Corrosivity	Cool 4 deg C	PSEG01	F11	TP-1-6FT	02/21/2018	9045C

Date/Time 02-21-18 3:42 PM
 Received by: JS
 Relinquished by: JS

Date/Time 02-21-18 6:00 PM
 Received by: JS
 Relinquished by: JS