

Daily Analysis Runlog For Sequence/QC Batch ID # LB93443

Review By		Review On				
SubDirectory		LB93443		Test : Nitrite		
STD. NAME		STD REF.#				
ICAL Standard		WP62183				
ICV Standard		WP62185				
CCV Standard		WP62184				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP62185				
Chk Standard		WP61258				
Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0061PPM	0.0061PPM	CAL1	02/22/18 10:41		OK
2	0.0122PPM	0.0122PPM	CAL2	02/22/18 10:41		OK
3	0.0243PPM	0.0243PPM	CAL3	02/22/18 10:41		OK
4	0.0608PPM	0.0608PPM	CAL4	02/22/18 10:41		OK
5	0.1216PPM	0.1216PPM	CAL5	02/22/18 10:41		OK
6	0.304PPM	0.304PPM	CAL6	02/22/18 10:41		OK
7	0.608PPM	0.608PPM	CAL7	02/22/18 10:41		OK
8	ICV1	ICV1	ICV	02/22/18 11:09		OK
9	ICB1	ICB1	ICB	02/22/18 11:09		OK
10	CCV1	CCV1	CCV	02/22/18 11:09		OK
11	CCB1	CCB1	CCB	02/22/18 11:10		OK
12	LB93443BLW	LB93443BLW	MB	02/22/18 11:10		OK
13	LB93443BSW	LB93443BSW	LCS	02/22/18 11:10		OK
14	J1161-07	LOD-MDL-WATER-0	SAM	02/22/18 11:10		OK
15	J1161-09	MDL-WATER-03-QT	SAM	02/22/18 11:10		OK
16	CCV2	CCV2	CCV	02/22/18 11:10		OK
17	CCB2	CCB2	CCB	02/22/18 11:10		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB93443	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	LOD-MDL-WATER-01-QT1-2018	SDG No.:	LB93443
Lab Sample ID:	J1161-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	Nitrite	0.0026	U	1	0.003		0.006	mg/L		02/22/2018	SM4500-NO2 E



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB93443	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	MDL-WATER-03-QT1-2018	SDG No.:	LB93443
Lab Sample ID:	J1161-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	Nitrite	0.0024	U	1	0.003		0.006	mg/L		02/22/2018	SM4500-NO2 E



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	1/16/2018 12:00:00 AM
Project:	LB93443	Date Received:	1/16/2018 12:00:00 AM
Client Sample ID:	MDL-WATER-03-QT1-2018	SDG No.:	LB93443
Lab Sample ID:	J1161-09	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB93443

SDG No.: LB93443

Contract: _____ Lab Code: CHEM

Case No.: LB93443

SAS No.: LB93443

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	Qual	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Nitrite	0.3	0.3	100		90 - 110		02/22/2018	11:09	LB93443
CCV2	Nitrite	0.31	0.3	103.3		90 - 110		02/22/2018	11:10	LB93443



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93443

SDG No.: LB93443

Contract: _____

Lab Code: CHEM

Case No.: LB93443

SAS No.: LB93443

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Nitrite	-0.0014	+/-0.006	U		0.006		02/22/2018	11:09	LB93443



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>LB93443</u>		SDG No.: <u>LB93443</u>								
Contract: _____	Lab Code: <u>CHEM</u>		Case No.: <u>LB93443</u>			SAS No.: <u>LB93443</u>				
Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Nitrite	-0.001	+/-0.006	U		0.006		02/22/2018	11:10	LB93443
CCB2	Nitrite	-0.0009	+/-0.006	U		0.006		02/22/2018	11:10	LB93443



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB93443

SDG No.: LB93443

Contract: Lab Code: CHEM

Case No.: LB93443

SAS No.: LB93443

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB93443

SDG No.: LB93443

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
LB93443BLW	Nitrite	WATER -0.0004	<0.006	Batch Number: U		0.006		Prep Date: 02/22/2018	11:10	LB93443



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB93443

SDG No.: LB93443

Contract:

Lab Code: CHEM

Case No.: LB93443

SAS No.: LB93443

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB93443BSW Nitrite	mg/L	0.3	0.31		103.3	90 - 110	