


7470A

LB93445
INSTRUMENT ID: CV2

Linear

 μ Abs.:

38010



A= 0.0000e+000

B= 2.6389e-004

C= -1.8500e-002

Rho= 0.9999148

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	196	0.000	196				
0.2	0.200	0.234	0.034	958	0.0 %	958				
2.5	2.500	2.448	-0.052	9348	0.0 %	9348				
5.0	5.000	4.921	-0.079	18718	0.0 %	18718				
7.5	7.500	7.551	0.051	28684	0.0 %	28684				
10.0	10.000	10.012	0.012	38010	0.0 %	38010				

LB93445 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Feb 2018 11:27:54

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	196	-	0.000	7470A	PPB	22 Feb 2018 11:36:10	S	Std
0.2 - 1	50.2	958	-	0.200	7470A	PPB	22 Feb 2018 11:38:18	S	Std
2.5 - 1	52.5	9348	-	2.500	7470A	PPB	22 Feb 2018 11:40:27	S	Std
5.0 - 1	55	18718	-	5.000	7470A	PPB	22 Feb 2018 11:42:34	S	Std
7.5 - 1	57.5	28684	-	7.500	7470A	PPB	22 Feb 2018 11:44:43	S	Std
10.0 - 1	59	38010	-	10.000	7470A	PPB	22 Feb 2018 11:46:51	S	Std
ICV12 - 1	ICV12	15265	4.0098	-	7470A	PPB	22 Feb 2018 11:49:27	U	SMPL
ICB12 - 1	ICB12	84	0.0037	-	7470A	PPB	22 Feb 2018 11:51:34	U	SMPL
CCV95 - 1	CCV95	18563	4.8801	-	7470A	PPB	22 Feb 2018 11:53:41	U	SMPL
CCB95 - 1	CCB95	98	0.0074	-	7470A	PPB	22 Feb 2018 11:55:47	U	SMPL
CRA - 1	CRA	980	0.2401	-	7470A	PPB	22 Feb 2018 11:57:54	U	SMPL
HighStd - 1	HighStd	37911	9.9859	-	7470A	PPB	22 Feb 2018 12:00:01	U	SMPL
ChkStd - 1	ChkStd	25898	6.8158	-	7470A	PPB	22 Feb 2018 12:02:08	U	SMPL
PB106732BL - 1	PBW	-114	-0.0486	-	7470A	PPB	22 Feb 2018 12:04:16	U	SMPL
PB106732BS - 1	LCSW	14185	3.7248	-	7470A	PPB	22 Feb 2018 12:06:25	U	SMPL
PB106732TB - 1	PB106703TB	-28	-0.0259	-	7470A	PPB	22 Feb 2018 12:08:33	U	SMPL
J1391-02 - 1	AU-05-022018-A	63	-0.0019	-	7470A	PPB	22 Feb 2018 12:10:41	U	SMPL
J1391-04 - 1	AU-05-022018-B	38	-0.0085	-	7470A	PPB	22 Feb 2018 12:12:49	U	SMPL
J1398-04 - 1	OK-02-022018	26	-0.0116	-	7470A	PPB	22 Feb 2018 12:14:58	U	SMPL
J1398-04DUP - 1	OK-02-022018DUP	120	0.0132	-	7470A	PPB	22 Feb 2018 12:17:05	U	SMPL
CCV96 - 1	CCV96	18256	4.7991	-	7470A	PPB	22 Feb 2018 12:19:12	U	SMPL
CCB96 - 1	CCB96	107	0.0097	-	7470A	PPB	22 Feb 2018 12:21:19	U	SMPL
J1398-04MS - 1	OK-02-022018MS	14134	3.7114	-	7470A	PPB	22 Feb 2018 12:23:26	U	SMPL
J1398-04MSD - 1	OK-02-022018MSD	13744	3.6084	-	7470A	PPB	22 Feb 2018 12:25:33	U	SMPL
J1403-06 - 1	SU-04-022018	69	-0.0003	-	7470A	PPB	22 Feb 2018 12:27:40	U	SMPL
J1404-06 - 1	TR-05-022018	163	0.0245	-	7470A	PPB	22 Feb 2018 12:29:47	U	SMPL
J1622-02 - 1	MAD-SUB-TRENCH	164	0.0248	-	7470A	PPB	22 Feb 2018 12:31:54	U	SMPL
J1623-04 - 1	TP-1-2	66	-0.0011	-	7470A	PPB	22 Feb 2018 12:34:02	U	SMPL
J1623-08 - 1	TP-3-4	66	-0.0011	-	7470A	PPB	22 Feb 2018 12:36:09	U	SMPL
J1623-12 - 1	TP-5-6	75	0.0013	-	7470A	PPB	22 Feb 2018 12:38:17	U	SMPL
J1623-16 - 1	TP-7-8	92	0.0058	-	7470A	PPB	22 Feb 2018 12:40:26	U	SMPL
J1623-20 - 1	TP-9-10	122	0.0137	-	7470A	PPB	22 Feb 2018 12:42:33	U	SMPL
CCV97 - 1	CCV97	18134	4.7669	-	7470A	PPB	22 Feb 2018 12:44:40	U	SMPL
CCB97 - 1	CCB97	106	0.0095	-	7470A	PPB	22 Feb 2018 12:46:46	U	SMPL
J1624-02 - 1	TP-1	108	0.0100	-	7470A	PPB	22 Feb 2018 12:48:54	U	SMPL
J1624-04 - 1	TP-2	123	0.0140	-	7470A	PPB	22 Feb 2018 12:51:01	U	SMPL
J1624-06 - 1	TP-3	44	-0.0069	-	7470A	PPB	22 Feb 2018 12:53:08	U	SMPL
PB106733BL - 1	PBW	177	0.0282	-	7470A	PPB	22 Feb 2018 12:55:15	U	SMPL
PB106733BS - 1	LCSW	14053	3.6900	-	7470A	PPB	22 Feb 2018 12:57:23	U	SMPL
PB106733TB - 1	PB106733TB	106	0.0095	-	7470A	PPB	22 Feb 2018 12:59:31	U	SMPL
J1626-02 - 1	SILVERLAKE-LAYDOWN-YARD	207	0.0361	-	7470A	PPB	22 Feb 2018 13:01:39	U	SMPL
J1626-02DUP - 1	SILVERLAKE-LAYDOWN-YARDDUP	140	0.0184	-	7470A	PPB	22 Feb 2018 13:03:47	U	SMPL
J1626-02MS - 1	SILVERLAKE-LAYDOWN-YARDMS	14225	3.7354	-	7470A	PPB	22 Feb 2018 13:05:56	U	SMPL
J1626-02MSD - 1	SILVERLAKE-LAYDOWN-YARDMSD	13167	3.4562	-	7470A	PPB	22 Feb 2018 13:08:03	U	SMPL
CCV98 - 1	CCV98	18553	4.8775	-	7470A	PPB	22 Feb 2018 13:10:10	U	SMPL
CCB98 - 1	CCB98	86	0.0042	-	7470A	PPB	22 Feb 2018 13:12:17	U	SMPL
PB106734BL - 1	PBW	64	-0.0016	-	7470A	PPB	22 Feb 2018 13:14:24	U	SMPL
PB106734BS - 1	LCSW	12573	3.2994	-	7470A	PPB	22 Feb 2018 13:16:31	U	SMPL
J1626-01 - 1	SILVERLAKE-LAYDOWN-YARD	-174	-0.0644	-	7470A	PPB	22 Feb 2018 13:18:39	U	SMPL
J1626-01DUP - 1	SILVERLAKE-LAYDOWN-YARDDUP	-81	-0.0399	-	7470A	PPB	22 Feb 2018 13:20:45	U	SMPL
J1626-01MS - 1	SILVERLAKE-LAYDOWN-YARDMS	12469	3.2720	-	7470A	PPB	22 Feb 2018 13:22:53	U	SMPL
J1626-01MSD - 1	SILVERLAKE-LAYDOWN-YARDMSD	12145	3.1865	-	7470A	PPB	22 Feb 2018 13:25:00	U	SMPL
CCV99 - 1	CCV99	18277	4.8047	-	7470A	PPB	22 Feb 2018 13:27:08	U	SMPL
CCB99 - 1	CCB99	70	-0.0000	-	7470A	PPB	22 Feb 2018 13:29:17	U	SMPL
J1398-04LX5 - 1		164	0.0248	-	7470A	PPB	22 Feb 2018 13:31:25	U	SMPL
J1398-04A - 1		15133	3.9750	-	7470A	PPB	22 Feb 2018 13:33:32	U	SMPL
J1626-02LX5 - 1		108	0.0100	-	7470A	PPB	22 Feb 2018 13:35:39	U	SMPL
J1626-02A - 1		15310	4.0217	-	7470A	PPB	22 Feb 2018 13:37:46	U	SMPL
J1626-01LX5 - 1		50	-0.0053	-	7470A	PPB	22 Feb 2018 13:39:54	U	SMPL
J1626-01A - 1		15078	3.9605	-	7470A	PPB	22 Feb 2018 13:42:00	U	SMPL
CCV100 - 1	CCV100	18552	4.8772	-	7470A	PPB	22 Feb 2018 13:44:08	U	SMPL
CCB100 - 1	CCB100	-286	-0.0940	-	7470A	PPB	22 Feb 2018 13:46:15	U	SMPL