

7471B


LB 93458

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42899



A= 0.0000e+000

B= 2.3372e-004

C= -4.7896e-002

Rho= 0.9999547

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.047	-0.047	4	0.000	4				
0.2	0.200	0.202	0.002	1070	0.0 %	1070				
2.5	2.500	2.565	0.065	11181	0.0 %	11181				
5.0	5.000	4.985	-0.015	21535	0.0 %	21535				
7.5	7.500	7.516	0.016	32361	0.0 %	32361				
10	10.000	9.979	-0.021	42899	0.0 %	42899				

LB93458 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Feb 2018 15:31:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000	7471B	PPB	22 Feb 2018 15:33:53	S	Std
0.2 - 1	50.2	1070	-	0.2000	7471B	PPB	22 Feb 2018 15:36:10	S	Std
2.5 - 1	52.5	11181	-	2.5000	7471B	PPB	22 Feb 2018 15:38:28	S	Std
5.0 - 1	55	21535	-	5.0000	7471B	PPB	22 Feb 2018 15:40:45	S	Std
7.5 - 1	57.5	32361	-	7.5000	7471B	PPB	22 Feb 2018 15:43:03	S	Std
10 - 1	510	42899	-	10.0000	7471B	PPB	22 Feb 2018 15:45:19	S	Std
ICV13 - 1	ICV13	17344	4.0058	-	7471B	PPB	22 Feb 2018 15:48:53	U	SMPL
ICB13 - 1	ICB13	-127	-0.0776	-	7471B	PPB	22 Feb 2018 15:51:08	U	SMPL
CCV02 - 1	CCV02	21467	4.9694	-	7471B	PPB	22 Feb 2018 15:53:24	U	SMPL
CCB02 - 1	CCB02	-132	-0.0787	-	7471B	PPB	22 Feb 2018 15:55:40	U	SMPL
CRA - 1	CRA	1068	0.2017	-	7471B	PPB	22 Feb 2018 15:57:56	U	SMPL
HighStd - 1	HighStd	42812	9.9582	-	7471B	PPB	22 Feb 2018 16:00:12	U	SMPL
ChkStd - 1	ChkStd	29510	6.8492	-	7471B	PPB	22 Feb 2018 16:02:28	U	SMPL
PB106743BL - 1	PBS	10	-0.0456	-	7471B	PPB	22 Feb 2018 16:04:46	U	SMPL
PB106743BS - 1	LCSS	16287	3.7587	-	7471B	PPB	22 Feb 2018 16:07:04	U	SMPL
J1396-01 - 1	HD-01-022118	1744	0.3597	-	7471B	PPB	22 Feb 2018 16:09:21	U	SMPL
J1634-01 - 1	RO-72-11931	1022	0.1910	-	7471B	PPB	22 Feb 2018 16:11:38	U	SMPL
J1635-01 - 1	TP-1-3FT	691	0.1136	-	7471B	PPB	22 Feb 2018 16:13:55	U	SMPL
J1635-03 - 1	TP-1-6FT	431	0.0528	-	7471B	PPB	22 Feb 2018 16:16:13	U	SMPL
J1635-03DUP - 1	TP-1-6FTDUP	474	0.0629	-	7471B	PPB	22 Feb 2018 16:18:29	U	SMPL
CCV03 - 1	CCV03	21527	4.9834	-	7471B	PPB	22 Feb 2018 16:20:45	U	SMPL
CCB03 - 1	CCB03	-152	-0.0834	-	7471B	PPB	22 Feb 2018 16:23:00	U	SMPL
J1635-03MS - 1	TP-1-6FTMS	16699	3.8550	-	7471B	PPB	22 Feb 2018 16:25:16	U	SMPL
J1635-03MSD - 1	TP-1-6FTMSD	16358	3.7753	-	7471B	PPB	22 Feb 2018 16:27:33	U	SMPL
J1638-01 - 1	3-4-DENSE-GRADE-AGG-RCA	10858	2.4899	-	7471B	PPB	22 Feb 2018 16:29:49	U	SMPL
J1639-01 - 1	1-1-2-DENSE-GRADE-AGG-RCA	2148	0.4541	-	7471B	PPB	22 Feb 2018 16:32:05	U	SMPL
J1640-01 - 1	2-2-DENSE-GRADE-AGG-RCA	426	0.0517	-	7471B	PPB	22 Feb 2018 16:34:22	U	SMPL
J1635-03LX5 - 1		25	-0.0421	-	7471B	PPB	22 Feb 2018 16:36:39	U	SMPL
J1635-03A - 1		17058	3.9389	-	7471B	PPB	22 Feb 2018 16:38:56	U	SMPL
CCV04 - 1	CCV04	21480	4.9725	-	7471B	PPB	22 Feb 2018 16:41:14	U	SMPL
CCB04 - 1	CCB04	-152	-0.0834	-	7471B	PPB	22 Feb 2018 16:43:32	U	SMPL