



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB93464

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	02/22/2018	17:00
2	CAL2	1	Water			7.01	02/22/2018	17:01
3	CAL3	1	Water			10.03	02/22/2018	17:02
4	ICV	1	Water			7.00	02/22/2018	17:05
5	CCV1	1	Water			2.01	02/22/2018	17:07
6	J1398-06	1	Solid	20.02	20	7.21	02/22/2018	17:09
7	J1398-06DUP	1	Solid	20.03	20	7.22	02/22/2018	17:10
8	J1648-02	1	Solid	20.05	20	9.04	02/22/2018	17:11
9	J1652-03	1	Solid	20.04	20	7.94	02/22/2018	17:13
10	J1652-06	1	Solid	20.03	20	7.91	02/22/2018	17:15
11	J1652-09	1	Solid	20.00	20	7.82	02/22/2018	17:16
12	J1652-12	1	Solid	20.02	20	8.17	02/22/2018	17:19
13	J1660-05	1	Solid	20.03	20	7.54	02/22/2018	17:21
14	J1660-10	1	Solid	20.05	20	7.88	02/22/2018	17:25
15	J1660-15	1	Solid	20.01	20	8.21	02/22/2018	17:29
16	CCV2	1	Water			12.01	02/22/2018	17:30
17	J1660-20	1	Solid	20.03	20	7.01	02/22/2018	17:31
18	CCV3	1	Water			2.01	02/22/2018	17:32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1660

WorkList ID : 108847

Date : 2/22/2018 4:36:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2018	Solid	J1398-06	Corrosivity	Cool 4 deg C	PSEG05	F12	OK-01-022218	02/22/2018	9045C
02/22/2018	Solid	J1648-02	Corrosivity	Cool 4 deg C	PSEG01	F21	RO-124031-RO124035	02/22/2018	9045C
02/22/2018	Solid	J1652-03	Corrosivity	Cool 4 deg C	PSEG01	F21	RT-4651	02/22/2018	9045C
02/22/2018	Solid	J1652-06	Corrosivity	Cool 4 deg C	PSEG01	F21	RT-2898	02/22/2018	9045C
02/22/2018	Solid	J1652-09	Corrosivity	Cool 4 deg C	PSEG01	F21	RT-4248	02/22/2018	9045C
02/22/2018	Solid	J1652-12	Corrosivity	Cool 4 deg C	PSEG01	F21	RT-3026	02/22/2018	9045C
02/22/2018	Solid	J1660-05	Corrosivity	Cool 4 deg C	PSEG01	F22	RT-4632	02/22/2018	9045C
02/22/2018	Solid	J1660-10	Corrosivity	Cool 4 deg C	PSEG01	F22	RT-1488	02/22/2018	9045C
02/22/2018	Solid	J1660-15	Corrosivity	Cool 4 deg C	PSEG01	F22	RT-3025	02/22/2018	9045C
02/22/2018	Solid	J1660-20	Corrosivity	Cool 4 deg C	PSEG01	F22	RT-2342	02/22/2018	9045C

Date/Time 02-22-18 4:37 PM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 02-22-18 6:30 PM
 Received by: [Signature]
 Relinquished by: [Signature]