



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93472

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 99.0

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/22/2018	16:40
2	CAL2	1	Water			7.00	02/22/2018	16:42
3	CAL3	1	Water			10.01	02/22/2018	16:44
4	ICV	1	Water			7.00	02/22/2018	16:46
5	CCV1	1	Water			2.00	02/22/2018	16:48
6	J1654-01	1	Solid	20.02	20	6.20	02/22/2018	16:50
7	J1654-01DUP	1	Solid	20.01	20	6.21	02/22/2018	16:52
8	CCV2	1	Water			12.00	02/22/2018	16:54

WORKLIST(Hardcopy Internal Chain)

LB 93472

Worklist Name : pHJ1654

Worklist ID : 108844

Date : 2/22/2018 4:30:20 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2018	Solid	J1654-01	pH	Cool 4 deg C	TULL02	D41	REID-TOPSOIL	02/22/2018	9045C

Date/Time

02/22/18 4:35 PM

Received by:

SP

Relinquished by:

CP

Date/Time

02/22/18 5:10 PM

Received by:

CP

Relinquished by:

SP