

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB93477

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP61786
calibration std. phosphate 1 ppm	WP62226
calibration std. phosphate 0.5 ppm	WP62225
calibration std. phosphate 0.3 ppm	WP62224
calibration std. 0 ppm	WP62221
phosphate ICV std.	WP62228
phosphate CCV std.	WP62227
calibration std. phosphate 0.1 ppm	WP62223
calibration std. phosphate 0.05 ppm	WP62222
Combined reagent	WP62231
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0007Slope: 0.6498Regression: 0.999842

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	02/23/2018	11:38
2	CAL2	0.05	1	50	50	0.033	02/23/2018	11:38
3	CAL3	0.10	1	50	50	0.066	02/23/2018	11:39
4	CAL4	0.30	1	50	50	0.186	02/23/2018	11:39
5	CAL5	0.50	1	50	50	0.329	02/23/2018	11:40
6	CAL6	1.00	1	50	50	0.649	02/23/2018	11:40



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB93477

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.494	02/23/2018	11:41
2	ICB		1	50	50	0.000	0.001	02/23/2018	11:41
3	CCV1	0.50	1	50	50	0.326	0.503	02/23/2018	11:42
4	CCB1		1	50	50	0.000	0.001	02/23/2018	11:42
5	PB106798BL		1	50	50	0.000	0.001	02/23/2018	11:43
6	PB106798BS	0.50	1	50	50	0.307	0.474	02/23/2018	11:43
7	J1650-01		1	50	50	1.663	2.560	02/23/2018	11:44
8	J1650-01DUP		1	50	50	1.689	2.600	02/23/2018	11:44
9	J1650-01MS	0.50	1	50	50	1.747	2.690	02/23/2018	11:45
10	J1650-01MSD	0.50	1	50	50	1.711	2.634	02/23/2018	11:45
11	J1650-01		20	50	50	0.512	0.789	02/23/2018	13:01
12	J1650-01DUP		20	50	50	0.537	0.827	02/23/2018	13:01
13	J1650-01MS	0.50	20	50	50	0.521	0.803	02/23/2018	13:02
14	J1650-01MSD	0.50	20	50	50	0.523	0.806	02/23/2018	13:02
15	CCV2	0.50	1	50	50	0.324	0.500	02/23/2018	13:03
16	CCB2		1	50	50	0.000	0.001	02/23/2018	13:03