

LB 93478


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

36736



A= 0.0000e+000

B= 2.7562e-004

C= -1.9142e-002

Rho= 0.9997732

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	116	0.000	116				
0.2	0.200	0.199	-0.001	791	0.0 %	791				
2.5	2.500	2.571	0.071	9397	0.0 %	9397				
5.0	5.000	4.944	-0.056	18007	0.0 %	18007				
7.5	7.500	7.368	-0.132	26801	0.0 %	26801				
10.0	10.000	10.106	0.106	36736	0.0 %	36736				

LB93478 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Feb 2018 11:06:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	116	-	0.0000	7470A	PPB	23 Feb 2018 11:12:02	S	Std
0.2 - 1	50.2	791	-	0.2000	7470A	PPB	23 Feb 2018 11:14:10	S	Std
2.5 - 1	52.5	9397	-	2.5000	7470A	PPB	23 Feb 2018 11:16:18	S	Std
5.0 - 1	55	18007	-	5.0000	7470A	PPB	23 Feb 2018 11:18:26	S	Std
7.5 - 1	57.5	26801	-	7.5000	7470A	PPB	23 Feb 2018 11:20:34	S	Std
10.0 - 1	510	36736	-	10.0000	7470A	PPB	23 Feb 2018 11:22:42	S	Std
ICV14 - 1	ICV14	14319	3.9274	-	7470A	PPB	23 Feb 2018 11:25:37	U	SMPL
ICB14 - 1	ICB14	-84	-0.0423	-	7470A	PPB	23 Feb 2018 11:27:43	U	SMPL
CCV05 - 1	CCV05	18643	5.1192	-	7470A	PPB	23 Feb 2018 11:29:50	U	SMPL
CCB05 - 1	CCB05	-86	-0.0428	-	7470A	PPB	23 Feb 2018 11:31:57	U	SMPL
CRA - 1	CRA	777	0.1950	-	7470A	PPB	23 Feb 2018 11:34:04	U	SMPL
HighStd - 1	HighStd	37208	10.2360	-	7470A	PPB	23 Feb 2018 11:36:11	U	SMPL
ChkStd - 1	ChkStd	25617	7.0413	-	7470A	PPB	23 Feb 2018 11:38:18	U	SMPL
PB106768BL - 1	PBW	-93	-0.0448	-	7470A	PPB	23 Feb 2018 11:40:26	U	SMPL
PB106768BS - 1	LCSW	13131	3.6000	-	7470A	PPB	23 Feb 2018 11:42:36	U	SMPL
PB106768TB - 1	PB106739TB	-1	-0.0194	-	7470A	PPB	23 Feb 2018 11:44:43	U	SMPL
J1396-02 - 1	HD-01-022118	623	0.1526	-	7470A	PPB	23 Feb 2018 11:46:52	U	SMPL
J1634-05 - 1	RO-72-11931	106	0.0101	-	7470A	PPB	23 Feb 2018 11:49:00	U	SMPL
J1635-02 - 1	TP-1-3FT	77	0.0021	-	7470A	PPB	23 Feb 2018 11:51:08	U	SMPL
J1635-02DUP - 1	TP-1-3FTDUP	73	0.0010	-	7470A	PPB	23 Feb 2018 11:53:15	U	SMPL
CCV06 - 1	CCV06	18922	5.1961	-	7470A	PPB	23 Feb 2018 11:55:21	U	SMPL
CCB06 - 1	CCB06	-77	-0.0404	-	7470A	PPB	23 Feb 2018 11:57:28	U	SMPL
J1635-02MS - 1	TP-1-3FTMS	13451	3.6882	-	7470A	PPB	23 Feb 2018 11:59:36	U	SMPL
J1635-02MSD - 1	TP-1-3FTMSD	12851	3.5228	-	7470A	PPB	23 Feb 2018 12:01:42	U	SMPL
J1635-04 - 1	TP-1-6FT	-13	-0.0227	-	7470A	PPB	23 Feb 2018 12:03:48	U	SMPL
J1637-01 - 1	RBR200034-RBR20035-RT3425	90	0.0057	-	7470A	PPB	23 Feb 2018 12:05:56	U	SMPL
J1637-02 - 1	RT4259-RT3090-RBR200030-RBR4225	164	0.0261	-	7470A	PPB	23 Feb 2018 12:08:03	U	SMPL
PB106769BL - 1	PBW	72	0.0007	-	7470A	PPB	23 Feb 2018 12:10:10	U	SMPL
PB106769BS - 1	LCSW	14087	3.8635	-	7470A	PPB	23 Feb 2018 12:12:18	U	SMPL
PB106769TB - 1	PB106741TB	20	-0.0136	-	7470A	PPB	23 Feb 2018 12:14:26	U	SMPL
J1634-07 - 1	FRAC-TANK-257928	74	0.0013	-	7470A	PPB	23 Feb 2018 12:16:35	U	SMPL
J1634-07DUP - 1	FRAC-TANK-257928DUP	114	0.0123	-	7470A	PPB	23 Feb 2018 12:18:42	U	SMPL
CCV07 - 1	CCV07	18724	5.1415	-	7470A	PPB	23 Feb 2018 12:20:49	U	SMPL
CCB07 - 1	CCB07	-67	-0.0376	-	7470A	PPB	23 Feb 2018 12:22:56	U	SMPL
J1634-07MS - 1	FRAC-TANK-257928MS	13544	3.7138	-	7470A	PPB	23 Feb 2018 12:25:05	U	SMPL
J1634-07MSD - 1	FRAC-TANK-257928MSD	14030	3.8478	-	7470A	PPB	23 Feb 2018 12:27:11	U	SMPL
J1637-03 - 1	FRAC-TANK-A124	0	-0.0191	-	7470A	PPB	23 Feb 2018 12:29:18	U	SMPL
PB106770BL - 1	PBW	36	-0.0092	-	7470A	PPB	23 Feb 2018 12:31:26	U	SMPL
PB106770BS - 1	LCSW	13387	3.6705	-	7470A	PPB	23 Feb 2018 12:33:33	U	SMPL
J1634-06 - 1	FRAC-TANK-257928	-219	-0.0795	-	7470A	PPB	23 Feb 2018 12:35:41	U	SMPL
J1634-06DUP - 1	FRAC-TANK-257928DUP	-127	-0.0541	-	7470A	PPB	23 Feb 2018 12:37:49	U	SMPL
J1634-06MS - 1	FRAC-TANK-257928MS	14399	3.9495	-	7470A	PPB	23 Feb 2018 12:39:57	U	SMPL
J1634-06MSD - 1	FRAC-TANK-257928MSD	13479	3.6959	-	7470A	PPB	23 Feb 2018 12:42:05	U	SMPL
CCV08 - 1	CCV08	17485	4.8000	-	7470A	PPB	23 Feb 2018 12:44:12	U	SMPL
CCB08 - 1	CCB08	-15	-0.0233	-	7470A	PPB	23 Feb 2018 12:46:19	U	SMPL
J1635-02LX5 - 1		77	0.0021	-	7470A	PPB	23 Feb 2018 12:48:28	U	SMPL
J1635-02A - 1		13755	3.7720	-	7470A	PPB	23 Feb 2018 12:50:34	U	SMPL
J1634-07LX5 - 1		-28	-0.0269	-	7470A	PPB	23 Feb 2018 12:52:41	U	SMPL
J1634-07A - 1		14001	3.8398	-	7470A	PPB	23 Feb 2018 12:54:49	U	SMPL
J1634-06LX5 - 1		-55	-0.0343	-	7470A	PPB	23 Feb 2018 12:56:56	U	SMPL
J1634-06A - 1		14022	3.8456	-	7470A	PPB	23 Feb 2018 12:59:03	U	SMPL
CCV09 - 1	CCV09	18014	4.9458	-	7470A	PPB	23 Feb 2018 13:01:11	U	SMPL
CCB09 - 1	CCB09	-39	-0.0299	-	7470A	PPB	23 Feb 2018 13:03:19	U	SMPL