


7471B

LB 93489
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45168



A= 0.0000e+000

B= 2.2096e-004

C= -5.4556e-002

Rho= 0.9999051

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.037	-0.037	79	0.000	79				
0.2	0.200	0.198	-0.002	1142	0.0 %	1142				
2.5	2.500	2.525	0.025	11675	0.0 %	11675				
5.0	5.000	4.999	-0.001	22873	0.0 %	22873				
7.5	7.500	7.589	0.089	34594	0.0 %	34594				
10	10.000	9.926	-0.074	45168	0.0 %	45168				

LB93489 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Feb 2018 15:13:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	79	-	0.00007471B	PPB		23 Feb 2018 15:16:16	S	Std
0.2 - 1	50.2	1142	-	0.20007471B	PPB		23 Feb 2018 15:18:34	S	Std
2.5 - 1	52.5	11675	-	2.50007471B	PPB		23 Feb 2018 15:20:52	S	Std
5.0 - 1	55	22873	-	5.00007471B	PPB		23 Feb 2018 15:23:09	S	Std
7.5 - 1	57.5	34594	-	7.50007471B	PPB		23 Feb 2018 15:25:26	S	Std
10 - 1	510	45168	-	10.00007471B	PPB		23 Feb 2018 15:27:42	S	Std
ICV15 - 1	ICV15	17439	3.7987	-7471B	PPB		23 Feb 2018 15:30:40	U	SMPL
ICB15 - 1	ICB15	-48	-0.0652	-7471B	PPB		23 Feb 2018 15:32:55	U	SMPL
CCV10 - 1	CCV10	23051	5.0387	-7471B	PPB		23 Feb 2018 15:35:11	U	SMPL
CCB10 - 1	CCB10	-49	-0.0654	-7471B	PPB		23 Feb 2018 15:37:27	U	SMPL
CRA - 1	CRA	1127	0.1945	-7471B	PPB		23 Feb 2018 15:39:42	U	SMPL
HighStd - 1	HighStd	45189	9.9303	-7471B	PPB		23 Feb 2018 15:41:58	U	SMPL
ChkStd - 1	ChkStd	31916	6.9975	-7471B	PPB		23 Feb 2018 15:44:14	U	SMPL
PB106776BL - 1	PBS	-101	-0.0769	-7471B	PPB		23 Feb 2018 15:46:32	U	SMPL
PB106776BS - 1	LCSS	15668	3.4074	-7471B	PPB		23 Feb 2018 15:48:50	U	SMPL
J1398-05 - 1	OK-01-022218	1476	0.2716	-7471B	PPB		23 Feb 2018 15:51:07	U	SMPL
J1648-01 - 1	RO-124031-RO124035	11503	2.4871	-7471B	PPB		23 Feb 2018 15:53:24	U	SMPL
J1649-01 - 1	CB-6-LINE-WW@4.5	9962	2.1466	-7471B	PPB		23 Feb 2018 15:55:41	U	SMPL
J1649-02 - 1	CB-6-LINE-EW@4.5	4395	0.9165	-7471B	PPB		23 Feb 2018 15:57:59	U	SMPL
J1649-03 - 1	OIL-SEP-EXC-NW@2.5	2888	0.5836	-7471B	PPB		23 Feb 2018 16:00:15	U	SMPL
CCV11 - 1	CCV11	23009	5.0294	-7471B	PPB		23 Feb 2018 16:02:32	U	SMPL
CCB11 - 1	CCB11	-47	-0.0649	-7471B	PPB		23 Feb 2018 16:04:48	U	SMPL
J1649-04 - 1	OIL-SEP-EXC-SW@2.5	2015	0.3907	-7471B	PPB		23 Feb 2018 16:07:04	U	SMPL
J1652-01 - 1	RT-4651	3985	0.8260	-7471B	PPB		23 Feb 2018 16:09:20	U	SMPL
J1652-04 - 1	RT-2898	4301	0.8958	-7471B	PPB		23 Feb 2018 16:11:36	U	SMPL
J1652-07 - 1	RT-4248	7319	1.5626	-7471B	PPB		23 Feb 2018 16:13:53	U	SMPL
J1652-07DUP - 1	RT-4248DUP	7107	1.5158	-7471B	PPB		23 Feb 2018 16:16:09	U	SMPL
J1652-07MS - 1	RT-4248MS	23474	5.1322	-7471B	PPB		23 Feb 2018 16:18:26	U	SMPL
J1652-07MSD - 1	RT-4248MSD	23368	5.1088	-7471B	PPB		23 Feb 2018 16:20:43	U	SMPL
J1652-10 - 1	RT-3026	7878	1.6861	-7471B	PPB		23 Feb 2018 16:23:01	U	SMPL
J1654-01 - 1	REID-TOPSOIL	3350	0.6856	-7471B	PPB		23 Feb 2018 16:25:20	U	SMPL
J1660-01 - 1	RT-4632	1911	0.3677	-7471B	PPB		23 Feb 2018 16:27:36	U	SMPL
CCV12 - 1	CCV12	23217	5.0754	-7471B	PPB		23 Feb 2018 16:29:52	U	SMPL
CCB12 - 1	CCB12	-37	-0.0627	-7471B	PPB		23 Feb 2018 16:32:08	U	SMPL
J1660-06 - 1	RT-1488	1780	0.3387	-7471B	PPB		23 Feb 2018 16:34:25	U	SMPL
J1660-11 - 1	RT-3025	5063	1.0641	-7471B	PPB		23 Feb 2018 16:36:41	U	SMPL
J1660-16 - 1	RT-2342	1622	0.3038	-7471B	PPB		23 Feb 2018 16:38:57	U	SMPL
J1652-07LX5 - 1		1453	0.2665	-7471B	PPB		23 Feb 2018 16:41:14	U	SMPL
J1652-07A - 1		24151	5.2818	-7471B	PPB		23 Feb 2018 16:43:30	U	SMPL
CCV13 - 1	CCV13	23024	5.0327	-7471B	PPB		23 Feb 2018 16:45:47	U	SMPL
CCB13 - 1	CCB13	-19	-0.0588	-7471B	PPB		23 Feb 2018 16:48:04	U	SMPL