

2B93496

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/23/2018 16:22

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.020	0.0	0.117	
ICB1	-0.010	0.0	0.009	
CCV1	0.982	0.0	0.113	
CCB1	-0.006	0.0	0.009	
PB106795BL	-0.002	0.0	0.010	
PB106795BS	1.012	0.0	0.116	
J1161-10	0.010	0.0	0.011	
J1642-01	0.030	0.0	0.013	
J1642-01DUP	0.031	0.0	0.013	
J1642-01MS	1.025	0.0	0.118	
J1642-01MSD	1.009	0.0	0.116	
J1642-02	-0.003	0.0	0.010	
CCV2	0.995	0.0	0.114	
CCB2	-0.001	0.0	0.010	
J1161-12	0.010	0.0	0.011	
CCV3	0.970	0.0	0.112	
CCB3	-0.005	0.0	0.010	

N	17
Mean	0.416
SD	0.5058
CV%	121.70

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Feb 23 15:18:11 2018

Fri Feb 23 16:21:57 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0522	mg/l	2/23/2018 15:18:11	
0.1PPM	A	Phenolics-	P	0.0986	mg/l	2/23/2018 15:18:12	
0.25PPM	A	Phenolics-	P	0.2509	mg/l	2/23/2018 15:18:13	
0.50PPM	A	Phenolics-	P	0.4996	mg/l	2/23/2018 15:18:14	
1.0PPM	A	Phenolics-	P	0.9973	mg/l	2/23/2018 15:18:15	
2.0PPM	A	Phenolics-	P	2.0013	mg/l	2/23/2018 15:18:16	
ICV1	S	Phenolics-	P	1.0199	mg/l	2/23/2018 15:42:28	
ICB1	S	Phenolics-	P	-0.0101	mg/l	2/23/2018 15:42:29	
CCV1	S	Phenolics-	P	0.9824	mg/l	2/23/2018 15:42:30	
CCB1	S	Phenolics-	P	-0.0063	mg/l	2/23/2018 15:42:31	
PB106795BL	S	Phenolics-	P	-0.0018	mg/l	2/23/2018 15:42:32	
PB106795BS	S	Phenolics-	P	1.0122	mg/l	2/23/2018 15:42:33	
J1161-10	S	Phenolics-	P	0.0098	mg/l	2/23/2018 15:42:34	
J1642-01	S	Phenolics-	P	0.0302	mg/l	2/23/2018 15:42:36	
J1642-01DUP	S	Phenolics-	P	0.0305	mg/l	2/23/2018 15:42:37	
J1642-01MS	S	Phenolics-	P	1.0247	mg/l	2/23/2018 15:42:38	
J1642-01MSD	S	Phenolics-	P	1.0086	mg/l	2/23/2018 15:42:39	
J1642-02	S	Phenolics-	P	-0.0031	mg/l	2/23/2018 15:47:22	
CCV2	S	Phenolics-	P	0.9951	mg/l	2/23/2018 15:47:23	
CCB2	S	Phenolics-	P	-0.0011	mg/l	2/23/2018 15:47:24	
J1161-12	S	Phenolics-	P	0.0103	mg/l	2/23/2018 16:21:55	
CCV3	S	Phenolics-	P	0.9701	mg/l	2/23/2018 16:21:56	
CCB3	S	Phenolics-	P	-0.0054	mg/l	2/23/2018 16:21:57	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

2/23/2018 15:19

Test Phenolics-


Accepted 2/23/2018 15:19

Factor 9.539

Bias 0.010

Coeff. of det. 0.999994

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.016	0.0522	0.0500	
2	Phenol-2	0.020	0.0986	0.1000	
3	Phenol-2	0.036	0.2509	0.2500	
4	Phenol-2	0.063	0.4996	0.5000	
5	Phenol-2	0.115	0.9973	1.0000	
6	Phenol-2	0.220	2.0013	2.0000	