



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB93497

Temp.: 22.4 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/23/2018	09:10
2	CAL2	1	Water			7.00	02/23/2018	09:12
3	CAL3	1	Water			10.01	02/23/2018	09:14
4	ICV	1	Water			7.00	02/23/2018	11:35
5	CCV1	1	Water			2.00	02/23/2018	11:37
6	J1669-01	1	Water			4.29	02/23/2018	11:39
7	J1669-01DUP	1	Water			4.30	02/23/2018	11:41
8	J1669-02	1	Water			4.49	02/23/2018	11:43
9	J1669-03	1	Water			4.70	02/23/2018	11:45
10	J1670-01	1	Water			4.27	02/23/2018	11:47
11	CCV2	1	Water			12.00	02/23/2018	11:50

WORKLIST(Hardcopy Internal Chain)

LB 93497

Worklist Name : PH022318

Worklist ID : 108895

Date : 2/23/2018 11:26:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/22/2018	Water	J1669-01	pH	Cool 4 deg C	STER13	E12	CELL-A-POOL-WATER	02/22/2018	SM 4500-PH B
02/22/2018	Water	J1669-02	pH	Cool 4 deg C	STER13	E12	CELL-B-POOL-WATER	02/22/2018	SM 4500-PH B
02/22/2018	Water	J1669-03	pH	Cool 4 deg C	STER13	E12	CELL-C-POOL-WATER	02/22/2018	SM 4500-PH B
02/22/2018	Water	J1670-01	pH	Cool 4 deg C	STER12		POOL-CELL-WATER	02/22/2018	SM 4500-PH B

Date/Time 02/23/18 11:30 AM

Received by: SP

Relinquished by: CP

Date/Time 02/23/18 12:00 PM

Received by: CP

Relinquished by: SP