

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 02/26/2018

DATE & TIME IN: 02/26/2018 08:15

OVEN TEMP: 103 °C

Run Number: LB93503

DATE & TIME OUT: 02/26/2018 13:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB93503BL	LB93503BL	1.8906	1.8906	100	1.8906	1.8906	1.8906	0	0
2	LB93503BS	LB93503BS	1.8589	1.8589	100	1.9095	1.9095	1.9095	0.0506	506
3	J1690-01	DSN001	1.8527	1.8527	300	1.8583	1.8583	1.8583	0.0056	18.7
4	J1690-01DUP	DSN001DUP	1.8638	1.8638	300	1.8693	1.8693	1.8693	0.0055	18.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS022618

WorkList ID : 108920

Date : 2/26/2018 7:18:34 AM

1293503

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2018	Water	J1690-01	TSS	Cool 4 deg C	NAT104	E41	DSN001	02/23/2018	SM2540 D

Date/Time 02/26/18, 07:45
 Received by: AS
 Relinquished by: CP

Date/Time 02/26/18, 08:25
 Received by: CP
 Relinquished by: AS