

7470A


LB93507

INSTRUMENT ID: LV2

Linear

 μ Abs.:

36700



A= 0.0000e+000

B= 2.7345e-004

C= -5.2728e-003

Rho= 0.9998499

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	24	0.000	24				
0.2	0.200	0.223	0.023	834	0.0 %	834				
2.5	2.500	2.434	-0.066	8919	0.0 %	8919				
5.0	5.000	5.102	0.102	18679	0.0 %	18679				
7.5	7.500	7.410	-0.090	27116	0.0 %	27116				
10.0	10.000	10.030	0.030	36700	0.0 %	36700				

LB93507 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Feb 2018 10:32:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	7470A	PPB	26 Feb 2018 10:46:53	S	Std
0.2 - 1	50.2	834	-	0.2000	7470A	PPB	26 Feb 2018 10:49:02	S	Std
2.5 - 1	52.5	8919	-	2.5000	7470A	PPB	26 Feb 2018 10:51:11	S	Std
5.0 - 1	55	18679	-	5.0000	7470A	PPB	26 Feb 2018 10:53:19	S	Std
7.5 - 1	57.5	27116	-	7.5000	7470A	PPB	26 Feb 2018 10:55:27	S	Std
10.0 - 1	50	36700	-	10.0000	7470A	PPB	26 Feb 2018 10:57:34	S	Std
ICV16 - 1	ICV16	14241	3.8889	-7470A	PPB	PPB	26 Feb 2018 11:00:09	U	SMPL
ICB16 - 1	ICB16	-88	-0.0293	-7470A	PPB	PPB	26 Feb 2018 11:02:15	U	SMPL
CCV14 - 1	CCV14	18246	4.9841	-7470A	PPB	PPB	26 Feb 2018 11:04:22	U	SMPL
CCB14 - 1	CCB14	-79	-0.0269	-7470A	PPB	PPB	26 Feb 2018 11:06:28	U	SMPL
CRA - 1	CRA	823	0.2198	-7470A	PPB	PPB	26 Feb 2018 11:08:36	U	SMPL
HighStd - 1	HighStd	36090	9.8635	-7470A	PPB	PPB	26 Feb 2018 11:10:42	U	SMPL
ChkStd - 1	ChkStd	26641	7.2797	-7470A	PPB	PPB	26 Feb 2018 11:12:49	U	SMPL
PB106803BL - 1	PBW	-199	-0.0597	-7470A	PPB	PPB	26 Feb 2018 11:14:58	U	SMPL
PB106803BS - 1	LCSW	14016	3.8274	-7470A	PPB	PPB	26 Feb 2018 11:17:07	U	SMPL
PB106803TB - 1	PB106774TB	-83	-0.0280	-7470A	PPB	PPB	26 Feb 2018 11:19:14	U	SMPL
J1398-06 - 1	OK-01-022218	-10	-0.0080	-7470A	PPB	PPB	26 Feb 2018 11:21:22	U	SMPL
J1648-02 - 1	RO-124031-RO124035	17	-0.0006	-7470A	PPB	PPB	26 Feb 2018 11:23:30	U	SMPL
J1652-03 - 1	RT-4651	9	-0.0028	-7470A	PPB	PPB	26 Feb 2018 11:25:39	U	SMPL
J1652-06 - 1	RT-2898	32	0.0035	-7470A	PPB	PPB	26 Feb 2018 11:27:46	U	SMPL
CCV15 - 1	CCV15	18341	5.0100	-7470A	PPB	PPB	26 Feb 2018 11:29:52	U	SMPL
CCB15 - 1	CCB15	-71	-0.0247	-7470A	PPB	PPB	26 Feb 2018 11:31:59	U	SMPL
J1652-09 - 1	RT-4248	26	0.0018	-7470A	PPB	PPB	26 Feb 2018 11:34:07	U	SMPL
J1652-12 - 1	RT-3026	48	0.0079	-7470A	PPB	PPB	26 Feb 2018 11:36:13	U	SMPL
J1660-05 - 1	RT-4632	54	0.0095	-7470A	PPB	PPB	26 Feb 2018 11:38:19	U	SMPL
J1660-10 - 1	RT-1488	55	0.0098	-7470A	PPB	PPB	26 Feb 2018 11:40:26	U	SMPL
J1660-15 - 1	RT-3025	28	0.0024	-7470A	PPB	PPB	26 Feb 2018 11:42:34	U	SMPL
J1660-20 - 1	RT-2342	23	0.0010	-7470A	PPB	PPB	26 Feb 2018 11:44:41	U	SMPL
J1660-20DUP - 1	RT-2342DUP	10	-0.0025	-7470A	PPB	PPB	26 Feb 2018 11:46:49	U	SMPL
J1660-20MS - 1	RT-2342MS	13789	3.7653	-7470A	PPB	PPB	26 Feb 2018 11:48:58	U	SMPL
J1660-20MSD - 1	RT-2342MSD	13676	3.7344	-7470A	PPB	PPB	26 Feb 2018 11:51:06	U	SMPL
PB106804BL - 1	PBW	-81	-0.0274	-7470A	PPB	PPB	26 Feb 2018 11:53:13	U	SMPL
CCV16 - 1	CCV16	18739	5.1189	-7470A	PPB	PPB	26 Feb 2018 11:55:20	U	SMPL
CCB16 - 1	CCB16	-86	-0.0288	-7470A	PPB	PPB	26 Feb 2018 11:57:27	U	SMPL
PB106804BS - 1	LCSW	14777	4.0355	-7470A	PPB	PPB	26 Feb 2018 11:59:34	U	SMPL
J1642-05 - 1	MW-9	-41	-0.0165	-7470A	PPB	PPB	26 Feb 2018 12:01:42	U	SMPL
J1642-05DUP - 1	MW-9DUP	-89	-0.0296	-7470A	PPB	PPB	26 Feb 2018 12:03:49	U	SMPL
J1642-05MS - 1	MW-9MS	13446	3.6715	-7470A	PPB	PPB	26 Feb 2018 12:05:56	U	SMPL
J1642-05MSD - 1	MW-9MSD	13938	3.8061	-7470A	PPB	PPB	26 Feb 2018 12:08:03	U	SMPL
J1642-06 - 1	FB022118	-117	-0.0373	-7470A	PPB	PPB	26 Feb 2018 12:10:10	U	SMPL
J1660-20LX5 - 1		8	-0.0031	-7470A	PPB	PPB	26 Feb 2018 12:12:19	U	SMPL
J1660-20A - 1		13777	3.7620	-7470A	PPB	PPB	26 Feb 2018 12:14:27	U	SMPL
J1642-05LX5 - 1		-88	-0.0293	-7470A	PPB	PPB	26 Feb 2018 12:16:35	U	SMPL
J1642-05A - 1		13554	3.7010	-7470A	PPB	PPB	26 Feb 2018 12:18:42	U	SMPL
CCV17 - 1	CCV17	18424	5.0327	-7470A	PPB	PPB	26 Feb 2018 12:20:49	U	SMPL
CCB17 - 1	CCB17	-85	-0.0285	-7470A	PPB	PPB	26 Feb 2018 12:22:57	U	SMPL