


7471B

LB93573
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45413



A= 0.0000e+000

B= 2.2139e-004

C= -8.1497e-003

Rho= 0.9999416

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-8	0.000	-8				
0.2	0.200	0.200	0.000	941	0.0 %	941				
2.5	2.500	2.553	0.053	11569	0.0 %	11569				
5.0	5.000	4.973	-0.027	22498	0.0 %	22498				
7.5	7.500	7.438	-0.062	33636	0.0 %	33636				
10	10.000	10.046	0.046	45413	0.0 %	45413				

LB93513 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Feb 2018 14:59:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.00007471B	PPB		26 Feb 2018 15:03:47	S	Std
0.2 - 1	50.2	941	-	0.20007471B	PPB		26 Feb 2018 15:06:05	S	Std
2.5 - 1	52.5	11569	-	2.50007471B	PPB		26 Feb 2018 15:08:23	S	Std
5.0 - 1	55	22498	-	5.00007471B	PPB		26 Feb 2018 15:10:40	S	Std
7.5 - 1	57.5	33636	-	7.50007471B	PPB		26 Feb 2018 15:12:58	S	Std
10 - 1	510	45413	-	10.00007471B	PPB		26 Feb 2018 15:17:49	S	Std
ICV17 - 1	ICV17	17077	3.7725	-7471B	PPB		26 Feb 2018 15:20:34	U	SMPL
ICB17 - 1	ICB17	-183	-0.0487	-7471B	PPB		26 Feb 2018 15:22:49	U	SMPL
CCV18 - 1	CCV18	21948	4.8508	-7471B	PPB		26 Feb 2018 15:25:05	U	SMPL
CCB18 - 1	CCB18	-155	-0.0425	-7471B	PPB		26 Feb 2018 15:27:21	U	SMPL
CRA - 1	CRA	905	0.1922	-7471B	PPB		26 Feb 2018 15:29:37	U	SMPL
HighStd - 1	HighStd	45930	10.1601	-7471B	PPB		26 Feb 2018 15:31:52	U	SMPL
ChkStd - 1	ChkStd	31245	6.9091	-7471B	PPB		26 Feb 2018 15:34:08	U	SMPL
PB106815BL - 1	PBS	-243	-0.0619	-7471B	PPB		26 Feb 2018 15:36:26	U	SMPL
PB106815BS - 1	LCSS	15030	3.3193	-7471B	PPB		26 Feb 2018 15:38:44	U	SMPL
J1395-03 - 1	NB-01-022318	2217	0.4827	-7471B	PPB		26 Feb 2018 15:41:01	U	SMPL
J1397-09 - 1	CL-02-022318-A	2837	0.6199	-7471B	PPB		26 Feb 2018 15:43:19	U	SMPL
J1397-11 - 1	CL-02-022318-B	1155	0.2476	-7471B	PPB		26 Feb 2018 15:45:36	U	SMPL
J1397-13 - 1	CL-02-022318-C	1171	0.2511	-7471B	PPB		26 Feb 2018 15:47:54	U	SMPL
J1397-15 - 1	CL-02-022318-D	861	0.1825	-7471B	PPB		26 Feb 2018 15:50:10	U	SMPL
CCV19 - 1	CCV19	21730	4.8026	-7471B	PPB		26 Feb 2018 15:52:26	U	SMPL
CCB19 - 1	CCB19	-155	-0.0425	-7471B	PPB		26 Feb 2018 15:54:42	U	SMPL
J1397-17 - 1	CL-02-022318-E	11304	2.4944	-7471B	PPB		26 Feb 2018 15:56:58	U	SMPL
J1672-01 - 1	72-11977-72-11982	2187	0.4760	-7471B	PPB		26 Feb 2018 15:59:14	U	SMPL
J1675-01 - 1	VNJ-223	469	0.0957	-7471B	PPB		26 Feb 2018 16:01:30	U	SMPL
J1686-01 - 1	T4-CONCRETE	292	0.0565	-7471B	PPB		26 Feb 2018 16:03:45	U	SMPL
J1688-01 - 1	T-3-ROAD	80	0.0096	-7471B	PPB		26 Feb 2018 16:06:02	U	SMPL
J1688-07 - 1	T-3-SOIL	893	0.1895	-7471B	PPB		26 Feb 2018 16:08:19	U	SMPL
J1688-07DUP - 1	T-3-SOILDUP	977	0.2081	-7471B	PPB		26 Feb 2018 16:10:37	U	SMPL
J1688-07MS - 1	T-3-SOILMS	17437	3.8522	-7471B	PPB		26 Feb 2018 16:12:54	U	SMPL
J1688-07MSD - 1	T-3-SOILMSD	16884	3.7297	-7471B	PPB		26 Feb 2018 16:15:12	U	SMPL
PB106822BL - 1	PBS	-135	-0.0380	-7471B	PPB		26 Feb 2018 16:17:28	U	SMPL
CCV20 - 1	CCV20	21628	4.7800	-7471B	PPB		26 Feb 2018 16:19:44	U	SMPL
CCB20 - 1	CCB20	-181	-0.0482	-7471B	PPB		26 Feb 2018 16:22:01	U	SMPL
PB106822BS - 1	LCSS	15819	3.4940	-7471B	PPB		26 Feb 2018 16:24:18	U	SMPL
J1691-01 - 1	B1	6713	1.4780	-7471B	PPB		26 Feb 2018 16:26:34	U	SMPL
J1691-01DUP - 1	B1DUP	7444	1.6399	-7471B	PPB		26 Feb 2018 16:28:49	U	SMPL
J1691-01MS - 1	B1MS	23582	5.2126	-7471B	PPB		26 Feb 2018 16:31:06	U	SMPL
J1691-01MSD - 1	B1MSD	23277	5.1451	-7471B	PPB		26 Feb 2018 16:33:22	U	SMPL
J1691-05 - 1	B2	14945	3.3005	-7471B	PPB		26 Feb 2018 16:35:40	U	SMPL
J1691-09 - 1	B3	16625	3.6724	-7471B	PPB		26 Feb 2018 16:37:57	U	SMPL
J1691-13 - 1	B4	50690	11.2139	-7471B	PPB		26 Feb 2018 16:40:15	U	SMPL
J1691-17 - 1	B5	31720	7.0142	-7471B	PPB		26 Feb 2018 16:42:33	U	SMPL
J1691-21 - 1	B6	10287	2.2693	-7471B	PPB		26 Feb 2018 16:44:51	U	SMPL
CCV21 - 1	CCV21	21569	4.7669	-7471B	PPB		26 Feb 2018 16:47:09	U	SMPL
CCB21 - 1	CCB21	-171	-0.0460	-7471B	PPB		26 Feb 2018 16:49:25	U	SMPL
J1691-25 - 1	B7	16554	3.6567	-7471B	PPB		26 Feb 2018 16:51:42	U	SMPL
J1691-29 - 1	B8	22436	4.9589	-7471B	PPB		26 Feb 2018 16:53:59	U	SMPL
J1691-33 - 1	B9	30764	6.8026	-7471B	PPB		26 Feb 2018 16:56:15	U	SMPL
J1691-37 - 1	B10	4924	1.0820	-7471B	PPB		26 Feb 2018 16:58:31	U	SMPL
J1691-41 - 1	B11	5983	1.3164	-7471B	PPB		26 Feb 2018 17:00:49	U	SMPL
J1691-45 - 1	B12	22625	5.0007	-7471B	PPB		26 Feb 2018 17:03:05	U	SMPL
J1691-49 - 1	B13	22557	4.9857	-7471B	PPB		26 Feb 2018 17:05:23	U	SMPL
J1691-53 - 1	B14	8040	1.7718	-7471B	PPB		26 Feb 2018 17:07:41	U	SMPL
J1691-57 - 1	B15	12528	2.7654	-7471B	PPB		26 Feb 2018 17:09:59	U	SMPL
J1691-61 - 1	B16	1492	0.3222	-7471B	PPB		26 Feb 2018 17:12:15	U	SMPL
CCV22 - 1	CCV22	22478	4.9682	-7471B	PPB		26 Feb 2018 17:14:31	U	SMPL
CCB22 - 1	CCB22	-185	-0.0491	-7471B	PPB		26 Feb 2018 17:16:48	U	SMPL
J1688-07LX5 - 1		155	0.0262	-7471B	PPB		26 Feb 2018 17:19:05	U	SMPL
J1688-07A - 1		17183	3.7959	-7471B	PPB		26 Feb 2018 17:21:21	U	SMPL
J1691-01LX5 - 1		1135	0.2431	-7471B	PPB		26 Feb 2018 17:23:37	U	SMPL
J1691-01A - 1		23806	5.2622	-7471B	PPB		26 Feb 2018 17:25:54	U	SMPL
J1692-13DLX2-1		24830	5.4889	-7471B	PPB		26 Feb 2018 17:28:10	U	SMPL
CCV23 - 1	CCV23	21788	4.8154	-7471B	PPB		26 Feb 2018 17:30:27	U	SMPL
CCB23 - 1	CCB23	-206	-0.0538	-7471B	PPB		26 Feb 2018 17:32:45	U	SMPL