



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93529

Temp.: 22.5 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/26/2018	17:15
2	CAL2	1	Water			7.01	02/26/2018	17:16
3	CAL3	1	Water			10.03	02/26/2018	17:18
4	ICV	1	Water			7.00	02/26/2018	17:20
5	CCV1	1	Water			2.02	02/26/2018	17:21
6	J1400-08	1	Solid	20.02	20	9.50	02/26/2018	17:22
7	J1400-08DUP	1	Solid	20.03	20	9.51	02/26/2018	17:24
8	J1400-10	1	Solid	20.05	20	8.96	02/26/2018	17:25
9	J1400-12	1	Solid	20.01	20	9.17	02/26/2018	17:26
10	J1697-03	1	Solid	20.03	20	6.60	02/26/2018	17:29
11	J1698-04	1	Solid	20.05	20	5.76	02/26/2018	17:30
12	J1698-08	1	Solid	20.04	20	8.07	02/26/2018	17:31
13	J1698-12	1	Solid	20.02	20	7.60	02/26/2018	17:32
14	J1698-16	1	Solid	20.04	20	9.40	02/26/2018	17:35
15	J1705-06	1	Solid	20.03	20	8.60	02/26/2018	17:36
16	CCV2	1	Water			12.02	02/26/2018	17:39

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1400

WorkList ID : 108967

Date : 2/26/2018 4:15:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2018	Solid	J1400-08	Corrosivity	Cool 4 deg C	PSEG05		EO-02-022618-A	02/26/2018	9045C
02/26/2018	Solid	J1400-10	Corrosivity	Cool 4 deg C	PSEG05		EO-02-022618-B	02/26/2018	9045C
02/26/2018	Solid	J1400-12	Corrosivity	Cool 4 deg C	PSEG05		EO-02-022618-C	02/26/2018	9045C
02/26/2018	Solid	J1697-03	Corrosivity	Cool 4 deg C	PSEG02	F41	T2-T3	02/26/2018	9045C
02/26/2018	Solid	J1698-04	Corrosivity	Cool 4 deg C	PSEG01	F42	B1	02/26/2018	9045C
02/26/2018	Solid	J1698-08	Corrosivity	Cool 4 deg C	PSEG01	F42	B2	02/26/2018	9045C
02/26/2018	Solid	J1698-12	Corrosivity	Cool 4 deg C	PSEG01	F42	B3	02/26/2018	9045C
02/26/2018	Solid	J1698-16	Corrosivity	Cool 4 deg C	PSEG01	F42	B4	02/26/2018	9045C
02/26/2018	Solid	J1705-06	Corrosivity	Cool 4 deg C	PSEG01	F51	100MVAR	02/26/2018	9045C

Date/Time 02-26-18 4:20 PM
 Received by: JD
 Relinquished by: CP

Date/Time 02-26-18 7:00 PM
 Received by: SD
 Relinquished by: CP