



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93530

Temp.: 18.6 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	02/26/2018	18:00
2	CAL2	1	Water			7.01	02/26/2018	18:01
3	CAL3	1	Water			10.02	02/26/2018	18:02
4	ICV	1	Water			7.00	02/26/2018	18:03
5	CCV1	1	Water			2.01	02/26/2018	18:05
6	J1705-08	1	Water			8.26	02/26/2018	18:06
7	J1705-08DUP	1	Water			8.27	02/26/2018	18:09
8	CCV2	1	Water			12.01	02/26/2018	18:10

1B03530

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1705 WorkList ID : 108969 Date : 2/26/2018 5:39:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2018	Water	J1705-08-18	Corrosivity	Cool 4 deg C	PSEG01	F51	100MVAR	02/26/2018	9040C

Date/Time
02-26-18 5:40 PM

Received by:
JP

Relinquished by:
JP

Date/Time
02-26-18 6:45 PM

Received by:
CP

Relinquished by:
JP