

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 02/27/2018

DATE & TIME IN: 02/27/2018 11:00

OVEN TEMP: 103 °C

Run Number: LB93540

DATE & TIME OUT: 02/27/2018 15:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB93540BL	LB93540BL	1.9242	1.9242	100	1.9242	1.9242	1.9242	0	0
2	LB93540BS	LB93540BS	1.8669	1.8669	100	1.9168	1.9168	1.9168	0.0499	499
3	J1692-01	1-BSMT-CITY-WATER	1.8595	1.8595	400	1.8599	1.8599	1.8599	0.0004	1
4	J1692-01DUP	1-BSMT-CITY-WATERDUP	1.8768	1.8768	400	1.8771	1.8771	1.8771	0.0003	0.7
5	J1692-03	3-16BE-ZN-2-KITCHENSINK	1.8984	1.8984	800	1.8991	1.8991	1.8991	0.0007	0.9
6	J1692-04	4-16BE-ZN-2-MBBTHSHWR	1.8768	1.8768	800	1.877	1.877	1.877	0.0002	0.3
7	J1692-05	5-18AW-ZN-2-KITCHENSINK	1.8644	1.8644	800	1.8651	1.8651	1.8651	0.0007	0.9
8	J1692-06	6-18AW-ZN-2-MBBTHSHWR	1.9116	1.9116	800	1.9121	1.9121	1.9121	0.0005	0.6
9	J1692-07	7-15AE-ZN-2-KITCHENSINK	1.8818	1.8818	800	1.8824	1.8824	1.8824	0.0006	0.7
10	J1692-08	8-15AE-ZN-2-MBBTHSHWR	1.9177	1.9177	800	1.9185	1.9185	1.9185	0.0008	1
11	J1692-09	9-15AW-ZN-2-KITCHENSINK	1.8712	1.8712	800	1.8717	1.8717	1.8717	0.0005	0.6
12	J1692-10	10-15AW-ZN-2-MBBTHSHWR	1.8705	1.8705	800	1.8708	1.8708	1.8708	0.0003	0.4
13	J1692-11	11-14AW-ZN-2-KITCHENSINK	1.8682	1.8682	800	1.869	1.869	1.869	0.0008	1
14	J1692-12	12-14AW-ZN-2-MBBTHSHWR	1.8652	1.8652	800	1.8658	1.8658	1.8658	0.0006	0.7
15	J1692-13	13-17AW-ZN-2-KITCHENSINK	1.8661	1.8661	300	1.8665	1.8665	1.8665	0.0004	1.3

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 02/27/2018

DATE & TIME IN: 02/27/2018 11:00

OVEN TEMP: 103 °C

Run Number: LB93540

DATE & TIME OUT: 02/27/2018 15:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J1692-13DUP	13-17AW-ZN-2-KITCHENSINKDUP	1.8996	1.8996	300	1.9	1.9	1.9	0.0004	1.3
17	J1692-14	14-17AW-ZN-2-MBBTHSHWR	1.9213	1.9213	800	1.9219	1.9219	1.9219	0.0006	0.7
18	J1692-15	15-47E-ZN-5-KITCHENSINK	1.8792	1.8792	800	1.8799	1.8799	1.8799	0.0007	0.9
19	J1692-16	16-47E-ZN-5-MBBTHSHWR	1.8674	1.8674	800	1.8688	1.8688	1.8688	0.0014	1.8
20	J1692-17	17-46E-ZN-5-KITCHENSINK	1.9039	1.9039	800	1.905	1.905	1.905	0.0011	1.4
21	J1692-18	18-46E-ZN-5-MBBTHSHWR	1.8687	1.8687	800	1.8695	1.8695	1.8695	0.0008	1
22	J1692-19	20-45W-ZN-5-KITCHENSINK	1.9321	1.9321	800	1.9328	1.9328	1.9328	0.0007	0.9
23	J1692-20	21-45W-ZN-5-MBBTHSHWR	1.9528	1.9528	800	1.9532	1.9532	1.9532	0.0004	0.5
24	J1693-01	22-52W-ZN5-KITCHENSINK	1.874	1.874	800	1.8749	1.8749	1.8749	0.0009	1.1
25	J1693-02	23-52W-ZN5-MBBTHSHWR	1.8684	1.8684	400	1.8689	1.8689	1.8689	0.0005	1.3
26	J1693-03	24-52E-ZN5-KITCHENSINK	1.8774	1.8774	700	1.8778	1.8778	1.8778	0.0004	0.6
27	J1693-04	25-52E-ZN5-MBBTHSHWR	1.861	1.861	500	1.8618	1.8618	1.8618	0.0008	1.6
28	J1693-05	26-PENTHOUSE53-ZN5-KITCHENS	1.924	1.924	600	1.9246	1.9246	1.9246	0.0006	1
29	J1693-06	27-PENTHOUSE53-ZN5-MBBTHSHW	1.8687	1.8687	400	1.8692	1.8692	1.8692	0.0005	1.3
30	J1693-06DUP	27-PENTHOUSE53-ZN5-MBBTHSHW	1.9464	1.9464	400	1.947	1.947	1.947	0.0006	1.5



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 02/27/2018DATE & TIME IN: 02/27/2018 11:00OVEN TEMP: 103 °CRun Number: LB93540DATE & TIME OUT: 02/27/2018 15:00OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
31	J1705-07	100MVAR	1.867	1.867	400	1.8748	1.8748	1.8748	0.0078	19.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS022718

WorkList ID : 108972

Date : 2/27/2018 7:27:48 AM

1393540

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2018	Water	J1692-01	TSS	Cool 4 deg C	DIST01		1-BSMT-CITY-WATER	02/23/2018	SM2540 D
03/02/2018	Water	J1692-03	TSS	Cool 4 deg C	DIST01		3-16BE-ZN-2-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-04	TSS	Cool 4 deg C	DIST01		4-16BE-ZN-2-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1692-05	TSS	Cool 4 deg C	DIST01		5-18AW-ZN-2-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-06	TSS	Cool 4 deg C	DIST01		6-18AW-ZN-2-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1692-07	TSS	Cool 4 deg C	DIST01		7-15AE-ZN-2-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-08	TSS	Cool 4 deg C	DIST01		8-15AE-ZN-2-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1692-09	TSS	Cool 4 deg C	DIST01		9-15AW-ZN-2-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-10	TSS	Cool 4 deg C	DIST01		10-15AW-ZN-2-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1692-11	TSS	Cool 4 deg C	DIST01		11-14AW-ZN-2-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-12	TSS	Cool 4 deg C	DIST01		12-14AW-ZN-2-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1692-13	TSS	Cool 4 deg C	DIST01		13-17AW-ZN-2-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-14	TSS	Cool 4 deg C	DIST01		14-17AW-ZN-2-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1692-15	TSS	Cool 4 deg C	DIST01		15-47E-ZN-5-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-16	TSS	Cool 4 deg C	DIST01		16-47E-ZN-5-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1692-17	TSS	Cool 4 deg C	DIST01		17-46E-ZN-5-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-18	TSS	Cool 4 deg C	DIST01		18-46E-ZN-5-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1692-19	TSS	Cool 4 deg C	DIST01		20-45W-ZN-5-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1692-20	TSS	Cool 4 deg C	DIST01		21-45W-ZN-5-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1693-01	TSS	Cool 4 deg C	DIST01	D62	22-52W-ZN5-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1693-02	TSS	Cool 4 deg C	DIST01	D62	23-52W-ZN5-MBBTHSHWR	02/23/2018	SM2540 D

Date/Time 02/27/18, 8:30

Received by: AS

Relinquished by: PC

Date/Time 02/27/18, 11:50

Received by: D.R.

Relinquished by: AS

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS022718

WorkList ID : 108972

Date : 2/27/2018 7:27:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2018	Water	J1693-03 <u>B</u>	TSS	Cool 4 deg C	DIST01	D62	24-52E-ZN5-KITCHENSINK	02/23/2018	SM2540 D
03/02/2018	Water	J1693-04 <u>B</u>	TSS	Cool 4 deg C	DIST01	D62	25-52E-ZN5-MBBTHSHWR	02/23/2018	SM2540 D
03/02/2018	Water	J1693-05 <u>C</u>	TSS	Cool 4 deg C	DIST01	D62	26-PENTHOUSE53-ZN5-KITC	02/23/2018	SM2540 D
03/02/2018	Water	J1693-06 <u>C</u>	TSS	Cool 4 deg C	DIST01	D62	27-PENTHOUSE53-ZN5-MBB	02/23/2018	SM2540 D
03/01/2018	Water	J1705-07 <u>J</u>	TSS	Cool 4 deg C	PSEG01	F51	100MVAR	02/26/2018	SM2540 D

Date/Time 02/27/18, 8:30
 Received by: AS
 Relinquished by: D.P.

Date/Time 02/27/18, 11:15
 Received by: D.P.
 Relinquished by: AS