


7470A

LB93541
INSTRUMENT ID: CY1

Linear

μ Abs.:

36497



A= 0.0000e+000

B= 2.7456e-004

C= 5.5457e-002

Rho= 0.9996105

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.061	0.061	20	0.000	20				
0.2	0.200	0.230	0.030	637	0.0 %	637				
2.5	2.500	2.528	0.028	9007	0.0 %	9007				
5.0	5.000	4.773	-0.227	17183	0.0 %	17183				
7.5	7.500	7.531	0.031	27228	0.0 %	27228				
10.0	10.000	10.076	0.076	36497	0.0 %	36497				

LB93541 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Feb 2018 09:15:59

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	7470A	PPB	27 Feb 2018 09:18:23	S	Std
0.2 - 1	50.2	637	-	0.2000	7470A	PPB	27 Feb 2018 09:20:31	S	Std
2.5 - 1	52.5	9007	-	2.5000	7470A	PPB	27 Feb 2018 09:22:40	S	Std
5.0 - 1	55	17183	-	5.0000	7470A	PPB	27 Feb 2018 09:24:48	S	Std
7.5 - 1	57.5	27228	-	7.5000	7470A	PPB	27 Feb 2018 09:26:55	S	Std
10.0 - 1	510	36497	-	10.0000	7470A	PPB	27 Feb 2018 09:29:03	S	Std
ICV18 - 1	ICV18	13568	3.7807	-	7470A	PPB	27 Feb 2018 09:31:51	U	SMPL
ICB18 - 1	ICB18	-165	0.0102	-	7470A	PPB	27 Feb 2018 09:33:57	U	SMPL
CCV24 - 1	CCV24	19009	5.2745	-	7470A	PPB	27 Feb 2018 09:36:05	U	SMPL
CCB24 - 1	CCB24	-187	0.0041	-	7470A	PPB	27 Feb 2018 09:38:11	U	SMPL
CRA - 1	CRA	508	0.1949	-	7470A	PPB	27 Feb 2018 09:45:00	U	SMPL
HighStd - 1	HighStd	37104	10.2427	-	7470A	PPB	27 Feb 2018 09:47:06	U	SMPL
ChkStd - 1	ChkStd	25304	7.0029	-	7470A	PPB	27 Feb 2018 09:49:12	U	SMPL
PB106852BL - 1	PBW	-240	-0.0104	-	7470A	PPB	27 Feb 2018 09:51:21	U	SMPL
PB106852BS - 1	LCSW	12773	3.5624	-	7470A	PPB	27 Feb 2018 09:53:30	U	SMPL
PB106852TB - 1	PB106813TB	-158	0.0121	-	7470A	PPB	27 Feb 2018 09:55:38	U	SMPL
J1395-04 - 1	NB-01-022318	-84	0.0324	-	7470A	PPB	27 Feb 2018 09:57:46	U	SMPL
J1395-04DUP - 1	NB-01-022318DUP	-68	0.0368	-	7470A	PPB	27 Feb 2018 09:59:54	U	SMPL
J1395-04MS - 1	NB-01-022318MS	13460	3.7510	-	7470A	PPB	27 Feb 2018 10:02:03	U	SMPL
J1395-04MSD - 1	NB-01-022318MSD	13187	3.6761	-	7470A	PPB	27 Feb 2018 10:04:09	U	SMPL
CCV25 - 1	CCV25	17855	4.9577	-	7470A	PPB	27 Feb 2018 10:06:16	U	SMPL
CCB25 - 1	CCB25	-177	0.0069	-	7470A	PPB	27 Feb 2018 10:08:22	U	SMPL
J1397-10 - 1	CL-02-022318-A	-73	0.0354	-	7470A	PPB	27 Feb 2018 10:10:29	U	SMPL
J1397-12 - 1	CL-02-022318-B	-80	0.0335	-	7470A	PPB	27 Feb 2018 10:12:36	U	SMPL
J1397-14 - 1	CL-02-022318-C	-80	0.0335	-	7470A	PPB	27 Feb 2018 10:14:43	U	SMPL
J1397-16 - 1	CL-02-022318-D	-48	0.0423	-	7470A	PPB	27 Feb 2018 10:16:50	U	SMPL
J1397-18 - 1	CL-02-022318-E	-47	0.0426	-	7470A	PPB	27 Feb 2018 10:18:57	U	SMPL
J1672-04 - 1	72-11977-72-11982	-60	0.0390	-	7470A	PPB	27 Feb 2018 10:21:05	U	SMPL
J1675-05 - 1	VNJ-223	-28	0.0478	-	7470A	PPB	27 Feb 2018 10:23:12	U	SMPL
J1686-06 - 1	T4-CONCRETE	-91	0.0305	-	7470A	PPB	27 Feb 2018 10:25:21	U	SMPL
J1688-06 - 1	T-3-ROAD	-85	0.0321	-	7470A	PPB	27 Feb 2018 10:27:30	U	SMPL
J1688-12 - 1	T-3-SOIL	-97	0.0288	-	7470A	PPB	27 Feb 2018 10:29:36	U	SMPL
CCV26 - 1	CCV26	18158	5.0409	-	7470A	PPB	27 Feb 2018 10:31:43	U	SMPL
CCB26 - 1	CCB26	-149	0.0145	-	7470A	PPB	27 Feb 2018 10:33:50	U	SMPL
PB106853BL - 1	PBW	-123	0.0217	-	7470A	PPB	27 Feb 2018 10:35:58	U	SMPL
PB106853BS - 1	LCSW	12614	3.5187	-	7470A	PPB	27 Feb 2018 10:38:05	U	SMPL
PB106853TB - 1	PB106816TB	-84	0.0324	-	7470A	PPB	27 Feb 2018 10:40:12	U	SMPL
J1691-04 - 1	B1	-31	0.0469	-	7470A	PPB	27 Feb 2018 10:42:19	U	SMPL
J1691-08 - 1	B2	-6	0.0538	-	7470A	PPB	27 Feb 2018 10:44:27	U	SMPL
J1691-12 - 1	B3	2	0.0560	-	7470A	PPB	27 Feb 2018 10:46:35	U	SMPL
J1691-16 - 1	B4	-38	0.0450	-	7470A	PPB	27 Feb 2018 10:48:43	U	SMPL
J1691-20 - 1	B5	-63	0.0382	-	7470A	PPB	27 Feb 2018 10:50:52	U	SMPL
J1691-24 - 1	B6	-27	0.0480	-	7470A	PPB	27 Feb 2018 10:53:00	U	SMPL
J1691-28 - 1	B7	-5	0.0541	-	7470A	PPB	27 Feb 2018 10:55:07	U	SMPL
CCV27 - 1	CCV27	17790	4.9398	-	7470A	PPB	27 Feb 2018 10:57:15	U	SMPL
CCB27 - 1	CCB27	-115	0.0239	-	7470A	PPB	27 Feb 2018 10:59:22	U	SMPL
J1691-32 - 1	B8	-58	0.0395	-	7470A	PPB	27 Feb 2018 11:01:30	U	SMPL
J1691-36 - 1	B9	-38	0.0450	-	7470A	PPB	27 Feb 2018 11:03:37	U	SMPL
J1691-40 - 1	B10	-66	0.0373	-	7470A	PPB	27 Feb 2018 11:05:44	U	SMPL
J1691-44 - 1	B11	16	0.0598	-	7470A	PPB	27 Feb 2018 11:07:51	U	SMPL
J1691-48 - 1	B12	-47	0.0426	-	7470A	PPB	27 Feb 2018 11:09:59	U	SMPL
J1691-52 - 1	B13	-27	0.0480	-	7470A	PPB	27 Feb 2018 11:12:06	U	SMPL
J1691-56 - 1	B14	-13	0.0519	-	7470A	PPB	27 Feb 2018 11:14:14	U	SMPL
J1691-60 - 1	B15	-14	0.0516	-	7470A	PPB	27 Feb 2018 11:16:22	U	SMPL
J1691-64 - 1	B16	4	0.0566	-	7470A	PPB	27 Feb 2018 11:18:31	U	SMPL
J1691-64DUP - 1	B16DUP	-12	0.0522	-	7470A	PPB	27 Feb 2018 11:20:38	U	SMPL
CCV28 - 1	CCV28	17656	4.9031	-	7470A	PPB	27 Feb 2018 11:22:45	U	SMPL
CCB28 - 1	CCB28	-117	0.0233	-	7470A	PPB	27 Feb 2018 11:24:52	U	SMPL
J1691-64MS - 1	B16MS	13563	3.7793	-	7470A	PPB	27 Feb 2018 11:27:00	U	SMPL
J1691-64MSD - 1	B16MSD	13168	3.6708	-	7470A	PPB	27 Feb 2018 11:29:07	U	SMPL
J1395-04LX5 - 1		-192	0.0027	-	7470A	PPB	27 Feb 2018 11:31:14	U	SMPL
J1395-04A - 1		13560	3.7785	-	7470A	PPB	27 Feb 2018 11:33:21	U	SMPL
J1691-64LX5 - 1		-168	0.0093	-	7470A	PPB	27 Feb 2018 11:35:28	U	SMPL
J1691-64A - 1		13314	3.7109	-	7470A	PPB	27 Feb 2018 11:37:36	U	SMPL
CCV29 - 1	CCV29	18076	5.0184	-	7470A	PPB	27 Feb 2018 11:39:44	U	SMPL
CCB29 - 1	CCB29	-110	0.0253	-	7470A	PPB	27 Feb 2018 11:41:52	U	SMPL