

7471B


LB93554

INSTRUMENT ID: CV4

Linear

 μ Abs.:

44144



A= 0.0000e+000

B= 2.2803e-004

C= 7.1248e-003

Rho= 0.9999144

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.028	0.028	91	0.000	91				
0.2	0.200	0.206	0.006	873	0.0 %	873				
2.5	2.500	2.498	-0.002	10924	0.0 %	10924				
5.0	5.000	4.981	-0.019	21812	0.0 %	21812				
7.5	7.500	7.414	-0.086	32481	0.0 %	32481				
10	10.000	10.073	0.073	44144	0.0 %	44144				

LB93554 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Feb 2018 14:39:20

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	91	-	0.0000	7471B	PPB	27 Feb 2018 14:42:15	S	Std
0.2 - 1	50.2	873	-	0.2000	7471B	PPB	27 Feb 2018 14:44:32	S	Std
2.5 - 1	52.5	10924	-	2.5000	7471B	PPB	27 Feb 2018 14:46:50	S	Std
5.0 - 1	55	21812	-	5.0000	7471B	PPB	27 Feb 2018 14:49:07	S	Std
7.5 - 1	57.5	32481	-	7.5000	7471B	PPB	27 Feb 2018 14:51:24	S	Std
10 - 1	510	44144	-	10.0000	7471B	PPB	27 Feb 2018 14:56:20	S	Std
ICV19 - 1	ICV19	17183	3.9253	-	7471B	PPB	27 Feb 2018 14:59:00	U	SMPL
ICB19 - 1	ICB19	-69	-0.0086	-	7471B	PPB	27 Feb 2018 15:01:16	U	SMPL
CCV30 - 1	CCV30	21843	4.9880	-	7471B	PPB	27 Feb 2018 15:03:32	U	SMPL
CCB30 - 1	CCB30	-58	-0.0061	-	7471B	PPB	27 Feb 2018 15:05:47	U	SMPL
CRA - 1	CRA	849	0.2007	-	7471B	PPB	27 Feb 2018 15:08:04	U	SMPL
HighStd - 1	HighStd	45229	10.3206	-	7471B	PPB	27 Feb 2018 15:10:20	U	SMPL
ChkStd - 1	ChkStd	30676	7.0021	-	7471B	PPB	27 Feb 2018 15:12:36	U	SMPL
PB106861BL - 1	PBS	-187	-0.0355	-	7471B	PPB	27 Feb 2018 15:14:54	U	SMPL
PB106861BS - 1	LCSS	15013	3.4305	-	7471B	PPB	27 Feb 2018 15:17:12	U	SMPL
J1400-07 - 1	EO-02-022618-A	9201	2.1052	-	7471B	PPB	27 Feb 2018 15:19:28	U	SMPL
J1400-07DUP - 1	EO-02-022618-ADUP	9516	2.1770	-	7471B	PPB	27 Feb 2018 15:24:22	U	SMPL
J1400-07MS - 1	EO-02-022618-AMS	25837	5.8987	-	7471B	PPB	27 Feb 2018 15:26:39	U	SMPL
J1400-07MSD - 1	EO-02-022618-AMSD	27509	6.2800	-	7471B	PPB	27 Feb 2018 15:28:57	U	SMPL
J1400-09 - 1	EO-02-022618-B	5878	1.3475	-	7471B	PPB	27 Feb 2018 15:31:14	U	SMPL
CCV31 - 1	CCV31	21427	4.8931	-	7471B	PPB	27 Feb 2018 15:33:31	U	SMPL
CCB31 - 1	CCB31	-61	-0.0068	-	7471B	PPB	27 Feb 2018 15:35:46	U	SMPL
J1400-11 - 1	EO-02-022618-C	7117	1.6300	-	7471B	PPB	27 Feb 2018 15:38:02	U	SMPL
J1697-01 - 1	T2-T3	8135	1.8621	-	7471B	PPB	27 Feb 2018 15:40:18	U	SMPL
J1698-01 - 1	B1	983	0.2313	-	7471B	PPB	27 Feb 2018 15:42:34	U	SMPL
J1698-05 - 1	B2	7860	1.7994	-	7471B	PPB	27 Feb 2018 15:44:50	U	SMPL
J1698-09 - 1	B3	21138	4.8272	-	7471B	PPB	27 Feb 2018 15:47:07	U	SMPL
J1698-13 - 1	B4	26128	5.9651	-	7471B	PPB	27 Feb 2018 15:49:23	U	SMPL
J1702-01 - 1	FK-SF-03-003-A	3625	0.8337	-	7471B	PPB	27 Feb 2018 15:51:41	U	SMPL
J1702-03 - 1	FK-SF-03-003-B	3346	0.7701	-	7471B	PPB	27 Feb 2018 15:53:59	U	SMPL
J1703-01 - 1	FK-GF-03-004-A	1721	0.3996	-	7471B	PPB	27 Feb 2018 15:56:17	U	SMPL
J1703-03 - 1	FK-GF-03-004-B	2886	0.6652	-	7471B	PPB	27 Feb 2018 15:58:33	U	SMPL
CCV32 - 1	CCV32	21435	4.8949	-	7471B	PPB	27 Feb 2018 16:00:49	U	SMPL
CCB32 - 1	CCB32	-84	-0.0120	-	7471B	PPB	27 Feb 2018 16:03:06	U	SMPL
J1705-01 - 1	100MVAR-230kV-FOUNDATION	26916	6.1447	-	7471B	PPB	27 Feb 2018 16:05:22	U	SMPL
J1705-03 - 1	100MVAR	16586	3.7892	-	7471B	PPB	27 Feb 2018 16:07:38	U	SMPL
J1705-12 - 1	230KV-CONCRETE-REACTOR	1200	0.2808	-	7471B	PPB	27 Feb 2018 16:09:55	U	SMPL
PB106862BL - 1	PBS	117	0.0338	-	7471B	PPB	27 Feb 2018 16:12:11	U	SMPL
PB106862BS - 1	LCSS	16726	3.8211	-	7471B	PPB	27 Feb 2018 16:14:28	U	SMPL
J1699-01 - 1	B1	883	0.2085	-	7471B	PPB	27 Feb 2018 16:16:45	U	SMPL
J1699-05 - 1	B8	2096	0.4851	-	7471B	PPB	27 Feb 2018 16:19:01	U	SMPL
J1699-09 - 1	B7	1233	0.2883	-	7471B	PPB	27 Feb 2018 16:21:19	U	SMPL
J1699-13 - 1	B5	1370	0.3195	-	7471B	PPB	27 Feb 2018 16:23:37	U	SMPL
J1699-17 - 1	B3	768	0.1823	-	7471B	PPB	27 Feb 2018 16:25:53	U	SMPL
J1699-33 - 1	CCV33	22094	5.0452	-	7471B	PPB	27 Feb 2018 16:28:09	U	SMPL
CCB33 - 1	CCB33	-36	-0.0011	-	7471B	PPB	27 Feb 2018 16:30:26	U	SMPL
J1699-21 - 1	B2	2498	0.5767	-	7471B	PPB	27 Feb 2018 16:32:42	U	SMPL
J1699-25 - 1	B4	966	0.2274	-	7471B	PPB	27 Feb 2018 16:34:59	U	SMPL
J1699-29 - 1	B6	1171	0.2741	-	7471B	PPB	27 Feb 2018 16:37:15	U	SMPL
J1699-33 - 1	B9	1472	0.3428	-	7471B	PPB	27 Feb 2018 16:39:31	U	SMPL
J1699-37 - 1	B11	2080	0.4814	-	7471B	PPB	27 Feb 2018 16:41:48	U	SMPL
J1699-41 - 1	B10	1186	0.2776	-	7471B	PPB	27 Feb 2018 16:44:05	U	SMPL
J1699-45 - 1	B14	1022	0.2402	-	7471B	PPB	27 Feb 2018 16:46:23	U	SMPL
J1699-49 - 1	B16	2207	0.5104	-	7471B	PPB	27 Feb 2018 16:48:40	U	SMPL
J1699-53 - 1	B23	882	0.2082	-	7471B	PPB	27 Feb 2018 16:50:58	U	SMPL
J1699-57 - 1	B26	941	0.2217	-	7471B	PPB	27 Feb 2018 16:53:14	U	SMPL
CCV34 - 1	CCV34	22021	5.0285	-	7471B	PPB	27 Feb 2018 16:55:30	U	SMPL
CCB34 - 1	CCB34	-72	-0.0093	-	7471B	PPB	27 Feb 2018 16:57:47	U	SMPL
J1699-61 - 1	B21	1152	0.2698	-	7471B	PPB	27 Feb 2018 17:00:03	U	SMPL
J1699-65 - 1	B19	699	0.1665	-	7471B	PPB	27 Feb 2018 17:02:19	U	SMPL
J1699-69 - 1	B20	1250	0.2922	-	7471B	PPB	27 Feb 2018 17:04:35	U	SMPL
J1699-73 - 1	B27	856	0.2023	-	7471B	PPB	27 Feb 2018 17:06:52	U	SMPL
J1699-77 - 1	B24	1111	0.2605	-	7471B	PPB	27 Feb 2018 17:09:08	U	SMPL
J1699-81 - 1	B22	318	0.0796	-	7471B	PPB	27 Feb 2018 17:11:25	U	SMPL
J1699-81DUP - 1	B22DUP	318	0.0796	-	7471B	PPB	27 Feb 2018 17:13:43	U	SMPL
J1699-81MS - 1	B22MS	16711	3.8177	-	7471B	PPB	27 Feb 2018 17:16:00	U	SMPL
J1699-81MSD - 1	B22MSD	16877	3.8556	-	7471B	PPB	27 Feb 2018 17:18:18	U	SMPL
CCV35 - 1	CCV35	21999	5.0235	-	7471B	PPB	27 Feb 2018 17:20:35	U	SMPL
CCB35 - 1	CCB35	-53	-0.0050	-	7471B	PPB	27 Feb 2018 17:22:53	U	SMPL
I1400-07LX5 - 1		1864	0.4322	-	7471B	PPB	27 Feb 2018 17:25:11	U	SMPL
I1400-07A - 1		26911	6.1436	-	7471B	PPB	27 Feb 2018 17:27:28	U	SMPL
I1699-81LX5 - 1		-3	0.0064	-	7471B	PPB	27 Feb 2018 17:29:44	U	SMPL
I1699-81A - 1		16525	3.7753	-	7471B	PPB	27 Feb 2018 17:32:02	U	SMPL
CCV36 - 1	CCV36	22020	5.0283	-	7471B	PPB	27 Feb 2018 17:34:19	U	SMPL
CCB36 - 1	CCB36	-66	-0.0079	-	7471B	PPB	27 Feb 2018 17:36:36	U	SMPL