



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB93572

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/27/2018	17:20
2	CAL2	1	Water			7.00	02/27/2018	17:21
3	CAL3	1	Water			10.03	02/27/2018	17:23
4	ICV	1	Water			6.99	02/27/2018	17:25
5	CCV1	1	Water			2.01	02/27/2018	17:26
6	J1399-10	1	Solid	20.02	20	8.04	02/27/2018	17:29
7	J1399-10DUP	1	Solid	20.04	20	8.05	02/27/2018	17:30
8	J1399-12	1	Solid	20.02	20	8.30	02/27/2018	17:31
9	J1403-08	1	Solid	20.04	20	7.94	02/27/2018	17:32
10	J1708-04	1	Solid	20.06	20	5.94	02/27/2018	17:33
11	J1709-03	1	Solid	20.04	20	9.27	02/27/2018	17:34
12	J1710-03	1	Solid	20.02	20	7.54	02/27/2018	17:35
13	J1710-06	1	Solid	20.03	20	8.39	02/27/2018	17:36
14	J1710-09	1	Solid	20.01	20	8.51	02/27/2018	17:38
15	J1712-04	1	Solid	20.04	20	8.98	02/27/2018	17:40
16	CCV2	1	Water			12.02	02/27/2018	17:41

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion j1712

WorkList ID : 109029

Date : 2/27/2018 2:33:21 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2018	Solid	J1399-10	Corrosivity	Cool 4 deg C	PSEG05		OR-03-022718-A	02/27/2018	9045C
02/27/2018	Solid	J1399-12	Corrosivity	Cool 4 deg C	PSEG05		OR-03-022718-B	02/27/2018	9045C
02/27/2018	Solid	J1403-08	Corrosivity	Cool 4 deg C	PSEG05		SU-03-022718	02/27/2018	9045C
02/27/2018	Solid	J1708-04	Corrosivity	Cool 4 deg C	PSEG01	F52	RO-222-226	02/27/2018	9045C
02/27/2018	Solid	J1709-03	Corrosivity	Cool 4 deg C	PSEG01	F61	DRUM-1-2	02/27/2018	9045C
02/27/2018	Solid	J1710-03	Corrosivity	Cool 4 deg C	PSEG01	F61	RO-242	02/27/2018	9045C
02/27/2018	Solid	J1710-06	Corrosivity	Cool 4 deg C	PSEG01	F61	RO-201	02/27/2018	9045C
02/27/2018	Solid	J1710-09	Corrosivity	Cool 4 deg C	PSEG01	F61	RO-215	02/27/2018	9045C
02/27/2018	Solid	J1712-04	Corrosivity	Cool 4 deg C	PSEG01	F61	2014-TRANS-LAUREL	02/27/2018	9045C

Date/Time 02-27-18 3:00 PM

Received by: JO

Relinquished by:

Date/Time 02-27-18 7:20 PM

Received by: JO

Relinquished by: JO