



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB93573

Temp.: 16.4 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	02/27/2018	17:50
2	CAL2	1	Water			7.01	02/27/2018	17:51
3	CAL3	1	Water			10.02	02/27/2018	18:00
4	ICV	1	Water			7.00	02/27/2018	18:01
5	CCV1	1	Water			2.02	02/27/2018	18:02
6	J1709-05	1	Water			9.56	02/27/2018	18:03
7	J1709-05DUP	1	Water			9.57	02/27/2018	18:05
8	J1715-05	1	Water			6.75	02/27/2018	18:06
9	CCV2	1	Water			12.01	02/27/2018	18:07

LB 93573

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1715 WorkList ID : 109033 Date : 2/27/2018 5:31:21 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2018	Water	J1709-05-18	Corrosivity	Cool 4 deg C	PSEG01	F61	OILY-WATER	02/27/2018	9040C
02/27/2018	Water	J1715-05-18	Corrosivity	Cool 4 deg C	PSEG02	F62	500kV-OILY-WATER	02/27/2018	9040C

Date/Time 02/27/18 3:30 AM
 Received by: JD
 Relinquished by: JD

Date/Time 02/27/18 7:00 AM
 Received by: JD
 Relinquished by: JD