

7470A


LB 93576

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38099



A= 0.0000e+000

B= 2.6482e-004

C= -2.0746e-002

Rho= 0.9999036

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	118	0.000	118				
0.2	0.200	0.236	0.036	969	0.0 %	969				
2.5	2.500	2.464	-0.036	9381	0.0 %	9381				
5.0	5.000	5.012	0.012	19003	0.0 %	19003				
7.5	7.500	7.410	-0.090	28058	0.0 %	28058				
10.0	10.000	10.069	0.069	38099	0.0 %	38099				

LB93576 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Feb 2018 10:12:16

Sample ID	Extended ID	μ Abs	Conc	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	118	-	0.00007470A	PPB	28 Feb 2018 10:15:33	S	Std
0.2 - 1	50.2	969	-	0.20007470A	PPB	28 Feb 2018 10:17:41	S	Std
2.5 - 1	52.5	9381	-	2.50007470A	PPB	28 Feb 2018 10:19:49	S	Std
5.0 - 1	55	19003	-	5.00007470A	PPB	28 Feb 2018 10:21:57	S	Std
7.5 - 1	57.5	28058	-	7.50007470A	PPB	28 Feb 2018 10:24:05	S	Std
10.0 - 1	510	38099	-	10.00007470A	PPB	28 Feb 2018 10:28:47	S	Std
ICV20 - 1	ICV20	14165	3.7305	-7470A	PPB	28 Feb 2018 10:31:18	U	SMPL
ICB20 - 1	ICB20	-27	-0.0279	-7470A	PPB	28 Feb 2018 10:33:24	U	SMPL
CCV37 - 1	CCV37	18825	4.9645	-7470A	PPB	28 Feb 2018 10:35:31	U	SMPL
CCB37 - 1	CCB37	-5	-0.0221	-7470A	PPB	28 Feb 2018 10:37:37	U	SMPL
CRA - 1	CRA	863	0.2078	-7470A	PPB	28 Feb 2018 10:39:45	U	SMPL
HighStd - 1	HighStd	37873	10.0089	-7470A	PPB	28 Feb 2018 10:41:51	U	SMPL
ChkStd - 1	ChkStd	26586	7.0198	-7470A	PPB	28 Feb 2018 10:43:58	U	SMPL
PB106904BL - 1	PBW	-44	-0.0324	-7470A	PPB	28 Feb 2018 10:46:07	U	SMPL
PB106904BS - 1	LCSW	14441	3.8036	-7470A	PPB	28 Feb 2018 10:53:05	U	SMPL
PB106904TB - 1	PB106857TB	49	-0.0078	-7470A	PPB	28 Feb 2018 10:55:12	U	SMPL
J1400-08 - 1	EO-02-022618-A	102	0.0063	-7470A	PPB	28 Feb 2018 10:57:20	U	SMPL
J1400-10 - 1	EO-02-022618-B	-15	-0.0247	-7470A	PPB	28 Feb 2018 10:59:29	U	SMPL
J1400-12 - 1	EO-02-022618-C	4	-0.0197	-7470A	PPB	28 Feb 2018 11:01:37	U	SMPL
J1697-03 - 1	T2-T3	-2	-0.0213	-7470A	PPB	28 Feb 2018 11:03:44	U	SMPL
CCV38 - 1	CCV38	19020	5.0162	-7470A	PPB	28 Feb 2018 11:05:51	U	SMPL
CCB38 - 1	CCB38	0	-0.0207	-7470A	PPB	28 Feb 2018 11:07:57	U	SMPL
J1698-04 - 1	B1	-23	-0.0268	-7470A	PPB	28 Feb 2018 11:10:05	U	SMPL
J1698-08 - 1	B2	-9	-0.0231	-7470A	PPB	28 Feb 2018 11:12:11	U	SMPL
J1698-12 - 1	B3	-49	-0.0337	-7470A	PPB	28 Feb 2018 11:14:18	U	SMPL
J1698-16 - 1	B4	7	-0.0189	-7470A	PPB	28 Feb 2018 11:16:25	U	SMPL
J1705-02 - 1	100MVAR-230KV-FOUNDATION	34	-0.0117	-7470A	PPB	28 Feb 2018 11:18:33	U	SMPL
J1705-06 - 1	100MVAR	67	-0.0030	-7470A	PPB	28 Feb 2018 11:20:40	U	SMPL
J1705-06DUP - 1	100MVARDUP	46	-0.0086	-7470A	PPB	28 Feb 2018 11:22:48	U	SMPL
J1705-06MS - 1	100MVARMS	13091	3.4460	-7470A	PPB	28 Feb 2018 11:24:56	U	SMPL
J1705-06MSD - 1	100MVARMSD	12673	3.3354	-7470A	PPB	28 Feb 2018 11:27:04	U	SMPL
J1705-13 - 1	230KV-CONCRETE-REACTOR	-119	-0.0523	-7470A	PPB	28 Feb 2018 11:29:11	U	SMPL
CCV39 - 1	CCV39	19079	5.0318	-7470A	PPB	28 Feb 2018 11:31:18	U	SMPL
CCB39 - 1	CCB39	-3	-0.0215	-7470A	PPB	28 Feb 2018 11:33:25	U	SMPL
PB106905BL - 1	PBW	98	0.0052	-7470A	PPB	28 Feb 2018 11:35:33	U	SMPL
PB106905BS - 1	LCSW	13990	3.6841	-7470A	PPB	28 Feb 2018 11:37:40	U	SMPL
PB106905TB - 1	PB106860TB	45	-0.0088	-7470A	PPB	28 Feb 2018 11:39:46	U	SMPL
J1705-08 - 1	100MVAR	103	0.0065	-7470A	PPB	28 Feb 2018 11:41:54	U	SMPL
J1705-08DUP - 1	100MVARDUP	67	-0.0030	-7470A	PPB	28 Feb 2018 11:44:01	U	SMPL
J1705-08MS - 1	100MVARMS	14613	3.8491	-7470A	PPB	28 Feb 2018 11:46:09	U	SMPL
J1705-08MSD - 1	100MVARMSD	14158	3.7286	-7470A	PPB	28 Feb 2018 11:50:46	U	SMPL
PB106906BL - 1	PBW	14	-0.0170	-7470A	PPB	28 Feb 2018 11:52:55	U	SMPL
PB106906BS - 1	LCSW	14482	3.8144	-7470A	PPB	28 Feb 2018 11:58:22	U	SMPL
J1705-07 - 1	100MVAR	-107	-0.0491	-7470A	PPB	28 Feb 2018 12:00:28	U	SMPL
CCV40 - 1	CCV40	19038	5.0212	-7470A	PPB	28 Feb 2018 12:02:35	U	SMPL
CCB40 - 1	CCB40	11	-0.0178	-7470A	PPB	28 Feb 2018 12:04:42	U	SMPL
J1705-07DUP - 1	100MVARDUP	-12	-0.0239	-7470A	PPB	28 Feb 2018 12:06:50	U	SMPL
J1705-07MS - 1	100MVARMS	14434	3.8017	-7470A	PPB	28 Feb 2018 12:11:31	U	SMPL
J1705-07MSD - 1	100MVARMSD	14345	3.7781	-7470A	PPB	28 Feb 2018 12:13:38	U	SMPL
J1705-06LX5 - 1		66	-0.0033	-7470A	PPB	28 Feb 2018 12:15:45	U	SMPL
J1705-06A - 1		13621	3.5864	-7470A	PPB	28 Feb 2018 12:17:53	U	SMPL
J1705-08LX5 - 1		73	-0.0014	-7470A	PPB	28 Feb 2018 12:20:01	U	SMPL
J1705-08A - 1		13733	3.6161	-7470A	PPB	28 Feb 2018 12:22:09	U	SMPL
J1705-07LX5 - 1		57	-0.0057	-7470A	PPB	28 Feb 2018 12:24:17	U	SMPL
J1705-07A - 1		13660	3.5967	-7470A	PPB	28 Feb 2018 12:26:26	U	SMPL
CCB41 - 1	CCB41	18873	4.9773	-7470A	PPB	28 Feb 2018 12:28:33	U	SMPL
CCB41 - 1	CCB41	6	-0.0192	-7470A	PPB	28 Feb 2018 12:30:39	U	SMPL

2/28/18