

LB 93593


7471B

INSTRUMENT ID : CY1

Linear

 μ Abs.:

43264



A= 0.0000e+000

B= 2.3248e-004

C= -7.2912e-002

Rho= 0.9998788

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.059	-0.059	60	0.000	60				
0.2	0.200	0.183	-0.017	1100	0.0 %	1100				
2.5	2.500	2.567	0.067	11355	0.0 %	11355				
5.0	5.000	5.086	0.086	22192	0.0 %	22192				
7.5	7.500	7.438	-0.062	32306	0.0 %	32306				
10	10.000	9.985	-0.015	43264	0.0 %	43264				

LB93593 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Feb 2018 18:43:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	60	-	0.00007471B	PPB		28 Feb 2018 18:45:33	S	Std
0.2 - 1	50.2	1100	-	0.20007471B	PPB		28 Feb 2018 18:47:51	S	Std
2.5 - 1	52.5	11355	-	2.50007471B	PPB		28 Feb 2018 18:50:08	S	Std
5.0 - 1	55	22192	-	5.00007471B	PPB		28 Feb 2018 18:52:25	S	Std
7.5 - 1	57.5	32306	-	7.50007471B	PPB		28 Feb 2018 18:54:42	S	Std
10 - 1	510	43264	-	10.00007471B	PPB		28 Feb 2018 18:59:34	S	Std
ICV22 - 1	ICV22	17365	3.9641	-7471B	PPB		28 Feb 2018 19:02:51	U	SMPL
ICB22 - 1	ICB22	-92	-0.0943	-7471B	PPB		28 Feb 2018 19:05:06	U	SMPL
CCV48 - 1	CCV48	22167	5.0805	-7471B	PPB		28 Feb 2018 19:07:23	U	SMPL
CCB48 - 1	CCB48	-130	-0.1031	-7471B	PPB		28 Feb 2018 19:09:39	U	SMPL
CRA - 1	CRA	1060	0.1735	-7471B	PPB		28 Feb 2018 19:11:56	U	SMPL
HighStd - 1	HighStd	43258	9.9838	-7471B	PPB		28 Feb 2018 19:14:12	U	SMPL
ChkStd - 1	ChkStd	31145	7.1678	-7471B	PPB		28 Feb 2018 19:16:27	U	SMPL
PB106930BL - 1	PBS	-11	-0.0755	-7471B	PPB		28 Feb 2018 19:24:15	U	SMPL
PB106930BS - 1	LCSS	16073	3.6638	-7471B	PPB		28 Feb 2018 19:26:32	U	SMPL
J1399-09 - 1	OR-03-022718-A	2495	0.5071	-7471B	PPB		28 Feb 2018 19:33:40	U	SMPL
J1399-09DUP - 1	OR-03-022718-ADUP	2216	0.4423	-7471B	PPB		28 Feb 2018 19:35:58	U	SMPL
J1399-09MS - 1	OR-03-022718-AMS	18877	4.3157	-7471B	PPB		28 Feb 2018 19:38:15	U	SMPL
J1399-09MSD - 1	OR-03-022718-AMSD	18153	4.1473	-7471B	PPB		28 Feb 2018 19:40:33	U	SMPL
J1399-11 - 1	OR-03-022718-B	4532	0.9807	-7471B	PPB		28 Feb 2018 19:42:49	U	SMPL
CCV49 - 1	CCV49	22441	5.1442	-7471B	PPB		28 Feb 2018 19:45:05	U	SMPL
CCB49 - 1	CCB49	-16	-0.0766	-7471B	PPB		28 Feb 2018 19:47:21	U	SMPL
J1403-07 - 1	SU-03-022718	2379	0.4802	-7471B	PPB		28 Feb 2018 19:49:38	U	SMPL
J1708-01 - 1	RO-222-226	2967	0.6169	-7471B	PPB		28 Feb 2018 19:51:54	U	SMPL
J1709-01 - 1	DRUM-1-2	64411	14.9015	-7471B	PPB		28 Feb 2018 19:54:10	U	SMPL
J1710-01 - 1	RO-242	12466	2.8252	-7471B	PPB		28 Feb 2018 19:56:27	U	SMPL
J1710-04 - 1	RO-201	24657	5.6594	-7471B	PPB		28 Feb 2018 19:58:45	U	SMPL
J1710-07 - 1	RO-215	16409	3.7419	-7471B	PPB		28 Feb 2018 20:01:02	U	SMPL
J1712-01 - 1	2014-TRANS-LAUREL	580	0.0619	-7471B	PPB		28 Feb 2018 20:03:20	U	SMPL
J1715-01 - 1	500KV-OILY-DEBRIS	-8	-0.0748	-7471B	PPB		28 Feb 2018 20:05:37	U	SMPL
J1716-02 - 1	SB-01-08-12	1255	0.2189	-7471B	PPB		28 Feb 2018 20:07:54	U	SMPL
J1716-04 - 1	SB-03-12-16	60882	14.0811	-7471B	PPB		28 Feb 2018 20:10:11	U	SMPL
CCV50 - 1	CCV50	22254	5.1008	-7471B	PPB		28 Feb 2018 20:12:27	U	SMPL
CCB50 - 1	CCB50	-52	-0.0850	-7471B	PPB		28 Feb 2018 20:14:45	U	SMPL
J1716-07 - 1	SB-08-10-17	19203	4.3914	-7471B	PPB		28 Feb 2018 20:17:01	U	SMPL
J1399-09LX5 - 1		342	0.0066	-7471B	PPB		28 Feb 2018 20:19:17	U	SMPL
J1399-09A - 1		19130	4.3745	-7471B	PPB		28 Feb 2018 20:21:34	U	SMPL
J1709-01DLX2 - 1		31956	7.3563	-7471B	PPB		28 Feb 2018 20:23:50	U	SMPL
J1715-04DLX2 - 1		30553	7.0301	-7471B	PPB		28 Feb 2018 20:26:07	U	SMPL
CCV51 - 1	CCV51	22347	5.1224	-7471B	PPB		28 Feb 2018 20:28:26	U	SMPL
CCB51 - 1	CCB51	-62	-0.0873	-7471B	PPB		28 Feb 2018 20:30:44	U	SMPL

J1716-04DLX2-1

AB

03/01/18