



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93595

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/28/2018	16:35
2	CAL2	1	Water			6.99	02/28/2018	16:36
3	CAL3	1	Water			10.03	02/28/2018	16:36
4	ICV	1	Water			7.00	02/28/2018	16:37
5	CCV1	1	Water			2.01	02/28/2018	16:39
6	J1402-08	1	Solid	20.02	20	5.84	02/28/2018	16:40
7	J1402-08DUP	1	Solid	20.03	20	5.85	02/28/2018	16:41
8	J1402-10	1	Solid	20.05	20	5.85	02/28/2018	16:42
9	J1402-12	1	Solid	20.03	20	7.50	02/28/2018	16:43
10	J1402-14	1	Solid	20.02	20	7.46	02/28/2018	16:44
11	J1716-05	1	Solid	20.02	20	8.85	02/28/2018	16:45
12	J1719-02	1	Solid	20.01	20	7.86	02/28/2018	16:46
13	J1730-03	1	Solid	20.05	20	6.05	02/28/2018	16:48
14	J1736-02	1	Solid	20.03	20	9.21	02/28/2018	16:50
15	J1738-03	1	Solid	20.05	20	10.47	02/28/2018	16:51
16	CCV2	1	Water			12.02	02/28/2018	17:00

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1735

WorkList ID : 109078

Date : 2/28/2018 2:50:57 PM

1393595

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/28/2018	Solid	J1402-08	Corrosivity	Cool 4 deg C	PSEG05		PL-01-022818-A	02/28/2018	9045C
02/28/2018	Solid	J1402-10	Corrosivity	Cool 4 deg C	PSEG05		PL-01-022818-B	02/28/2018	9045C
02/28/2018	Solid	J1402-12	Corrosivity	Cool 4 deg C	PSEG05		PL-01-022818-C	02/28/2018	9045C
02/28/2018	Solid	J1402-14	Corrosivity	Cool 4 deg C	PSEG05		PL-01-022818-D	02/28/2018	9045C
02/27/2018	Solid	J1716-05	Corrosivity	Cool 4 deg C	GZAG04	D12	WC-01-01-16	02/27/2018	9045C
02/28/2018	Solid	J1719-02	Corrosivity	Cool 4 deg C	PSEG01	G11	TP-2	02/28/2018	9045C
02/28/2018	Solid	J1730-03	Corrosivity	Cool 4 deg C	PSEG01	G11	ROLLOFF-20408	02/28/2018	9045C
02/28/2018	Solid	J1736-02	Corrosivity	Cool 4 deg C	PSEG02	G21	OILY-STONE	02/28/2018	9045C
02/28/2018	Solid	J1738-03	Corrosivity	Cool 4 deg C	PSEG01	G21	OILY-STONE	02/28/2018	9045C

Date/Time 02-28-18 3:00 PM
 Received by: JP
 Relinquished by: CP

Date/Time 02-28-18 7:20 PM
 Received by: CP
 Relinquished by: JP