



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93596

Temp.: 16.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/28/2018	17:30
2	CAL2	1	Water			7.00	02/28/2018	17:31
3	CAL3	1	Water			10.03	02/28/2018	17:32
4	ICV	1	Water			7.01	02/28/2018	17:33
5	CCV1	1	Water			2.01	02/28/2018	17:35
6	J1719-04	1	Water			6.78	02/28/2018	17:36
7	J1719-04DUP	1	Water			6.79	02/28/2018	17:37
8	J1736-07	1	Water			7.18	02/28/2018	17:38
9	CCV2	1	Water			12.01	02/28/2018	17:39

LB 03596

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1736

WorkList ID : 109079

Date : 2/28/2018 2:51:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/28/2018	Water	J1719-04	Corrosivity	Cool 4 deg C	PSEG01	G11	TP-2	02/28/2018	9040C
02/28/2018	Water	J1736-07	Corrosivity	Cool 4 deg C	PSEG02	G21	OILY-WATER	02/28/2018	9040C

Date/Time 02-28-18 3:00 PM
 Received by: JP
 Relinquished by: GP

Date/Time 02-28-18 7:20 PM
 Received by: GP
 Relinquished by: JP