



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB93597

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 5N sulfuric acid, Hexavalent Chromium ICV Std., etc.

Intercept: 0 Slope: 0.7685 Regression: 0.999989

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB93597

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		1.99	0.000	0.387	0.387	0.504	02/28/2018	12:03
2	ICB		1	100	100		2.05	0.000	0.000	0.000	0.000	02/28/2018	12:04
3	CCV1	0.5	1	100	100		2.31	0.000	0.387	0.387	0.504	02/28/2018	12:04
4	CCB1		1	100	100		1.77	0.000	0.000	0.000	0.000	02/28/2018	12:05
5	PB106932BL		1	2.50	100	7.10	2.35	0.000	0.000	0.000	0.000	02/28/2018	12:05
6	PB106932BS	20	1	2.54	100	7.03	2.24	0.000	0.365	0.365	0.475	02/28/2018	12:06
7	J1399-09		1	2.50	100	7.69	1.69	0.004	0.004	0.000	0.000	02/28/2018	12:06
8	J1399-11		1	2.51	100	7.55	1.73	0.016	0.016	0.000	0.000	02/28/2018	12:07
9	J1403-07		1	2.55	100	7.22	1.82	0.012	0.012	0.000	0.000	02/28/2018	12:07
10	J1697-01		1	2.50	100	7.60	2.11	0.039	0.039	0.000	0.000	02/28/2018	12:08
11	J1698-01		1	2.53	100	7.33	2.36	0.031	0.031	0.000	0.000	02/28/2018	12:08
12	J1698-05		1	2.55	100	7.16	2.45	0.028	0.028	0.000	0.000	02/28/2018	12:09
13	J1698-09		1	2.51	100	7.33	1.70	0.018	0.018	0.000	0.000	02/28/2018	12:09
14	J1698-13		1	2.50	100	7.14	2.01	0.013	0.015	0.002	0.003	02/28/2018	12:10
15	CCV2	0.5	1	100	100		1.93	0.000	0.388	0.388	0.505	02/28/2018	12:10
16	CCB2		1	100	100		1.58	0.000	0.000	0.000	0.000	02/28/2018	12:11
17	J1703-01		1	2.52	100	7.12	2.44	0.016	0.020	0.004	0.005	02/28/2018	12:11
18	J1703-03		1	2.51	100	7.05	2.39	0.007	0.007	0.000	0.000	02/28/2018	12:12
19	J1703-03DUP		1	2.55	100	7.19	2.47	0.006	0.006	0.000	0.000	02/28/2018	12:12
20	J1703-03MSPre	40	2	2.53	100	7.33	1.63	0.008	0.349	0.341	0.444	02/28/2018	12:13
21	J1703-03MS2Ins	1284	40	2.52	100	7.02	1.81	0.006	0.584	0.578	0.752	02/28/2018	12:13
22	J1703-03MS3Post	40	2	2.55	100	7.14	2.14	0.009	0.352	0.343	0.446	02/28/2018	12:14
23	CCV3	0.5	1	100	100		2.27	0.000	0.388	0.388	0.505	02/28/2018	12:14
24	CCB3		1	100	100		1.83	0.000	0.000	0.000	0.000	02/28/2018	12:15