



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB93602

Temp.: 16.6 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	02/28/2018	17:30
2	CAL2	1	Water			7.00	02/28/2018	17:31
3	CAL3	1	Water			10.03	02/28/2018	17:32
4	ICV	1	Water			7.01	02/28/2018	17:33
5	CCV1	1	Water			2.01	02/28/2018	17:35
6	J1719-03	1	Water			6.78	02/28/2018	17:36
7	J1719-03DUP	1	Water			6.79	02/28/2018	17:37
8	J1734-01	1	Water			7.65	02/28/2018	17:39
9	J1745-01	1	Water			4.85	02/28/2018	17:40
10	J1745-02	1	Water			8.83	02/28/2018	17:41
11	J1745-03	1	Water			5.92	02/28/2018	17:42
12	J1745-05	1	Water			7.58	02/28/2018	17:43
13	J1747-01	1	Water			7.21	02/28/2018	17:45
14	J1747-02	1	Water			8.70	02/28/2018	17:46
15	J1747-03	1	Water			7.24	02/28/2018	17:48
16	CCV2	1	Water			12.01	02/28/2018	17:49
17	J1747-04	1	Water			7.20	02/28/2018	17:50
18	J1747-05	1	Water			7.23	02/28/2018	17:51
19	CCV3	1	Water			2.01	02/28/2018	17:52

1903602

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1736

WorkList ID : 109080

Date : 2/28/2018 2:51:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/28/2018	Water	J1719-03	pH	Cool 4 deg C	PSEG01	G11	TP-2	02/27/2018	9040C
02/28/2018	Water	J1734-01	pH	Cool 4 deg C	WSP001	D52	PB-2-OW	02/28/2018	9040C
02/28/2018	Water	J1736-06	pH	Cool 4 deg C	PSEG02	G21	OILY-WATER	02/28/2018	9040C
02/28/2018	Water	J1745-01	pH	Cool 4 deg C	PSEG02	G22	19614	02/28/2018	9040C
02/28/2018	Water	J1745-02	pH	Cool 4 deg C	PSEG02	G22	19613	02/28/2018	9040C
02/28/2018	Water	J1745-03	pH	Cool 4 deg C	PSEG02	G22	19612	02/28/2018	9040C
02/28/2018	Water	J1745-05	pH	Cool 4 deg C	PSEG02	G22	15294	02/28/2018	9040C
02/28/2018	Water	J1747-01	pH	Cool 4 deg C	PSEG01	G31	19615-19616-19617-19618	02/28/2018	9040C
02/28/2018	Water	J1747-02	pH	Cool 4 deg C	PSEG01	G31	19615	02/28/2018	9040C
02/28/2018	Water	J1747-03	pH	Cool 4 deg C	PSEG01	G31	19616	02/28/2018	9040C
02/28/2018	Water	J1747-04	pH	Cool 4 deg C	PSEG01	G31	19617	02/28/2018	9040C
02/28/2018	Water	J1747-05	pH	Cool 4 deg C	PSEG01	G31	19618	02/28/2018	9040C

Date/Time 02-28-18 3:30 PM
 Received by: Jg
 Relinquished by: Jg

Date/Time 02-28-18 7:10 PM
 Received by: Jg
 Relinquished by: Jg