


7471B

LB 93610
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42475



A= 0.0000e+000

B= 2.3600e-004

C= -1.7573e-002

Rho= 0.9999256

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-7	0.000	-7				
0.2	0.200	0.257	0.057	1163	0.0 %	1163				
2.5	2.500	2.426	-0.074	10354	0.0 %	10354				
5.0	5.000	5.051	0.051	21477	0.0 %	21477				
7.5	7.500	7.479	-0.021	31766	0.0 %	31766				
10	10.000	10.006	0.006	42475	0.0 %	42475				

LB93610 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Mar 2018 14:14:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-7	-	0.0000	7471B	PPB	01 Mar 2018 14:19:28	S	Std
0.2 - 1	50.2	1163	-	0.2000	7471B	PPB	01 Mar 2018 14:21:46	S	Std
2.5 - 1	52.5	10354	-	2.5000	7471B	PPB	01 Mar 2018 14:24:03	S	Std
5.0 - 1	55	21477	-	5.0000	7471B	PPB	01 Mar 2018 14:26:21	S	Std
7.5 - 1	57.5	31766	-	7.5000	7471B	PPB	01 Mar 2018 14:28:38	S	Std
10 - 1	51.0	42475	-	10.0000	7471B	PPB	01 Mar 2018 14:30:55	S	Std
ICV24 - 1	ICV24	16499	3.8761	-	7471B	PPB	01 Mar 2018 14:34:22	U	SMPL
ICB24 - 1	ICB24	-81	-0.0367	-	7471B	PPB	01 Mar 2018 14:36:37	U	SMPL
CCV58 - 1	CCV58	22224	5.2272	-	7471B	PPB	01 Mar 2018 14:38:54	U	SMPL
CCB58 - 1	CCB58	-103	-0.0419	-	7471B	PPB	01 Mar 2018 14:41:09	U	SMPL
CRA - 1	CRA	1044	0.2288	-	7471B	PPB	01 Mar 2018 14:43:26	U	SMPL
HighStd - 1	HighStd	42361	9.9795	-	7471B	PPB	01 Mar 2018 14:45:42	U	SMPL
ChkStd - 1	ChkStd	29823	7.0206	-	7471B	PPB	01 Mar 2018 14:47:57	U	SMPL
PB106955BL - 1	PBS	-16	-0.0213	-	7471B	PPB	01 Mar 2018 14:52:50	U	SMPL
PB106955BS - 1	LCSS	17345	4.0758	-	7471B	PPB	01 Mar 2018 14:57:36	U	SMPL
J1720-01 - 1	SP-31	291	0.0511	-	7471B	PPB	01 Mar 2018 14:59:52	U	SMPL
J1720-03 - 1	SP-38	806	0.1726	-	7471B	PPB	01 Mar 2018 15:02:11	U	SMPL
J1720-05 - 1	SP-44	751	0.1597	-	7471B	PPB	01 Mar 2018 15:04:28	U	SMPL
J1720-07 - 1	SP-30	933	0.2026	-	7471B	PPB	01 Mar 2018 15:06:46	U	SMPL
J1720-09 - 1	SP-37	16992	3.9925	-	7471B	PPB	01 Mar 2018 15:09:02	U	SMPL
CCV59 - 1	CCV59	21808	5.1290	-	7471B	PPB	01 Mar 2018 15:11:18	U	SMPL
CCB59 - 1	CCB59	-72	-0.0346	-	7471B	PPB	01 Mar 2018 15:13:35	U	SMPL
J1720-11 - 1	SP-43	17051	4.0064	-	7471B	PPB	01 Mar 2018 15:15:52	U	SMPL
J1720-13 - 1	SP-49	17457	4.1022	-	7471B	PPB	01 Mar 2018 15:18:08	U	SMPL
J1720-15 - 1	SP-29	20996	4.9374	-	7471B	PPB	01 Mar 2018 15:20:25	U	SMPL
J1720-17 - 1	SP-36	19492	4.5825	-	7471B	PPB	01 Mar 2018 15:22:41	U	SMPL
J1720-19 - 1	SP-42	14749	3.4631	-	7471B	PPB	01 Mar 2018 15:24:59	U	SMPL
J1721-01 - 1	SP-48	16369	3.8455	-	7471B	PPB	01 Mar 2018 15:27:17	U	SMPL
J1721-03 - 1	SP-28	17775	4.1773	-	7471B	PPB	01 Mar 2018 15:29:34	U	SMPL
J1721-05 - 1	SP-35	19000	4.4664	-	7471B	PPB	01 Mar 2018 15:31:53	U	SMPL
J1721-07 - 1	SP-41	17407	4.0904	-	7471B	PPB	01 Mar 2018 15:34:11	U	SMPL
J1721-09 - 1	SP-47	16155	3.7950	-	7471B	PPB	01 Mar 2018 15:36:27	U	SMPL
CCV60 - 1	CCV60	21949	5.1623	-	7471B	PPB	01 Mar 2018 15:38:44	U	SMPL
CCB60 - 1	CCB60	-77	-0.0357	-	7471B	PPB	01 Mar 2018 15:41:00	U	SMPL
J1721-11 - 1	SP-50	19891	4.6766	-	7471B	PPB	01 Mar 2018 15:43:17	U	SMPL
J1721-13 - 1	SP-46	12014	2.8177	-	7471B	PPB	01 Mar 2018 15:45:33	U	SMPL
J1721-15 - 1	SP-40	18354	4.3139	-	7471B	PPB	01 Mar 2018 15:47:51	U	SMPL
J1721-17 - 1	SP-34	13305	3.1224	-	7471B	PPB	01 Mar 2018 15:50:07	U	SMPL
J1721-19 - 1	SP-27	17586	4.1327	-	7471B	PPB	01 Mar 2018 15:52:24	U	SMPL
J1721-19DUP - 1	SP-27DUP	17633	4.1438	-	7471B	PPB	01 Mar 2018 15:54:41	U	SMPL
J1721-19MS - 1	SP-27MS	31671	7.4567	-	7471B	PPB	01 Mar 2018 15:56:59	U	SMPL
J1721-19MSD - 1	SP-27MSD	32333	7.6129	-	7471B	PPB	01 Mar 2018 15:59:16	U	SMPL
PB106956BL - 1	PBS	-72	-0.0346	-	7471B	PPB	01 Mar 2018 16:01:34	U	SMPL
PB106956BS - 1	LCSS	15272	3.5866	-	7471B	PPB	01 Mar 2018 16:03:52	U	SMPL
CCV61 - 1	CCV61	21472	5.0497	-	7471B	PPB	01 Mar 2018 16:06:09	U	SMPL
CCB61 - 1	CCB61	-52	-0.0298	-	7471B	PPB	01 Mar 2018 16:08:25	U	SMPL
J1722-01 - 1	SP-24	14890	3.4964	-	7471B	PPB	01 Mar 2018 16:10:42	U	SMPL
J1722-03 - 1	SP-21	11811	2.7698	-	7471B	PPB	01 Mar 2018 16:12:58	U	SMPL
J1722-05 - 1	SP-18	7614	1.7793	-	7471B	PPB	01 Mar 2018 16:15:15	U	SMPL
J1722-07 - 1	SP-15	13956	3.2760	-	7471B	PPB	01 Mar 2018 16:17:31	U	SMPL
J1722-09 - 1	SP-5	14812	3.4780	-	7471B	PPB	01 Mar 2018 16:19:48	U	SMPL
J1722-11 - 1	SP-7	8441	1.9745	-	7471B	PPB	01 Mar 2018 16:22:05	U	SMPL
J1722-13 - 1	SP-9	11890	2.7884	-	7471B	PPB	01 Mar 2018 16:24:22	U	SMPL
J1722-15 - 1	SP-11	13018	3.0546	-	7471B	PPB	01 Mar 2018 16:26:39	U	SMPL
J1722-17 - 1	SP-14	13292	3.1193	-	7471B	PPB	01 Mar 2018 16:28:57	U	SMPL
J1722-19 - 1	SP-17	37133	8.7457	-	7471B	PPB	01 Mar 2018 16:31:13	U	SMPL
CCV62 - 1	CCV62	21433	5.0405	-	7471B	PPB	01 Mar 2018 16:33:29	U	SMPL
CCB62 - 1	CCB62	-46	-0.0284	-	7471B	PPB	01 Mar 2018 16:35:48	U	SMPL
J1723-01 - 1	SP-20	17382	4.0845	-	7471B	PPB	01 Mar 2018 16:38:05	U	SMPL
J1723-03 - 1	SP-23	10268	2.4056	-	7471B	PPB	01 Mar 2018 16:40:21	U	SMPL
J1723-05 - 1	SP-1	18501	4.3486	-	7471B	PPB	01 Mar 2018 16:42:38	U	SMPL
J1723-07 - 1	SP-2	26661	6.2743	-	7471B	PPB	01 Mar 2018 16:44:54	U	SMPL
J1723-09 - 1	SP-3	13310	3.1235	-	7471B	PPB	01 Mar 2018 16:47:11	U	SMPL
J1723-11 - 1	SP-4	17068	4.0104	-	7471B	PPB	01 Mar 2018 16:49:29	U	SMPL
J1723-13 - 1	SP-26	19286	4.5339	-	7471B	PPB	01 Mar 2018 16:51:46	U	SMPL
J1723-15 - 1	SP-33	18281	4.2967	-	7471B	PPB	01 Mar 2018 16:54:03	U	SMPL
J1723-17 - 1	SP-39	19804	4.6561	-	7471B	PPB	01 Mar 2018 16:56:22	U	SMPL
J1723-19 - 1	SP-45	17614	4.1393	-	7471B	PPB	01 Mar 2018 16:58:39	U	SMPL
CCV63 - 1	CCV63	21741	5.1132	-	7471B	PPB	01 Mar 2018 17:00:55	U	SMPL
CCB63 - 1	CCB63	-70	-0.0341	-	7471B	PPB	01 Mar 2018 17:03:12	U	SMPL
J1723-19DUP - 1	SP-45DUP	17661	4.1504	-	7471B	PPB	01 Mar 2018 17:08:03	U	SMPL
J1723-19MS - 1	SP-45MS	35753	8.4200	-	7471B	PPB	01 Mar 2018 17:10:19	U	SMPL
J1723-19MSD - 1	SP-45MSD	34750	8.1833	-	7471B	PPB	01 Mar 2018 17:12:36	U	SMPL
PB106957BL - 1	PBS	213	0.0327	-	7471B	PPB	01 Mar 2018 17:18:23	U	SMPL
PB106957BS - 1	LCSS	16314	3.8325	-	7471B	PPB	01 Mar 2018 17:20:39	U	SMPL
J1402-07 - 1	PL-01-022818-A	3197	0.7369	-	7471B	PPB	01 Mar 2018 17:22:56	U	SMPL
J1402-09 - 1	PL-01-022818-B	2331	0.5325	-	7471B	PPB	01 Mar 2018 17:25:13	U	SMPL
J1402-11 - 1	PL-01-022818-C	7139	1.6672	-	7471B	PPB	01 Mar 2018 17:27:31	U	SMPL
J1402-13 - 1	PL-01-022818-D	6155	1.4350	-	7471B	PPB	01 Mar 2018 17:29:49	U	SMPL
J1719-01 - 1	TP-2	67263	15.8563	-	7471B	PPB	01 Mar 2018 17:32:06	U	SMPL
CCV64 - 1	CCV64	21293	5.0075	-	7471B	PPB	01 Mar 2018 17:34:23	U	SMPL

LB93610 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Mar 2018 14:14:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB64 - 1	CCB64	-154	-0.0539	-7471B	PPB		01 Mar 2018 17:36:42	U	SMPL
J1724-01 - 1	SP-6	14291	3.3551	-7471B	PPB		01 Mar 2018 17:38:59	U	SMPL
J1724-03 - 1	SP-8	19575	4.6021	-7471B	PPB		01 Mar 2018 17:41:16	U	SMPL
J1724-05 - 1	SP-10	14615	3.4315	-7471B	PPB		01 Mar 2018 17:43:33	U	SMPL
J1724-07 - 1	SP-13	16704	3.9245	-7471B	PPB		01 Mar 2018 17:45:50	U	SMPL
J1724-09 - 1	SP-16	13028	3.0570	-7471B	PPB		01 Mar 2018 17:48:07	U	SMPL
J1724-11 - 1	SP-19	15227	3.5759	-7471B	PPB		01 Mar 2018 17:50:25	U	SMPL
J1724-13 - 1	SP-22	15372	3.6102	-7471B	PPB		01 Mar 2018 17:52:42	U	SMPL
J1724-15 - 1	SP-25	16396	3.8518	-7471B	PPB		01 Mar 2018 17:54:59	U	SMPL
J1724-17 - 1	SP-32	17616	4.1397	-7471B	PPB		01 Mar 2018 17:57:17	U	SMPL
J1724-19 - 1	SP-12	17702	4.1600	-7471B	PPB		01 Mar 2018 17:59:34	U	SMPL
CCV65 - 1	CCV65	21853	5.1397	-7471B	PPB		01 Mar 2018 18:01:51	U	SMPL
CCB65 - 1	CCB65	-128	-0.0478	-7471B	PPB		01 Mar 2018 18:04:08	U	SMPL
J1724-19DUP - 1	SP-12DUP	17767	4.1754	-7471B	PPB		01 Mar 2018 18:09:09	U	SMPL
J1724-19MS - 1	SP-12MS	37834	8.9111	-7471B	PPB		01 Mar 2018 18:11:26	U	SMPL
J1724-19MSD - 1	SP-12MSD	34303	8.0778	-7471B	PPB		01 Mar 2018 18:13:43	U	SMPL
J1730-01 - 1	ROLLOFF-20408	6778	1.5820	-7471B	PPB		01 Mar 2018 18:16:02	U	SMPL
J1735-01 - 1	5SDM	413	0.0799	-7471B	PPB		01 Mar 2018 18:18:20	U	SMPL
J1736-01 - 1	OILY-STONE	1065	0.2338	-7471B	PPB		01 Mar 2018 18:20:37	U	SMPL
J1736-03 - 1	OILY-DEBRIS	391	0.0747	-7471B	PPB		01 Mar 2018 18:22:55	U	SMPL
J1738-02 - 1	OILY-STONE	141	0.0157	-7471B	PPB		01 Mar 2018 18:25:12	U	SMPL
CCV66 - 1	CCV66	22200	5.2216	-7471B	PPB		01 Mar 2018 18:27:29	U	SMPL
CCB66 - 1	CCB66	-110	-0.0435	-7471B	PPB		01 Mar 2018 18:29:46	U	SMPL
J1721-19LX5 - 1		3247	0.7487	-7471B	PPB		01 Mar 2018 18:32:04	U	SMPL
J1721-19A - 1		33833	7.9669	-7471B	PPB		01 Mar 2018 18:34:21	U	SMPL
J1723-19LX5 - 1		3228	0.7442	-7471B	PPB		01 Mar 2018 18:36:38	U	SMPL
J1723-19A - 1		33808	7.9610	-7471B	PPB		01 Mar 2018 18:38:56	U	SMPL
J1724-19LX5 - 1		3254	0.7504	-7471B	PPB		01 Mar 2018 18:41:13	U	SMPL
J1724-19A - 1		36664	8.6350	-7471B	PPB		01 Mar 2018 18:43:32	U	SMPL
J1719-01DLX5 - 1		12238	2.8706	-7471B	PPB		01 Mar 2018 18:45:49	U	SMPL
CCV67 - 1	CCV67	22590	5.3136	-7471B	PPB		01 Mar 2018 18:48:07	U	SMPL
CCB67 - 1	CCB67	-86	-0.0379	-7471B	PPB		01 Mar 2018 18:50:24	U	SMPL