



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB93632

Temp.: 16.2 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/01/2018	18:30
2	CAL2	1	Water			7.00	03/01/2018	18:31
3	CAL3	1	Water			10.01	03/01/2018	18:32
4	ICV	1	Water			7.01	03/01/2018	18:33
5	CCV1	1	Water			2.01	03/01/2018	18:35
6	J1779-08	1	Water			7.20	03/01/2018	18:36
7	J1779-08DUP	1	Water			7.21	03/01/2018	18:37
8	CCV2	1	Water			12.02	03/01/2018	18:39

137 93632

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1779

WorkList ID : 109147

Date : 3/1/2018 6:35:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/01/2018	Water	J1779-08	pH	Cool 4 deg C	PSEG02	G31	DRUM-1-4	03/01/2018	9040C

Date/Time 03-01-18 6:30 PM
Received by: SP
Relinquished by: SP

Date/Time 03-01-18 7:00 PM
Received by: SP
Relinquished by: SP