



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB93638

Reagent/Standard	Lot/Log #
conductivity standard	WP59787

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB93638BL		1	22.4	0.91	0.96	03/02/2018	10:52
2	LB93638BS	1413	1	22.5	1341.00	1408.24	03/02/2018	11:04
3	J1763-01		1	22.3	2430.00	2562.13	03/02/2018	11:10
4	J1763-01DUP		1	22.3	2420.00	2551.59	03/02/2018	11:16

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

6B93638 WORKLIST(Hardcopy Internal Chain)

WorkList Name : con030218

WorkList ID : 109184

Date : 3/2/2018 9:43:35 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2018	Water	J1763-01	Conductance	Cool 4 deg C	ECHE01	D52	C193-50	02/24/2018	SM2510 B

Date/Time 3-2-18 10:30
 Received by: A/C
 Relinquished by: CP

Date/Time 3-2-18 11:20
 Received by: CP
 Relinquished by: AK