

7470A


LB 93641

INSTRUMENT ID: CV1

Linear

 μ Abs.:

3513E



A= 0.0000e+000

B= 2.8292e-004

C= 1.2386e-002

Rho= 0.9998421

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	-8	0.000	-8				
0.2	0.200	0.218	0.018	726	0.0 %	726				
2.5	2.500	2.486	-0.014	8742	0.0 %	8742				
5.0	5.000	4.910	-0.090	17312	0.0 %	17312				
7.5	7.500	7.622	0.122	26898	0.0 %	26898				
10.0	10.000	9.954	-0.046	35138	0.0 %	35138				

LB93641 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Mar 2018 10:31:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.0000	7470A	PPB	02 Mar 2018 10:44:35	S	Std
0.2 - 1	50.2	726	-	0.2000	7470A	PPB	02 Mar 2018 10:46:43	S	Std
2.5 - 1	52.5	8742	-	2.5000	7470A	PPB	02 Mar 2018 10:48:52	S	Std
5.0 - 1	55	17312	-	5.0000	7470A	PPB	02 Mar 2018 10:51:00	S	Std
7.5 - 1	57.5	26898	-	7.5000	7470A	PPB	02 Mar 2018 10:55:38	S	Std
10.0 - 1	510	35138	-	10.0000	7470A	PPB	02 Mar 2018 11:00:21	S	Std
ICV25 - 1	ICV25	12948	3.6757	-	7470A	PPB	02 Mar 2018 11:02:53	U	SMPL
ICB25 - 1	ICB25	-224	-0.0510	-	7470A	PPB	02 Mar 2018 11:05:01	U	SMPL
CCV68 - 1	CCV68	17481	4.9581	-	7470A	PPB	02 Mar 2018 11:07:08	U	SMPL
CCB68 - 1	CCB68	-244	-0.0566	-	7470A	PPB	02 Mar 2018 11:09:15	U	SMPL
CRA - 1	CRA	691	0.2079	-	7470A	PPB	02 Mar 2018 11:11:22	U	SMPL
HighStd - 1	HighStd	35584	10.0799	-	7470A	PPB	02 Mar 2018 11:13:29	U	SMPL
ChkStd - 1	ChkStd	24206	6.8608	-	7470A	PPB	02 Mar 2018 11:15:36	U	SMPL
PB106998BL - 1	PBW	-355	-0.0881	-	7470A	PPB	02 Mar 2018 11:17:45	U	SMPL
PB106998BS - 1	LCSW	12441	3.5322	-	7470A	PPB	02 Mar 2018 11:19:54	U	SMPL
PB106998TB - 1	PB106943TB	-168	-0.0351	-	7470A	PPB	02 Mar 2018 11:22:01	U	SMPL
J1402-08 - 1	PL-01-022818-A	-45	-0.0003	-	7470A	PPB	02 Mar 2018 11:24:10	U	SMPL
J1402-10 - 1	PL-01-022818-B	-20	0.0067	-	7470A	PPB	02 Mar 2018 11:26:19	U	SMPL
J1402-12 - 1	PL-01-022818-C	-8	0.0101	-	7470A	PPB	02 Mar 2018 11:28:27	U	SMPL
J1402-14 - 1	PL-01-022818-D	-13	0.0087	-	7470A	PPB	02 Mar 2018 11:30:34	U	SMPL
CCV69 - 1	CCV69	17552	4.9782	-	7470A	PPB	02 Mar 2018 11:32:41	U	SMPL
CCB69 - 1	CCB69	-230	-0.0527	-	7470A	PPB	02 Mar 2018 11:34:48	U	SMPL
J1699-44 - 1	B10	-41	0.0008	-	7470A	PPB	02 Mar 2018 11:36:55	U	SMPL
J1699-48 - 1	B14	-41	0.0008	-	7470A	PPB	02 Mar 2018 11:39:02	U	SMPL
J1699-52 - 1	B16	-32	0.0033	-	7470A	PPB	02 Mar 2018 11:41:09	U	SMPL
J1699-56 - 1	B23	-39	0.0014	-	7470A	PPB	02 Mar 2018 11:43:15	U	SMPL
J1699-60 - 1	B26	-39	0.0014	-	7470A	PPB	02 Mar 2018 11:45:23	U	SMPL
J1715-10 - 1	500KV-OILY-WATER	-28	0.0045	-	7470A	PPB	02 Mar 2018 11:47:31	U	SMPL
J1719-02 - 1	TP-2	45	0.0251	-	7470A	PPB	02 Mar 2018 11:49:39	U	SMPL
J1730-03 - 1	ROLLOFF-20408	-29	0.0042	-	7470A	PPB	02 Mar 2018 11:51:47	U	SMPL
J1733-01 - 1	RT-4223-RO-144	10	0.0152	-	7470A	PPB	02 Mar 2018 11:53:56	U	SMPL
J1733-02 - 1	CHVB0788	-4	0.0113	-	7470A	PPB	02 Mar 2018 11:56:03	U	SMPL
CCV70 - 1	CCV70	16869	4.7850	-	7470A	PPB	02 Mar 2018 11:58:10	U	SMPL
CCB70 - 1	CCB70	-211	-0.0473	-	7470A	PPB	02 Mar 2018 12:00:17	U	SMPL
J1735-02 - 1	5SDM	-58	-0.0040	-	7470A	PPB	02 Mar 2018 12:02:25	U	SMPL
J1736-02 - 1	OILY-STONE	-23	0.0059	-	7470A	PPB	02 Mar 2018 12:04:32	U	SMPL
J1738-03 - 1	OILY-STONE	-32	0.0033	-	7470A	PPB	02 Mar 2018 12:06:39	U	SMPL
J1738-03DUP - 1	OILY-STONEDUP	-27	0.0047	-	7470A	PPB	02 Mar 2018 12:08:46	U	SMPL
J1738-03MS - 1	OILY-STONEMS	11818	3.3560	-	7470A	PPB	02 Mar 2018 12:10:53	U	SMPL
J1738-03MSD - 1	OILY-STONEMSD	11716	3.3271	-	7470A	PPB	02 Mar 2018 12:13:01	U	SMPL
PB106999BL - 1	PBW	-161	-0.0332	-	7470A	PPB	02 Mar 2018 12:15:09	U	SMPL
PB106999BS - 1	LCSW	11887	3.3755	-	7470A	PPB	02 Mar 2018 12:17:17	U	SMPL
PB106999TB - 1	PB106958TB	-158	-0.0323	-	7470A	PPB	02 Mar 2018 12:19:26	U	SMPL
J1719-04 - 1	TP-2	-48	-0.0012	-	7470A	PPB	02 Mar 2018 12:21:33	U	SMPL
CCV71 - 1	CCV71	17187	4.8750	-	7470A	PPB	02 Mar 2018 12:23:40	U	SMPL
CCB71 - 1	CCB71	-220	-0.0499	-	7470A	PPB	02 Mar 2018 12:25:47	U	SMPL
J1719-04DUP - 1	TP-2DUP	-59	-0.0043	-	7470A	PPB	02 Mar 2018 12:27:56	U	SMPL
J1719-04MS - 1	TP-2MS	11505	3.2674	-	7470A	PPB	02 Mar 2018 12:30:03	U	SMPL
J1719-04MSD - 1	TP-2MSD	11893	3.3772	-	7470A	PPB	02 Mar 2018 12:32:10	U	SMPL
J1736-07 - 1	OILY-WATER	-184	-0.0397	-	7470A	PPB	02 Mar 2018 12:34:18	U	SMPL
J1745-02 - 1	19613	514	0.1578	-	7470A	PPB	02 Mar 2018 12:36:26	U	SMPL
J1745-03 - 1	19612	-154	-0.0312	-	7470A	PPB	02 Mar 2018 12:38:34	U	SMPL
J1745-05 - 1	15294	-35	0.0025	-	7470A	PPB	02 Mar 2018 12:40:42	U	SMPL
J1747-01 - 1	19615-19616-19617-19618	-25	0.0053	-	7470A	PPB	02 Mar 2018 12:42:50	U	SMPL
PB107002BL - 1	PBW	-27	0.0047	-	7470A	PPB	02 Mar 2018 12:44:59	U	SMPL
PB107002BS - 1	LCSW	11504	3.2671	-	7470A	PPB	02 Mar 2018 12:47:06	U	SMPL
CCV72 - 1	CCV72	17337	4.9174	-	7470A	PPB	02 Mar 2018 12:49:13	U	SMPL
CCB72 - 1	CCB72	-249	-0.0581	-	7470A	PPB	02 Mar 2018 12:51:21	U	SMPL
J1719-03 - 1	TP-2	870	0.2585	-	7470A	PPB	02 Mar 2018 12:53:29	U	SMPL
J1719-03DUP - 1	TP-2DUP	845	0.2515	-	7470A	PPB	02 Mar 2018 12:55:36	U	SMPL
J1719-03MS - 1	TP-2MS	14423	4.0930	-	7470A	PPB	02 Mar 2018 12:57:44	U	SMPL
J1719-03MSD - 1	TP-2MSD	14179	4.0239	-	7470A	PPB	02 Mar 2018 12:59:52	U	SMPL
J1736-06 - 1	OILY-WATER	-144	-0.0284	-	7470A	PPB	02 Mar 2018 13:02:01	U	SMPL
CCV73 - 1	CCV73	16984	4.8175	-	7470A	PPB	02 Mar 2018 13:04:10	U	SMPL
CCB73 - 1	CCB73	-190	-0.0414	-	7470A	PPB	02 Mar 2018 13:06:18	U	SMPL
J1738-03LX5 - 1		-47	-0.0009	-	7470A	PPB	02 Mar 2018 13:08:27	U	SMPL
J1738-03A - 1		12568	3.5681	-	7470A	PPB	02 Mar 2018 13:10:36	U	SMPL
J1719-04LX5 - 1		-151	-0.0303	-	7470A	PPB	02 Mar 2018 13:12:43	U	SMPL
J1719-04A - 1		12314	3.4963	-	7470A	PPB	02 Mar 2018 13:14:51	U	SMPL
J1719-03LX5 - 1		-94	-0.0142	-	7470A	PPB	02 Mar 2018 13:16:58	U	SMPL
J1719-03A - 1		15278	4.3349	-	7470A	PPB	02 Mar 2018 13:19:06	U	SMPL
CCV74 - 1	CCV74	17015	4.8263	-	7470A	PPB	02 Mar 2018 13:21:13	U	SMPL
CCB74 - 1	CCB74	-196	-0.0431	-	7470A	PPB	02 Mar 2018 13:23:23	U	SMPL