

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB93642

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP61786
calibration std. phosphate 1 ppm	WP62402
calibration std. phosphate 0.5 ppm	WP62401
calibration std. phosphate 0.3 ppm	WP62400
calibration std. 0 ppm	WP62397
phosphate ICV std.	WP62404
phosphate CCV std.	WP62403
calibration std. phosphate 0.1 ppm	WP62399
calibration std. phosphate 0.05 ppm	WP62398
Combined reagent	WP62406
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.001Slope: 0.6502Regression: 0.999847

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	03/02/2018	11:53
2	CAL2	0.05	1	50	50	0.033	03/02/2018	11:53
3	CAL3	0.10	1	50	50	0.065	03/02/2018	11:55
4	CAL4	0.30	1	50	50	0.186	03/02/2018	11:55
5	CAL5	0.50	1	50	50	0.329	03/02/2018	11:56
6	CAL6	1.00	1	50	50	0.649	03/02/2018	11:56



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB93642

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.503	03/02/2018	11:57
2	ICB		1	50	50	0.000	0.002	03/02/2018	11:57
3	CCV1	0.50	1	50	50	0.330	0.509	03/02/2018	11:58
4	CCB1		1	50	50	0.000	0.002	03/02/2018	11:59
5	PB107025BL		1	50	50	0.000	0.002	03/02/2018	11:59
6	PB107025BS	0.50	1	50	50	0.333	0.514	03/02/2018	12:01
7	J1690-01		1	50	50	0.105	0.163	03/02/2018	12:01
8	J1690-01DUP		1	50	50	0.096	0.149	03/02/2018	12:02
9	J1690-01MS	0.50	1	50	50	0.395	0.609	03/02/2018	12:03
10	J1690-01MSD	0.50	1	50	50	0.386	0.595	03/02/2018	12:03
11	J1711-01		1	50	50	0.020	0.032	03/02/2018	12:04
12	J1711-02		1	50	50	0.020	0.032	03/02/2018	12:04
13	CCV2	0.50	1	50	50	0.328	0.506	03/02/2018	12:05
14	CCB2		1	50	50	0.000	0.002	03/02/2018	12:05