



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93643

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 2.01 Control Limits[+/- 0.1].

True Value of CCV4 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/01/2018	17:10
2	CAL2	1	Water			7.00	03/01/2018	17:11
3	CAL3	1	Water			10.02	03/01/2018	17:12
4	ICV	1	Water			6.99	03/01/2018	17:15
5	CCV1	1	Water			2.01	03/01/2018	17:16
6	J1751-02	1	Solid	20.02	20	9.61	03/01/2018	17:19
7	J1751-02DUP	1	Solid	20.03	20	9.62	03/01/2018	17:20
8	J1751-04	1	Solid	20.05	20	8.98	03/01/2018	17:21
9	J1751-06	1	Solid	20.02	20	8.63	03/01/2018	17:22
10	J1751-08	1	Solid	20.03	20	9.11	03/01/2018	17:24
11	J1751-10	1	Solid	20.02	20	9.00	03/01/2018	17:25
12	J1751-12	1	Solid	20.01	20	8.92	03/01/2018	17:26
13	J1751-14	1	Solid	20.03	20	9.08	03/01/2018	17:27
14	J1751-16	1	Solid	20.03	20	9.50	03/01/2018	17:30
15	J1751-18	1	Solid	20.01	20	8.12	03/01/2018	17:31
16	CCV2	1	Water			12.01	03/01/2018	17:33
17	J1756-02	1	Solid	20.02	20	8.18	03/01/2018	17:34
18	J1756-04	1	Solid	20.03	20	8.32	03/01/2018	17:35
19	J1766-03	1	Solid	20.06	20	8.25	03/01/2018	17:36
20	J1767-04	1	Solid	20.04	20	10.94	03/01/2018	17:37

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J1777-01	1	Solid	20.00	20	8.02	03/01/2018	17:38
22	J1778-01	1	Solid	20.02	20	7.16	03/01/2018	17:39
23	J1778-02	1	Solid	20.03	20	7.02	03/01/2018	17:40
24	J1778-03	1	Solid	20.03	20	10.31	03/01/2018	17:42
25	J1779-03	1	Solid	20.02	20	5.91	03/01/2018	17:43
26	J1780-01	1	Solid	20.03	20	6.70	03/01/2018	17:45
27	CCV3	1	Water			2.01	03/01/2018	17:46
28	J1780-01DUP	1	Solid	20.05	20	6.71	03/01/2018	17:48
29	CCV4	1	Water			12.01	03/01/2018	17:50

LB93643

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j1779 WorkList ID : 109144 Date : 3/1/2018 4:14:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/01/2018	Solid	J1751-02	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-A	03/01/2018	9045C
03/01/2018	Solid	J1751-04	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-B	03/01/2018	9045C
03/01/2018	Solid	J1751-06	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-C	03/01/2018	9045C
03/01/2018	Solid	J1751-08	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-D	03/01/2018	9045C
03/01/2018	Solid	J1751-10	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-E	03/01/2018	9045C
03/01/2018	Solid	J1751-12	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-F	03/01/2018	9045C
03/01/2018	Solid	J1751-14	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-G	03/01/2018	9045C
03/01/2018	Solid	J1751-16	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-H	03/01/2018	9045C
03/01/2018	Solid	J1751-18	Corrosivity	Cool 4 deg C	PSEG05	G41	JC-02-030118-I	03/01/2018	9045C
03/01/2018	Solid	J1756-02	Corrosivity	Cool 4 deg C	PSEG05	G31	OR-01-030118-A	03/01/2018	9045C
03/01/2018	Solid	J1756-04	Corrosivity	Cool 4 deg C	PSEG05	G31	OR-01-030118-B	03/01/2018	9045C
03/01/2018	Solid	J1766-03	Corrosivity	Cool 4 deg C	PSEG02	G22	DRUM-1-2	03/01/2018	9045C
03/01/2018	Solid	J1767-04	Corrosivity	Cool 4 deg C	PSEG01	G22	RO-72-11998	03/01/2018	9045C
03/01/2018	Solid	J1777-01	Corrosivity	Cool 4 deg C	PSEG02	G31	COOK-RD-OILY-DEBRIS	03/01/2018	9045C
03/06/2018	Solid	J1778-01	Corrosivity	Cool 4 deg C	PSEG01	G31	TP-3	03/28/2018	9045C
03/06/2018	Solid	J1778-02	Corrosivity	Cool 4 deg C	PSEG01	G31	TP-4	03/28/2018	9045C
03/06/2018	Solid	J1778-03	Corrosivity	Cool 4 deg C	PSEG01	G31	TP-5	03/28/2018	9045C
03/01/2018	Solid	J1779-03	Corrosivity	Cool 4 deg C	PSEG02	G31	DRUM-5	03/01/2018	9045C
03/01/2018	Solid	J1780-01	Corrosivity	Cool 4 deg C	PSEG01	G32	DRUM-1-3	03/01/2018	9045C

Date/Time 03-01-18 4:14:02 PM

Received by: Jo

Relinquished by: Jo

Date/Time 03-01-18 7:25 PM

Received by: Jo

Relinquished by: Jo