



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 03/02/2018DATE & TIME IN: 03/01/2018 16:30OVEN TEMP: 103 °CRun Number: LB93648DATE & TIME OUT: 03/02/2018 10:00OVEN TEMP: 103 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB93648BL	LB93648BL	85.2386	85.2386	100	85.2386	85.2386	85.2386	0	0
2	J1734-01	PB-2-OW	75.1494	75.1494	100	75.4006	75.4006	75.4006	0.2512	2512
3	J1734-01DUP	PB-2-OWDUP	79.9925	79.9925	100	80.2435	80.2435	80.2435	0.251	2510

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS030118

WorkList ID : 109122

Date : 3/1/2018 10:57:13 AM

LB93648

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2018	Water	J1734-01	TS	Cool 4 deg C	WSPO01	D52	PB-2-OW	02/28/2018	SM2540 B

Date/Time

Received by:

Relinquished by:

03/01/18, 14:50
AS
CJP

Date/Time

Received by:

Relinquished by:

03/01/18, 16:45
CJP
AS