

7471B


L13 93 650

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41200



A= 0.0000e+000

B= 2.4302e-004

C= 2.5961e-002

Rho= 0.9997387

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	1	0.000	1				
0.2	0.200	0.229	0.029	834	0.0 %	834				
2.5	2.500	2.384	-0.116	9703	0.0 %	9703				
5.0	5.000	5.125	0.125	20983	0.0 %	20983				
7.5	7.500	7.397	-0.103	30332	0.0 %	30332				
10	10.000	10.039	0.039	41200	0.0 %	41200				

LB93650 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Mar 2018 14:43:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1	-	0.0000	7471B	PPB	02 Mar 2018 14:45:35	S	Std
0.2 - 1	50.2	834	-	0.2000	7471B	PPB	02 Mar 2018 14:47:52	S	Std
2.5 - 1	52.5	9703	-	2.5000	7471B	PPB	02 Mar 2018 14:50:10	S	Std
5.0 - 1	55	20983	-	5.0000	7471B	PPB	02 Mar 2018 14:55:02	S	Std
7.5 - 1	57.5	30332	-	7.5000	7471B	PPB	02 Mar 2018 14:57:19	S	Std
10 - 1	514	41200	-	10.0000	7471B	PPB	02 Mar 2018 14:59:37	S	Std
ICV26 - 1	ICV26	15611	3.8198	-	7471B	PPB	02 Mar 2018 15:02:17	U	SMPL
ICB26 - 1	ICB26	-230	-0.0299	-	7471B	PPB	02 Mar 2018 15:04:34	U	SMPL
CCV75 - 1	CCV75	20636	5.0410	-	7471B	PPB	02 Mar 2018 15:06:51	U	SMPL
CCB75 - 1	CCB75	-218	-0.0270	-	7471B	PPB	02 Mar 2018 15:09:06	U	SMPL
CRA - 1	CRA	802	0.2209	-	7471B	PPB	02 Mar 2018 15:11:23	U	SMPL
HighStd - 1	HighStd	41830	10.1916	-	7471B	PPB	02 Mar 2018 15:13:39	U	SMPL
ChkStd - 1	ChkStd	28413	6.9310	-	7471B	PPB	02 Mar 2018 15:15:55	U	SMPL
PB107005BL - 1	PBS	-323	-0.0525	-	7471B	PPB	02 Mar 2018 15:18:14	U	SMPL
PB107005BS - 1	LCSS	15412	3.7714	-	7471B	PPB	02 Mar 2018 15:20:32	U	SMPL
J1741-02 - 1	SB-09-10-17	50542	12.3088	-	7471B	PPB	02 Mar 2018 15:22:49	U	SMPL
J1766-01 - 1	DRUM-1-2	1730	0.4464	-	7471B	PPB	02 Mar 2018 15:25:07	U	SMPL
J1767-01 - 1	RO-72-11998-CONCRETE	439	0.1326	-	7471B	PPB	02 Mar 2018 15:27:27	U	SMPL
J1767-03 - 1	RO-72-11998	1285	0.3382	-	7471B	PPB	02 Mar 2018 15:29:44	U	SMPL
J1772-02 - 1	SB-02-06-10	9345	2.2970	-	7471B	PPB	02 Mar 2018 15:32:01	U	SMPL
CCV76 - 1	CCV76	20525	5.0140	-	7471B	PPB	02 Mar 2018 15:34:16	U	SMPL
CCB76 - 1	CCB76	-184	-0.0188	-	7471B	PPB	02 Mar 2018 15:36:32	U	SMPL
J1774-01 - 1	CLEAN-SAND-17	412	0.1261	-	7471B	PPB	02 Mar 2018 15:38:49	U	SMPL
J1774-02 - 1	CLEAN-SAND-18	260	0.0891	-	7471B	PPB	02 Mar 2018 15:41:05	U	SMPL
J1774-03 - 1	CLEAN-SAND-19	260	0.0891	-	7471B	PPB	02 Mar 2018 15:43:21	U	SMPL
J1778-04 - 1	TP-3	4630	1.1512	-	7471B	PPB	02 Mar 2018 15:45:37	U	SMPL
J1778-04DUP - 1	TP-3DUP	4659	1.1582	-	7471B	PPB	02 Mar 2018 15:47:53	U	SMPL
J1778-04MS - 1	TP-3MS	22785	5.5632	-	7471B	PPB	02 Mar 2018 15:52:46	U	SMPL
J1778-04MSD - 1	TP-3MSD	21952	5.3608	-	7471B	PPB	02 Mar 2018 15:55:03	U	SMPL
J1778-05 - 1	TP-4	1563	0.4058	-	7471B	PPB	02 Mar 2018 15:57:22	U	SMPL
J1778-06 - 1	TP-5	15430	3.7758	-	7471B	PPB	02 Mar 2018 15:59:41	U	SMPL
J1779-01 - 1	DRUM-5	-102	0.0012	-	7471B	PPB	02 Mar 2018 16:01:57	U	SMPL
CCV77 - 1	CCV77	20500	5.0079	-	7471B	PPB	02 Mar 2018 16:04:13	U	SMPL
CCB77 - 1	CCB77	-190	-0.0202	-	7471B	PPB	02 Mar 2018 16:06:29	U	SMPL
PB107006BL - 1	PBS	-14	0.0226	-	7471B	PPB	02 Mar 2018 16:08:46	U	SMPL
PB107006BS - 1	LCSS	15265	3.7357	-	7471B	PPB	02 Mar 2018 16:11:02	U	SMPL
J1751-01 - 1	JC-02-030118-A	3951	0.9861	-	7471B	PPB	02 Mar 2018 16:13:18	U	SMPL
J1751-03 - 1	JC-02-030118-B	19523	4.7705	-	7471B	PPB	02 Mar 2018 16:15:35	U	SMPL
J1751-05 - 1	JC-02-030118-C	6907	1.7045	-	7471B	PPB	02 Mar 2018 16:17:52	U	SMPL
J1751-07 - 1	JC-02-030118-D	3070	0.7720	-	7471B	PPB	02 Mar 2018 16:20:09	U	SMPL
J1751-09 - 1	JC-02-030118-E	6117	1.5125	-	7471B	PPB	02 Mar 2018 16:22:26	U	SMPL
J1751-11 - 1	JC-02-030118-F	7018	1.7315	-	7471B	PPB	02 Mar 2018 16:24:44	U	SMPL
J1751-13 - 1	JC-02-030118-G	2625	0.6639	-	7471B	PPB	02 Mar 2018 16:27:02	U	SMPL
J1751-15 - 1	JC-02-030118-H	1920	0.4926	-	7471B	PPB	02 Mar 2018 16:29:18	U	SMPL
CCV78 - 1	CCV78	21077	5.1482	-	7471B	PPB	02 Mar 2018 16:31:34	U	SMPL
CCB78 - 1	CCB78	-147	-0.0098	-	7471B	PPB	02 Mar 2018 16:33:50	U	SMPL
J1751-17 - 1	JC-02-030118-I	6027	1.4907	-	7471B	PPB	02 Mar 2018 16:36:08	U	SMPL
J1751-17DUP - 1	JC-02-030118-IDUP	5725	1.4173	-	7471B	PPB	02 Mar 2018 16:41:27	U	SMPL
J1751-17MS - 1	JC-02-030118-IMS	20871	5.0981	-	7471B	PPB	02 Mar 2018 16:43:43	U	SMPL
J1751-17MSD - 1	JC-02-030118-IMSD	22344	5.4561	-	7471B	PPB	02 Mar 2018 16:45:59	U	SMPL
J1756-01 - 1	OR-01-030118-A	1992	0.5101	-	7471B	PPB	02 Mar 2018 16:48:17	U	SMPL
J1756-03 - 1	OR-01-030118-B	3403	0.8530	-	7471B	PPB	02 Mar 2018 16:50:35	U	SMPL
J1778-04LX5 - 1		825	0.2265	-	7471B	PPB	02 Mar 2018 16:52:52	U	SMPL
J1778-04A - 1		20188	4.9321	-	7471B	PPB	02 Mar 2018 16:55:09	U	SMPL
J1751-17LX5 - 1		948	0.2563	-	7471B	PPB	02 Mar 2018 16:57:28	U	SMPL
J1751-17A - 1		22133	5.4048	-	7471B	PPB	02 Mar 2018 16:59:45	U	SMPL
CCV79 - 1	CCV79	20977	5.1239	-	7471B	PPB	02 Mar 2018 17:02:01	U	SMPL
CCB79 - 1	CCB79	-181	-0.0180	-	7471B	PPB	02 Mar 2018 17:04:18	U	SMPL
J1741-02DLX2 - 1		24387	5.9526	-	7471B	PPB	02 Mar 2018 17:23:15	U	SMPL
CCV80 - 1	CCV80	20721	5.0616	-	7471B	PPB	02 Mar 2018 17:25:32	U	SMPL
CCB80 - 1	CCB80	-219	-0.0273	-	7471B	PPB	02 Mar 2018 17:27:50	U	SMPL