



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93655

Temp.: 17.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/02/2018	15:00
2	CAL2	1	Water			7.01	03/02/2018	15:01
3	CAL3	1	Water			10.02	03/02/2018	15:03
4	ICV	1	Water			6.99	03/02/2018	15:05
5	CCV1	1	Water			2.02	03/02/2018	15:06
6	J1792-12	1	Water			6.92	03/02/2018	15:08
7	J1792-12DUP	1	Water			6.93	03/02/2018	15:10
8	J1792-14	1	Water			7.55	03/02/2018	15:11
9	J1792-16	1	Water			7.38	03/02/2018	15:12
10	CCV2	1	Water			12.02	03/02/2018	15:15

1593655

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion j1792			WorkList ID : 109202		Date : 3/2/2018 2:31:38 PM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/01/2018	Water	J1792-12 -13	Corrosivity	Cool 4 deg C	PSEG01	G32	TP-7	03/01/2018	9040C
03/01/2018	Water	J1792-14 -15	Corrosivity	Cool 4 deg C	PSEG01	G32	TP-8	03/01/2018	9040C
03/01/2018	Water	J1792-16 -17	Corrosivity	Cool 4 deg C	PSEG01	G32	TP-13	03/01/2018	9040C

Date/Time 03/02/18 21:35 PM
Received by: JP
Relinquished by: JP

Date/Time 03/02/18 4:00 PM
Received by: JP
Relinquished by: JP