



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB93659

Temp.: 15.2 °C

BalanceID: WC SC-4

SLOPE.: 97.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/02/2018	15:50
2	CAL2	1	Water			7.01	03/02/2018	15:51
3	CAL3	1	Water			10.03	03/02/2018	16:00
4	ICV	1	Water			6.99	03/02/2018	16:01
5	CCV1	1	Water			2.01	03/02/2018	16:02
6	J1797-01	1	Solid	20.02	20	10.09	03/02/2018	16:03
7	J1798-01	1	Solid	20.04	20	7.50	03/02/2018	16:05
8	J1798-01DUP	1	Solid	20.02	20	7.51	03/02/2018	16:09
9	CCV2	1	Water			12.01	03/02/2018	16:10

2393659

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1787

WorkList ID : 109207

Date : 3/2/2018 3:27:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2018	Solid	J1797-01-G	pH	Cool 4 deg C	AREC01	D21	SI-1	03/02/2018	9045C
03/02/2018	Solid	J1798-01-F	pH	Cool 4 deg C	AREC01	E52	TRK-STONE	03/02/2018	9045C

Date/Time 03-02-18 23:01PM
Received by: J R
Relinquished by:

Date/Time 03-02-18 4:40PM
Received by: CD
Relinquished by: J0