



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : _____

Run Number: LB93661

Temp.: 21.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/02/2018	13:05
2	CAL2	1	Water			7.00	03/02/2018	13:07
3	CAL3	1	Water			10.01	03/02/2018	13:09
4	ICV	1	Water			7.00	03/02/2018	13:11
5	CCV1	1	Water			2.00	03/02/2018	13:13
6	J1784-01	1	Water			6.39	03/02/2018	13:15
7	J1784-01DUP	1	Water			6.40	03/02/2018	13:17
8	CCV2	1	Water			12.00	03/02/2018	13:20

WORKLIST(Hardcopy Internal Chain)

LB93661

WorkList Name : PH1784

WorkList ID : 109193

Date : 3/2/2018 11:53:23 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/28/2018	Water	J1784-01	pH	Cool 4 deg C	STER04	D31	POOL-WATER	02/28/2018	SM 4500-PH B

Date/Time

03/02/18 1:00 PM

Received by:

SP

Relinquished by:

CP

Date/Time

03/02/18 1:45 PM

Received by:

CP

Relinquished by:

SP