



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB93662

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/02/2018	14:30
2	CAL2	1	Water			7.00	03/02/2018	14:31
3	CAL3	1	Water			10.02	03/02/2018	14:32
4	ICV	1	Water			6.99	03/02/2018	14:33
5	CCV1	1	Water			2.02	03/02/2018	14:35
6	J1785-02	1	Solid	20.02	20	8.17	03/02/2018	14:36
7	J1785-03	1	Solid	20.03	20	8.39	03/02/2018	14:39
8	J1785-04	1	Solid	20.02	20	8.15	03/02/2018	14:40
9	J1785-05	1	Solid	20.02	20	8.15	03/02/2018	14:40
10	J1785-06	1	Solid	20.02	20	8.15	03/02/2018	14:40
11	CCV2	1	Water			12.01	03/02/2018	14:41

1893662

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH J1785

WorkList ID : 109198

Date : 3/2/2018 2:12:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2018	Solid	J1785-02	pH	Cool 4 deg C	USEP01	A11	MDALH8	02/26/2018	9045C
02/26/2018	Solid	J1785-03	pH	Cool 4 deg C	USEP01	A11	MDALH9	02/26/2018	9045C
02/26/2018	Solid	J1785-04	pH	Cool 4 deg C	USEP01	A11	MDALJ0	02/26/2018	9045C
02/26/2018	Solid	J1785-05	pH	Cool 4 deg C	USEP01	A11	MDALJ0D	02/26/2018	9045C
02/26/2018	Solid	J1785-06	pH	Cool 4 deg C	USEP01	A11	MDALJ0S	02/26/2018	9045C

Date/Time 03-02-18 2:30 PM
 Received by: JP
 Relinquished by: CS

Date/Time 03-02-18 4:10 PM
 Received by: JP
 Relinquished by: CS