



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB93663

Temp.: 22.5 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/02/2018	16:30
2	CAL2	1	Water			7.01	03/02/2018	16:31
3	CAL3	1	Water			10.02	03/02/2018	16:32
4	ICV	1	Water			7.00	03/02/2018	16:33
5	CCV1	1	Water			2.01	03/02/2018	16:35
6	J1761-02	1	Solid	20.02	20	9.26	03/02/2018	16:36
7	J1783-04	1	Solid	20.02	20	7.91	03/02/2018	16:37
8	J1786-05	1	Solid	20.03	20	9.95	03/02/2018	16:38
9	J1786-06	1	Solid	20.05	20	9.83	03/02/2018	16:40
10	J1792-02	1	Solid	20.02	20	8.20	03/02/2018	16:41
11	J1792-04	1	Solid	20.03	20	8.57	03/02/2018	16:42
12	J1792-06	1	Solid	20.05	20	8.40	03/02/2018	16:43
13	J1792-08	1	Solid	20.02	20	9.05	03/02/2018	16:45
14	J1792-10	1	Solid	20.04	20	8.92	03/02/2018	16:46
15	J1792-10DUP	1	Solid	20.02	20	8.93	03/02/2018	16:47
16	CCV2	1	Water			12.01	03/02/2018	16:48

WORKLIST(Hardcopy Internal Chain)

1393663

WorkList Name : corrosivity J1783 WorkList ID : 109201

Date : 3/2/2018 2:30:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2018	Solid	J1761-02	Corrosivity	Cool 4 deg C	PSEG05	G32	TR-04-0302018	03/02/2018	9045C
03/02/2018	Solid	J1783-04	Corrosivity	Cool 4 deg C	PSEG01	G31	SLUDGE	03/02/2018	9045C
03/02/2018	Solid	J1786-05	Corrosivity	Cool 4 deg C	TULL02	D61	SAMPLE-1-24+31-WEST	03/02/2018	9045C
03/02/2018	Solid	J1786-06	Corrosivity	Cool 4 deg C	TULL02	D61	SAMPLE-2-25+71-WEST	03/02/2018	9045C
03/01/2018	Solid	J1792-02	Corrosivity	Cool 4 deg C	PSEG01	G32	TP-6	03/01/2018	9045C
03/01/2018	Solid	J1792-04	Corrosivity	Cool 4 deg C	PSEG01	G32	TP-7	03/01/2018	9045C
03/01/2018	Solid	J1792-06	Corrosivity	Cool 4 deg C	PSEG01	G32	TP-8	03/01/2018	9045C
03/01/2018	Solid	J1792-08	Corrosivity	Cool 4 deg C	PSEG01	G32	TP-12	03/01/2018	9045C
03/01/2018	Solid	J1792-10	Corrosivity	Cool 4 deg C	PSEG01	G32	TP-13	03/01/2018	9045C

Date/Time 03-02-18 2:35 PM
 Received by: JP
 Relinquished by: CB

Date/Time 03-02-18 5:20 PM
 Received by: CB
 Relinquished by: SP