



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB93665

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/02/2018	10:40
2	J1779-09		1	>100.0	>212.0	03/02/2018	10:55
3	J1779-09DUP		1	>100.0	>212.0	03/02/2018	11:20

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP3218

WorkList ID : 109166

Date : 3/2/2018 7:57:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/07/2018	Solid	J1745-10	Flash Point	Cool 4 deg C	PSEG02	G22	15478	02/28/2018	1010A
03/06/2018	✓ Water	J1779-09	Flash Point	Cool 4 deg C	PSEG02	G31	DRUM-1-4	03/01/2018	1010A

Date/Time 3/2/18 10:15 AM
 Received by: KS
 Relinquished by: CP

Date/Time 3/2/18 11:30 AM
 Received by: CP
 Relinquished by: KS

LB 93665