


245.1

LB 93669
INSTRUMENT ID: CV2

Linear

 μ Abs.:

31884



A= 0.0000e+000

B= 3.1129e-004

C= -5.7392e-003

Rho= 0.9999095

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.036	-0.036	-98	0.000	-98				
0.05	0.050	0.043	-0.007	157	0.0 %	157				
0.2	0.200	0.188	-0.012	622	0.0 %	622				
2.5	2.500	2.503	0.003	8060	0.0 %	8060				
5.0	5.000	5.084	0.084	16352	0.0 %	16352				
7.5	7.500	7.548	0.048	24266	0.0 %	24266				
10.0	10.000	9.919	-0.081	31884	0.0 %	31884				

LB93669 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Mar 2018 17:49:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-98	-	0.0000	245.1	PPB	02 Mar 2018 17:55:06	S	Std
0.05 - 1	50.05	157	-	0.0500	245.1	PPB	02 Mar 2018 17:57:14	S	Std
0.2 - 1	50.2	622	-	0.2000	245.1	PPB	02 Mar 2018 17:59:22	S	Std
2.5 - 1	52.5	8060	-	2.5000	245.1	PPB	02 Mar 2018 18:01:30	S	Std
5.0 - 1	55	16352	-	5.0000	245.1	PPB	02 Mar 2018 18:03:38	S	Std
7.5 - 1	57.5	24266	-	7.5000	245.1	PPB	02 Mar 2018 18:05:45	S	Std
10.0 - 1	510	31884	-	10.0000	245.1	PPB	02 Mar 2018 18:10:43	S	Std
ICV27 - 1	ICV27	12428	3.8630	-245.1		PPB	02 Mar 2018 18:13:21	U	SMPL
ICB27 - 1	ICB27	-218	-0.0736	-245.1		PPB	02 Mar 2018 18:15:27	U	SMPL
CCV81 - 1	CCV81	16167	5.0269	-245.1		PPB	02 Mar 2018 18:17:35	U	SMPL
CCB81 - 1	CCB81	-235	-0.0789	-245.1		PPB	02 Mar 2018 18:19:42	U	SMPL
CRA - 1	CRA	191	0.0537	-245.1		PPB	02 Mar 2018 18:21:50	U	SMPL
HighStd - 1	HighStd	32572	10.1336	-245.1		PPB	02 Mar 2018 18:23:56	U	SMPL
ChkStd - 1	ChkStd	22227	6.9133	-245.1		PPB	02 Mar 2018 18:26:03	U	SMPL
PB107027BL - 1	PBW	-283	-0.0938	-245.1		PPB	02 Mar 2018 18:28:13	U	SMPL
PB107027BS - 1	LCSW	12395	3.8527	-245.1		PPB	02 Mar 2018 18:30:22	U	SMPL
J1734-01 - 1	PB-2-OW	-349	-0.1144	-245.1		PPB	02 Mar 2018 18:32:29	U	SMPL
J1734-01DUP - 1	PB-2-OWDUP	-196	-0.0668	-245.1		PPB	02 Mar 2018 18:34:37	U	SMPL
J1734-01MS - 1	PB-2-OWMS	13967	4.3421	-245.1		PPB	02 Mar 2018 18:36:46	U	SMPL
J1734-01MSD - 1	PB-2-OWMSD	13694	4.2571	-245.1		PPB	02 Mar 2018 18:38:55	U	SMPL
CCV82 - 1	CCV82	15771	4.9036	-245.1		PPB	02 Mar 2018 18:41:01	U	SMPL
CCB82 - 1	CCB82	-237	-0.0795	-245.1		PPB	02 Mar 2018 18:43:08	U	SMPL
J1734-01LX5 - 1		-108	-0.0394	-245.1		PPB	02 Mar 2018 18:45:16	U	SMPL
J1734-01A - 1		15736	4.8927	-245.1		PPB	02 Mar 2018 18:47:22	U	SMPL
CCV83 - 1	CCV83	15783	4.9074	-245.1		PPB	02 Mar 2018 18:49:29	U	SMPL
CCB83 - 1	CCB83	-219	-0.0739	-245.1		PPB	02 Mar 2018 18:51:36	U	SMPL