


7470A

LB 93681
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35147



A= 0.0000e+000

B= 2.8384e-004

C= 3.1059e-002

Rho= 0.9998227

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	-17	0.000	-17				
0.2	0.200	0.231	0.031	704	0.0 %	704				
2.5	2.500	2.496	-0.004	8686	0.0 %	8686				
5.0	5.000	4.857	-0.143	17002	0.0 %	17002				
7.5	7.500	7.582	0.082	26604	0.0 %	26604				
10.0	10.000	10.007	0.007	35147	0.0 %	35147				

LB93681 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 05 Mar 2018 09:17:40					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-17	-	0.00007470A	PPB		05 Mar 2018 09:29:54	S	Std
0.2 - 1	50.2	704	-	0.20007470A	PPB		05 Mar 2018 09:32:02	S	Std
2.5 - 1	52.5	8686	-	2.50007470A	PPB		05 Mar 2018 09:34:11	S	Std
5.0 - 1	55	17002	-	5.00007470A	PPB		05 Mar 2018 09:36:19	S	Std
7.5 - 1	57.5	26604	-	7.50007470A	PPB		05 Mar 2018 09:38:26	S	Std
10.0 - 1	510	35147	-	10.00007470A	PPB		05 Mar 2018 09:40:34	S	Std
ICV28 - 1	ICV28	13430	3.8430	-7470A	PPB		05 Mar 2018 09:43:12	U	SMPL
ICB28 - 1	ICB28	-169	-0.0169	-7470A	PPB		05 Mar 2018 09:45:18	U	SMPL
CCV84 - 1	CCV84	17463	4.9877	-7470A	PPB		05 Mar 2018 09:47:25	U	SMPL
CCB84 - 1	CCB84	-209	-0.0283	-7470A	PPB		05 Mar 2018 09:49:32	U	SMPL
CRA - 1	CRA	688	0.2263	-7470A	PPB		05 Mar 2018 09:51:39	U	SMPL
HighStd - 1	HighStd	35281	10.0452	-7470A	PPB		05 Mar 2018 09:53:46	U	SMPL
ChkStd - 1	ChkStd	23972	6.8353	-7470A	PPB		05 Mar 2018 09:55:53	U	SMPL
PB107037BL - 1	PBW	-331	-0.0629	-7470A	PPB		05 Mar 2018 09:58:01	U	SMPL
PB107037BS - 1	LCSW	13050	3.7352	-7470A	PPB		05 Mar 2018 10:00:10	U	SMPL
PB107037TB - 1	PB106995TB	-175	-0.0186	-7470A	PPB		05 Mar 2018 10:02:18	U	SMPL
J1699-64 - 1	B21	-119	-0.0027	-7470A	PPB		05 Mar 2018 10:04:27	U	SMPL
J1699-64DUP - 1	B21DUP	-68	0.0118	-7470A	PPB		05 Mar 2018 10:06:35	U	SMPL
J1699-64MS - 1	B21MS	13154	3.7647	-7470A	PPB		05 Mar 2018 10:08:43	U	SMPL
J1699-64MSD - 1	B21MSD	12201	3.4942	-7470A	PPB		05 Mar 2018 10:10:50	U	SMPL
CCV85 - 1	CCV85	17534	5.0079	-7470A	PPB		05 Mar 2018 10:12:57	U	SMPL
CCB85 - 1	CCB85	-203	-0.0266	-7470A	PPB		05 Mar 2018 10:15:04	U	SMPL
J1699-68 - 1	B19	-74	0.0101	-7470A	PPB		05 Mar 2018 10:17:12	U	SMPL
J1699-72 - 1	B20	-55	0.0154	-7470A	PPB		05 Mar 2018 10:19:18	U	SMPL
J1699-76 - 1	B27	-43	0.0189	-7470A	PPB		05 Mar 2018 10:21:25	U	SMPL
J1699-80 - 1	B24	-105	0.0013	-7470A	PPB		05 Mar 2018 10:23:33	U	SMPL
J1699-84 - 1	B22	-57	0.0149	-7470A	PPB		05 Mar 2018 10:25:40	U	SMPL
J1718-06 - 1	B-1(5)	8	0.0333	-7470A	PPB		05 Mar 2018 10:27:48	U	SMPL
J1718-07 - 1	B-2(5)	-72	0.0106	-7470A	PPB		05 Mar 2018 10:29:56	U	SMPL
J1718-08 - 1	B-2(10)	-91	0.0052	-7470A	PPB		05 Mar 2018 10:32:04	U	SMPL
J1718-09 - 1	B-3(5)	-85	0.0069	-7470A	PPB		05 Mar 2018 10:34:12	U	SMPL
J1718-10 - 1	B-3(10)	-65	0.0126	-7470A	PPB		05 Mar 2018 10:36:19	U	SMPL
CCV86 - 1	CCV86	17798	5.0828	-7470A	PPB		05 Mar 2018 10:38:25	U	SMPL
CCB86 - 1	CCB86	-150	-0.0115	-7470A	PPB		05 Mar 2018 10:40:32	U	SMPL
J1767-02 - 1	RO-72-11998-CONCRETE	-68	0.0118	-7470A	PPB		05 Mar 2018 10:42:40	U	SMPL
J1767-04 - 1	RO-72-11998	-80	0.0084	-7470A	PPB		05 Mar 2018 10:44:46	U	SMPL
J1777-01 - 1	COOK-RD-OILY-DEBRIS	-69	0.0115	-7470A	PPB		05 Mar 2018 10:46:53	U	SMPL
J1778-01 - 1	TP-3	-110	-0.0002	-7470A	PPB		05 Mar 2018 10:49:00	U	SMPL
J1778-02 - 1	TP-4	-64	0.0129	-7470A	PPB		05 Mar 2018 10:51:08	U	SMPL
J1778-03 - 1	TP-5	-61	0.0137	-7470A	PPB		05 Mar 2018 10:53:15	U	SMPL
PB107038BL - 1	PBW	-32	0.0220	-7470A	PPB		05 Mar 2018 10:55:24	U	SMPL
PB107038BS - 1	LCSW	13699	3.9194	-7470A	PPB		05 Mar 2018 10:57:32	U	SMPL
PB107038TB - 1	PB107003TB	-133	-0.0067	-7470A	PPB		05 Mar 2018 10:59:40	U	SMPL
J1751-02 - 1	JC-02-030118-A	-101	0.0024	-7470A	PPB		05 Mar 2018 11:01:48	U	SMPL
CCV87 - 1	CCV87	17414	4.9738	-7470A	PPB		05 Mar 2018 11:03:55	U	SMPL
CCB87 - 1	CCB87	-184	-0.0212	-7470A	PPB		05 Mar 2018 11:06:02	U	SMPL
J1751-02DUP - 1	JC-02-030118-ADUP	-73	0.0103	-7470A	PPB		05 Mar 2018 11:08:11	U	SMPL
J1751-02MS - 1	JC-02-030118-AMS	13538	3.8737	-7470A	PPB		05 Mar 2018 11:10:18	U	SMPL
J1751-02MSD - 1	JC-02-030118-AMSD	12602	3.6080	-7470A	PPB		05 Mar 2018 11:12:26	U	SMPL
J1751-04 - 1	JC-02-030118-B	-175	-0.0186	-7470A	PPB		05 Mar 2018 11:14:34	U	SMPL
J1751-06 - 1	JC-02-030118-C	-85	0.0069	-7470A	PPB		05 Mar 2018 11:16:41	U	SMPL
J1751-08 - 1	JC-02-030118-D	-56	0.0152	-7470A	PPB		05 Mar 2018 11:18:49	U	SMPL
J1751-10 - 1	JC-02-030118-E	-31	0.0223	-7470A	PPB		05 Mar 2018 11:20:57	U	SMPL
J1751-12 - 1	JC-02-030118-F	-62	0.0135	-7470A	PPB		05 Mar 2018 11:23:05	U	SMPL
J1751-14 - 1	JC-02-030118-G	-13	0.0274	-7470A	PPB		05 Mar 2018 11:25:14	U	SMPL
J1751-16 - 1	JC-02-030118-H	-28	0.0231	-7470A	PPB		05 Mar 2018 11:27:22	U	SMPL
CCV88 - 1	CCV88	17026	4.8637	-7470A	PPB		05 Mar 2018 11:29:28	U	SMPL
CCB88 - 1	CCB88	-197	-0.0249	-7470A	PPB		05 Mar 2018 11:31:35	U	SMPL
J1751-18 - 1	JC-02-030118-I	-35	0.0211	-7470A	PPB		05 Mar 2018 11:33:43	U	SMPL
J1756-02 - 1	OR-01-030118-A	-33	0.0217	-7470A	PPB		05 Mar 2018 11:35:50	U	SMPL
J1756-04 - 1	OR-01-030118-B	-28	0.0231	-7470A	PPB		05 Mar 2018 11:37:57	U	SMPL
J1766-03 - 1	DRUM-1-2	-90	0.0055	-7470A	PPB		05 Mar 2018 11:40:04	U	SMPL
J1779-03 - 1	DRUM-5	-91	0.0052	-7470A	PPB		05 Mar 2018 11:42:12	U	SMPL
J1780-01 - 1	DRUM-1-3	-163	-0.0152	-7470A	PPB		05 Mar 2018 11:44:19	U	SMPL
PB107048BL - 1	PBW	25	0.0382	-7470A	PPB		05 Mar 2018 11:46:28	U	SMPL
PB107048BS - 1	LCSW	13663	3.9092	-7470A	PPB		05 Mar 2018 11:48:36	U	SMPL
PB107048TB - 1	PB107004TB	-130	-0.0058	-7470A	PPB		05 Mar 2018 11:50:45	U	SMPL
J1779-09 - 1	DRUM-1-4	-98	0.0032	-7470A	PPB		05 Mar 2018 11:52:52	U	SMPL
CCV89 - 1	CCV89	17390	4.9670	-7470A	PPB		05 Mar 2018 11:55:00	U	SMPL
CCB89 - 1	CCB89	-187	-0.0220	-7470A	PPB		05 Mar 2018 11:57:07	U	SMPL
J1779-09DUP - 1	DRUM-1-4DUP	-88	0.0061	-7470A	PPB		05 Mar 2018 11:59:16	U	SMPL
J1779-09MS - 1	DRUM-1-4MS	12525	3.5861	-7470A	PPB		05 Mar 2018 12:01:23	U	SMPL
J1779-09MSD - 1	DRUM-1-4MSD	12732	3.6449	-7470A	PPB		05 Mar 2018 12:03:31	U	SMPL
J1699-64LX5 - 1		-157	-0.0135	-7470A	PPB		05 Mar 2018 12:05:39	U	SMPL
J1699-64A - 1		13863	3.9659	-7470A	PPB		05 Mar 2018 12:07:46	U	SMPL
J1751-02LX5 - 1		-161	-0.0146	-7470A	PPB		05 Mar 2018 12:09:54	U	SMPL
J1751-02A - 1		13696	3.9185	-7470A	PPB		05 Mar 2018 12:12:02	U	SMPL
J1779-09LX5 - 1		-183	-0.0209	-7470A	PPB		05 Mar 2018 12:14:10	U	SMPL
J1779-09A - 1		13238	3.7885	-7470A	PPB		05 Mar 2018 12:16:19	U	SMPL
CCV90 - 1	CCV90	17829	5.0916	-7470A	PPB		05 Mar 2018 12:18:27	U	SMPL
CCB90 - 1	CCB90	-197	-0.0249	-7470A	PPB		05 Mar 2018 12:20:35	U	SMPL