

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/05/2018

DATE & TIME IN: 03/05/2018 09:00

OVEN TEMP: 103 °C

Run Number: LB93682

DATE & TIME OUT: 03/05/2018 13:30

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB93682BL	LB93682BL	1.8987	1.8987	100	1.8987	1.8987	1.8987	0	0
2	LB93682BS	LB93682BS	1.9042	1.9042	100	1.9542	1.9542	1.9542	0.05	500
3	J1790-01	001M-WILLETS-PT-BLVD (FEB)	1.9252	1.9252	100	1.9357	1.9357	1.9357	0.0105	105
4	J1790-01DUP	001M-WILLETS-PT-BLVD (FEB) D	1.924	1.924	100	1.9341	1.9341	1.9341	0.0101	101
5	J1790-02	002M-35TH-AVE (FEB)	1.9232	1.9232	100	1.9349	1.9349	1.9349	0.0117	117
6	J1792-11	TP-7	1.9234	1.9234	800	1.926	1.926	1.926	0.0026	3.3
7	J1792-13	TP-8	1.8871	1.8871	700	1.903	1.903	1.903	0.0159	22.7
8	J1792-15	TP-13	1.9137	1.9137	500	1.916	1.916	1.916	0.0023	4.6
9	J1794-01	030118-SIDI-F-D-M	1.9001	1.9001	100	1.9222	1.9222	1.9222	0.0221	221
10	J1794-02	030118-SIDI-F-U-M	1.9178	1.9178	200	1.9304	1.9304	1.9304	0.0126	63
11	J1794-03	030118-JETT-F-D-M	1.91	1.91	100	1.9311	1.9311	1.9311	0.0211	211
12	J1794-04	030118-JETT-F-U-M	1.9273	1.9273	300	1.9449	1.9449	1.9449	0.0176	58.7
13	J1796-01	COMP	1.9218	1.9218	100	1.9327	1.9327	1.9327	0.0109	109

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS030518

WorkList ID : 109234

Date : 3/5/2018 7:53:43 AM

1093682

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/06/2018	Water	J1790-01	TSS	Cool 4 deg C	TULL01	E42	001M-WILLETTS-PT-BLVD(FEE)	03/01/2018	SM2540 D
03/06/2018	Water	J1790-02	TSS	Cool 4 deg C	TULL01	E42	002M-35TH-AVE(FEB)	03/01/2018	SM2540 D
03/07/2018	Water	J1792-11	TSS	Cool 4 deg C	PSEG01	G32	TP-7	03/01/2018	SM2540 D
03/07/2018	Water	J1792-13	TSS	Cool 4 deg C	PSEG01	G32	TP-8	03/01/2018	SM2540 D
03/07/2018	Water	J1792-15	TSS	Cool 4 deg C	PSEG01	G32	TP-13	03/01/2018	SM2540 D
03/08/2018	Water	J1794-01	TSS	Cool 4 deg C	HDR103	D42	030118-SIDI-F-D-M	03/01/2018	SM2540 D
03/08/2018	Water	J1794-02	TSS	Cool 4 deg C	HDR103	D42	030118-SIDI-F-U-M	03/01/2018	SM2540 D
03/08/2018	Water	J1794-03	TSS	Cool 4 deg C	HDR103	D42	030118-JETT-F-D-M	03/01/2018	SM2540 D
03/08/2018	Water	J1794-04	TSS	Cool 4 deg C	HDR103	D42	030118-JETT-F-U-M	03/01/2018	SM2540 D
03/09/2018	Water	J1796-01	TSS	Cool 4 deg C	ARAM01	E32	COMP	03/02/2018	SM2540 D

Date/Time 03/05/18, 8:00
 Received by: AS
 Relinquished by: CP

Date/Time 03/05/18, 9:10
 Received by: AS
 Relinquished by: AS